

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE121119\
 Data File : BE100855.D
 Acq On : 11 Dec 2019 20:34
 Operator : JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Dec 12 04:43:30 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE120419.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Dec 04 17:18:11 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	94	0.00
2	1,4-Dioxane	0.729	0.838	-15.0	108	0.00
3	n-Nitrosodimethylamine	0.594	0.640	-7.7	116	0.00
4 S	2-Fluorophenol	0.949	1.080	-13.8	114	0.00
5 S	Phenol-d6	1.433	1.718	-19.9	121	0.00
6	bis(2-Chloroethyl)ether	1.385	1.604	-15.8	115	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	104	0.00
8 S	Nitrobenzene-d5	0.190	0.263	-38.4#	165#	0.00
9	Naphthalene	4.250	4.835	-13.8	123	0.00
10	Hexachlorobutadiene	0.166	0.184	-10.8	119	0.00
11 SURR	2-Methylnaphthalene-d10	0.580	0.667	-15.0	128	0.00
12	2-Methylnaphthalene	0.664	0.765	-15.2	128	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	109	0.00
14 S	2,4,6-Tribromophenol	0.044	0.067	-52.3#	188#	-0.01
15 S	2-Fluorobiphenyl	1.587	1.671	-5.3	120	-0.01
16	Acenaphthylene	1.655	1.835	-10.9	140	0.00
17	Acenaphthene	1.140	1.293	-13.4	136	0.00
18	Fluorene	1.471	1.655	-12.5	138	-0.01
19 I	Phenanthrene-d10	1.000	1.000	0.0	104	-0.01
20	4-Bromophenyl-phenylether	0.196	0.219	-11.7	128	-0.01
21	Hexachlorobenzene	0.213	0.241	-13.1	125	0.00
22	Pentachlorophenol	0.015	0.019	-26.7#	704#	-0.01
23	Phenanthrene	1.212	1.354	-11.7	123	0.00
24	Anthracene	1.013	1.058	-4.4	126	0.00
25 SURR	Fluoranthene-d10	4.728	4.365	7.7	106	-0.01
26	Fluoranthene	1.446	1.383	4.4	110	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	81	0.00
28	Pvrene	1.300	1.635	-25.8#	109	-0.01
29 S	Terphenyl-d14	0.958	1.094	-14.2	97	0.00
30	Benzo(a)anthracene	1.179	1.306	-10.8	101	-0.01
31	Chrysene	1.385	1.677	-21.1	101	0.00
32	Bis(2-ethylhexyl)phthalate	1.069	1.210	-13.2	121	-0.01
33	Indeno(1,2,3-cd)pyrene	1.207	1.464	-21.3	111	-0.03
34 I	Perylene-d12	1.000	1.000	0.0	91	-0.02
35	Benzo(b)fluoranthene	1.095	1.172	-7.0	103	-0.01
36	Benzo(k)fluoranthene	1.228	1.331	-8.4	104	-0.01
37 C	Benzo(a)pyrene	0.948	0.983	-3.7	104	-0.02
38	Dibenzo(a,h)anthracene	0.943	1.010	-7.1	110	-0.03
39	Benzo(a,h,i)perylene	1.009	1.119	-10.9	108	-0.03

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE121119\
Data File : BE100855.D
Acq On : 11 Dec 2019 20:34
Operator : JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Dec 12 04:43:30 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE120419.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Dec 04 17:18:11 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			