

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE121119\
 Data File : BE100855.D
 Acq On : 11 Dec 2019 20:34
 Operator : JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Dec 12 04:43:30 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE120419.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Dec 04 17:18:11 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	94	0.00
2	1,4-Dioxane	0.400	0.460	-15.0	108	0.00
3	n-Nitrosodimethylamine	0.400	0.431	-7.7	116	0.00
4 S	2-Fluorophenol	0.400	0.455	-13.7	114	0.00
5 S	Phenol-d6	0.400	0.480	-20.0	121	0.00
6	bis(2-Chloroethyl)ether	0.400	0.463	-15.8	115	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	104	0.00
8 S	Nitrobenzene-d5	0.400	0.552	-38.0#	165	0.00
9	Naphthalene	0.400	0.455	-13.7	123	0.00
10	Hexachlorobutadiene	0.400	0.443	-10.7	119	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.460	-15.0	128	0.00
12	2-Methylnaphthalene	0.400	0.461	-15.3	128	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	109	0.00
14 S	2,4,6-Tribromophenol	0.400	0.602	-50.5#	188	-0.01
15 S	2-Fluorobiphenyl	0.400	0.421	-5.2	120	-0.01
16	Acenaphthylene	0.400	0.443	-10.7	140	0.00
17	Acenaphthene	0.400	0.454	-13.5	136	0.00
18	Fluorene	0.400	0.450	-12.5	138	-0.01
19 I	Phenanthrene-d10	0.400	0.400	0.0	104	-0.01
20	4-Bromophenyl-phenylether	0.400	0.446	-11.5	128	-0.01
21	Hexachlorobenzene	0.400	0.452	-13.0	125	0.00
22	Pentachlorophenol	0.400	0.523	-30.8#	704	-0.01
23	Phenanthrene	0.400	0.447	-11.7	123	0.00
24	Anthracene	0.400	0.418	-4.5	126	0.00
25 SURR	Fluoranthene-d10	0.400	0.369	7.8	106	-0.01
26	Fluoranthene	0.400	0.383	4.3	110	0.00
27 I	Chrysene-d12	0.400	0.400	0.0	81	0.00
28	Pvrene	0.400	0.503	-25.7#	109	-0.01
29 S	Terphenyl-d14	0.400	0.457	-14.2	97	0.00
30	Benzo(a)anthracene	0.400	0.443	-10.7	101	-0.01
31	Chrysene	0.400	0.484	-21.0	101	0.00
32	Bis(2-ethylhexyl)phthalate	0.400	0.581	-45.2#	121	-0.01
33	Indeno(1,2,3-cd)pyrene	0.400	0.485	-21.2	111	-0.03
34 I	Perylene-d12	0.400	0.400	0.0	91	-0.02
35	Benzo(b)fluoranthene	0.400	0.428	-7.0	103	-0.01
36	Benzo(k)fluoranthene	0.400	0.433	-8.2	104	-0.01
37 C	Benzo(a)pyrene	0.400	0.415	-3.7	104	-0.02
38	Dibenzo(a,h)anthracene	0.400	0.428	-7.0	110	-0.03
39	Benzo(a,h,i)perylene	0.400	0.444	-11.0	108	-0.03

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE121119\
Data File : BE100855.D
Acq On : 11 Dec 2019 20:34
Operator : JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Dec 12 04:43:30 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE120419.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Dec 04 17:18:11 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			