

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE121119\
 Data File : BE100857.D
 Acq On : 11 Dec 2019 23:03
 Operator : JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Dec 12 04:43:44 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE120419.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Dec 04 17:18:11 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	93	-0.01
2	1,4-Dioxane	0.729	0.814	-11.7	104	-0.01
3	n-Nitrosodimethylamine	0.594	0.611	-2.9	110	-0.01
4 S	2-Fluorophenol	0.949	1.186	-25.0	124	-0.01
5 S	Phenol-d6	1.433	1.895	-32.2#	133	0.00
6	bis(2-Chloroethyl)ether	1.385	1.596	-15.2	113	-0.01
7 I	Naphthalene-d8	1.000	1.000	0.0	104	0.00
8 S	Nitrobenzene-d5	0.190	0.264	-38.9#	166#	0.00
9	Naphthalene	4.250	4.812	-13.2	123	0.00
10	Hexachlorobutadiene	0.166	0.182	-9.6	118	0.00
11 SURR	2-Methylnaphthalene-d10	0.580	0.659	-13.6	126	0.00
12	2-Methylnaphthalene	0.664	0.777	-17.0	129	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	113	0.00
14 S	2,4,6-Tribromophenol	0.044	0.073	-65.9#	211#	-0.01
15 S	2-Fluorobiphenyl	1.587	1.637	-3.2	122	-0.01
16	Acenaphthylene	1.655	1.823	-10.2	144	0.00
17	Acenaphthene	1.140	1.281	-12.4	139	0.00
18	Fluorene	1.471	1.617	-9.9	139	-0.01
19 I	Phenanthrene-d10	1.000	1.000	0.0	105	-0.01
20	4-Bromophenyl-phenylether	0.196	0.218	-11.2	128	-0.01
21	Hexachlorobenzene	0.213	0.248	-16.4	130	0.00
22	Pentachlorophenol	0.015	0.032	-113.3#	1189#	-0.01
23	Phenanthrene	1.212	1.384	-14.2	127	0.00
24	Anthracene	1.013	1.063	-4.9	128	0.00
25 SURR	Fluoranthene-d10	4.728	4.403	6.9	108	0.00
26	Fluoranthene	1.446	1.406	2.8	113	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	84	0.00
28	Pvrene	1.300	1.636	-25.8#	113	0.00
29 S	Terphenyl-d14	0.958	1.087	-13.5	100	0.00
30	Benzo(a)anthracene	1.179	1.347	-14.2	108	-0.01
31	Chrysene	1.385	1.675	-20.9	105	0.00
32	Bis(2-ethylhexyl)phthalate	1.069	1.247	-16.7	130	-0.01
33	Indeno(1,2,3-cd)pyrene	1.207	1.492	-23.6	118	-0.02
34 I	Perylene-d12	1.000	1.000	0.0	97	-0.01
35	Benzo(b)fluoranthene	1.095	1.174	-7.2	110	-0.01
36	Benzo(k)fluoranthene	1.228	1.321	-7.6	110	-0.01
37 C	Benzo(a)pyrene	0.948	0.980	-3.4	110	-0.01
38	Dibenzo(a,h)anthracene	0.943	1.022	-8.4	119	-0.03
39	Benzo(a,h,i)perylene	1.009	1.096	-8.6	112	-0.02

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE121119\
Data File : BE100857.D
Acq On : 11 Dec 2019 23:03
Operator : JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Dec 12 04:43:44 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE120419.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Dec 04 17:18:11 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			