

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE121319\
 Data File : BE100878.D
 Acq On : 13 Dec 2019 9:35
 Operator : JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Dec 13 10:49:01 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE120419.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Dec 13 10:42:38 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	96	0.00
2	1,4-Dioxane	0.729	0.870	-19.3	115	0.00
3	n-Nitrosodimethylamine	0.594	0.684	-15.2	127	0.00
4 S	2-Fluorophenol	0.949	1.132	-19.3	122	0.00
5 S	Phenol-d6	1.433	1.686	-17.7	122	0.00
6	bis(2-Chloroethyl)ether	1.385	1.595	-15.2	117	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	98	0.00
8 S	Nitrobenzene-d5	0.190	0.270	-42.1#	160#	0.00
9	Naphthalene	4.250	4.904	-15.4	118	0.00
10	Hexachlorobutadiene	0.166	0.186	-12.0	114	0.00
11 SURR	2-Methylnaphthalene-d10	0.580	0.655	-12.9	118	0.00
12	2-Methylnaphthalene	0.664	0.774	-16.6	122	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	100	0.00
14 S	2,4,6-Tribromophenol	0.044	0.068	-54.5#	176#	0.00
15 S	2-Fluorobiphenyl	1.587	1.716	-8.1	113	0.00
16	Acenaphthylene	1.655	1.842	-11.3	129	0.00
17	Acenaphthene	1.140	1.327	-16.4	128	0.00
18	Fluorene	1.471	1.676	-13.9	128	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	96	0.00
20	4-Bromophenyl-phenylether	0.196	0.221	-12.8	120	0.00
21	Hexachlorobenzene	0.213	0.256	-20.2	123	0.00
22	Pentachlorophenol	0.015	0.023	-53.3#	762#	0.00
23	Phenanthrene	1.212	1.393	-14.9	118	0.00
24	Anthracene	1.013	1.074	-6.0	119	0.00
25 SURR	Fluoranthene-d10	4.728	4.558	3.6	103	0.00
26	Fluoranthene	1.446	1.462	-1.1	108	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	80	0.00
28	Pvrene	1.300	1.617	-24.4	107	0.00
29 S	Terphenyl-d14	0.958	1.101	-14.9	96	0.00
30	Benzo(a)anthracene	1.179	1.333	-13.1	102	0.00
31	Chrysene	1.385	1.693	-22.2	101	0.00
32	Bis(2-ethylhexyl)phthalate	1.069	1.162	-8.7	115	0.00
33	Indeno(1,2,3-cd)pyrene	1.207	1.420	-17.6	107	0.00
34 I	Perylene-d12	1.000	1.000	0.0	89	0.00
35	Benzo(b)fluoranthene	1.095	1.158	-5.8	100	0.00
36	Benzo(k)fluoranthene	1.228	1.398	-13.8	107	0.00
37 C	Benzo(a)pyrene	0.948	0.973	-2.6	101	0.00
38	Dibenzo(a,h)anthracene	0.943	0.996	-5.6	106	0.00
39	Benzo(a,h,i)perylene	1.009	1.096	-8.6	103	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE121319\
Data File : BE100878.D
Acq On : 13 Dec 2019 9:35
Operator : JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Dec 13 10:49:01 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE120419.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Dec 13 10:42:38 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			