

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE121319\
 Data File : BE100897.D
 Acq On : 13 Dec 2019 22:13
 Operator : JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Dec 14 03:07:57 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE120419.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Dec 13 10:42:38 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	95	0.00
2	1,4-Dioxane	0.729	0.740	-1.5	97	0.00
3	n-Nitrosodimethylamine	0.594	0.602	-1.3	111	0.00
4 S	2-Fluorophenol	0.949	1.025	-8.0	110	0.00
5 S	Phenol-d6	1.433	1.559	-8.8	112	0.00
6	bis(2-Chloroethyl)ether	1.385	1.351	2.5	98	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	101	0.00
8 S	Nitrobenzene-d5	0.190	0.251	-32.1#	152#	0.00
9	Naphthalene	4.250	4.099	3.6	101	0.01
10	Hexachlorobutadiene	0.166	0.155	6.6	97	0.00
11 SURR	2-Methylnaphthalene-d10	0.580	0.554	4.5	102	0.00
12	2-Methylnaphthalene	0.664	0.627	5.6	101	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	105	0.01
14 S	2,4,6-Tribromophenol	0.044	0.063	-43.2#	171#	0.00
15 S	2-Fluorobiphenyl	1.587	1.461	7.9	102	0.00
16	Acenaphthylene	1.655	1.487	10.2	110	0.00
17	Acenaphthene	1.140	1.071	6.1	109	0.00
18	Fluorene	1.471	1.377	6.4	111	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	104	0.00
20	4-Bromophenyl-phenylether	0.196	0.176	10.2	102	0.00
21	Hexachlorobenzene	0.213	0.203	4.7	105	0.00
22	Pentachlorophenol	0.015	0.019	-26.7#	700#	0.00
23	Phenanthrene	1.212	1.147	5.4	104	0.00
24	Anthracene	1.013	0.880	13.1	105	0.00
25 SURR	Fluoranthene-d10	4.728	3.849	18.6	93	0.00
26	Fluoranthene	1.446	1.191	17.6	95	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	87	0.00
28	Pvrene	1.300	1.306	-0.5	94	0.00
29 S	Terphenyl-d14	0.958	0.932	2.7	89	0.00
30	Benzo(a)anthracene	1.179	1.156	2.0	96	0.00
31	Chrysene	1.385	1.347	2.7	88	0.01
32	Bis(2-ethylhexyl)phthalate	1.069	1.334	-24.8	144	0.00
33	Indeno(1,2,3-cd)pyrene	1.207	1.214	-0.6	99	0.00
34 I	Perylene-d12	1.000	1.000	0.0	104	0.00
35	Benzo(b)fluoranthene	1.095	0.892	18.5	89	0.00
36	Benzo(k)fluoranthene	1.228	1.119	8.9	99	0.00
37 C	Benzo(a)pyrene	0.948	0.759	19.9	91	0.00
38	Dibenzo(a,h)anthracene	0.943	0.802	15.0	99	0.00
39	Benzo(a,h,i)perylene	1.009	0.871	13.7	95	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE121319\
Data File : BE100897.D
Acq On : 13 Dec 2019 22:13
Operator : JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Dec 14 03:07:57 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE120419.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Dec 13 10:42:38 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			