

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE121319\
 Data File : BE100913.D
 Acq On : 14 Dec 2019 8:07
 Operator : JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Dec 16 00:44:38 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE120419.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Dec 13 10:42:38 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	97	0.00
2	1,4-Dioxane	0.729	0.698	4.3	93	0.00
3	n-Nitrosodimethylamine	0.594	0.621	-4.5	116	0.00
4 S	2-Fluorophenol	0.949	1.148	-21.0	125	0.00
5 S	Phenol-d6	1.433	1.728	-20.6	126	0.00
6	bis(2-Chloroethyl)ether	1.385	1.383	0.1	102	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	106	0.00
8 S	Nitrobenzene-d5	0.190	0.280	-47.4#	179#	0.00
9	Naphthalene	4.250	4.058	4.5	106	0.01
10	Hexachlorobutadiene	0.166	0.153	7.8	101	0.00
11 SURR	2-Methylnaphthalene-d10	0.580	0.569	1.9	111	0.00
12	2-Methylnaphthalene	0.664	0.648	2.4	110	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	119	0.01
14 S	2,4,6-Tribromophenol	0.044	0.088	-100.0#	270#	-0.01
15 S	2-Fluorobiphenyl	1.587	1.391	12.4	110	0.00
16	Acenaphthylene	1.655	1.573	5.0	132	0.00
17	Acenaphthene	1.140	1.085	4.8	125	0.00
18	Fluorene	1.471	1.407	4.4	128	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	119	0.00
20	4-Bromophenyl-phenylether	0.196	0.184	6.1	123	0.00
21	Hexachlorobenzene	0.213	0.197	7.5	117	0.00
22	Pentachlorophenol	0.015	0.039	-160.0#	1620#	-0.01
23	Phenanthrene	1.212	1.135	6.4	119	0.00
24	Anthracene	1.013	0.947	6.5	130	0.00
25 SURR	Fluoranthene-d10	4.728	4.466	5.5	125	0.00
26	Fluoranthene	1.446	1.362	5.8	124	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	114	0.00
28	Pvrene	1.300	1.354	-4.2	128	0.00
29 S	Terphenyl-d14	0.958	0.967	-0.9	121	0.00
30	Benzo(a)anthracene	1.179	1.284	-8.9	140	0.00
31	Chrysene	1.385	1.318	4.8	112	0.01
32	Bis(2-ethylhexyl)phthalate	1.069	1.826	-70.8#	258#	0.00
33	Indeno(1,2,3-cd)pyrene	1.207	1.361	-12.8	146	0.00
34 I	Perylene-d12	1.000	1.000	0.0	139	0.00
35	Benzo(b)fluoranthene	1.095	0.967	11.7	129	0.00
36	Benzo(k)fluoranthene	1.228	1.119	8.9	133	0.00
37 C	Benzo(a)pyrene	0.948	0.853	10.0	137	0.00
38	Dibenzo(a,h)anthracene	0.943	0.883	6.4	146	0.00
39	Benzo(a,h,i)perylene	1.009	0.937	7.1	137	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE121319\
Data File : BE100913.D
Acq On : 14 Dec 2019 8:07
Operator : JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 35 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Dec 16 00:44:38 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE120419.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Dec 13 10:42:38 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			