

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE121319\
 Data File : BE100913.D
 Acq On : 14 Dec 2019 8:07
 Operator : JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Dec 16 00:44:38 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE120419.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Dec 13 10:42:38 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	97	0.00
2	1,4-Dioxane	0.400	0.383	4.3	93	0.00
3	n-Nitrosodimethylamine	0.400	0.418	-4.5	116	0.00
4 S	2-Fluorophenol	0.400	0.484	-21.0	125	0.00
5 S	Phenol-d6	0.400	0.482	-20.5	126	0.00
6	bis(2-Chloroethyl)ether	0.400	0.400	0.0	102	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	106	0.00
8 S	Nitrobenzene-d5	0.400	0.587	-46.7#	179	0.00
9	Naphthalene	0.400	0.382	4.5	106	0.01
10	Hexachlorobutadiene	0.400	0.370	7.5	101	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.393	1.8	111	0.00
12	2-Methylnaphthalene	0.400	0.390	2.5	110	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	119	0.01
14 S	2,4,6-Tribromophenol	0.400	0.788	-97.0#	270	-0.01
15 S	2-Fluorobiphenyl	0.400	0.351	12.3	110	0.00
16	Acenaphthylene	0.400	0.380	5.0	132	0.00
17	Acenaphthene	0.400	0.380	5.0	125	0.00
18	Fluorene	0.400	0.383	4.3	128	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	119	0.00
20	4-Bromophenyl-phenylether	0.400	0.374	6.5	123	0.00
21	Hexachlorobenzene	0.400	0.369	7.8	117	0.00
22	Pentachlorophenol	0.400	1.051	-162.7#	1620	-0.01
23	Phenanthrene	0.400	0.375	6.3	119	0.00
24	Anthracene	0.400	0.374	6.5	130	0.00
25 SURR	Fluoranthene-d10	0.400	0.378	5.5	125	0.00
26	Fluoranthene	0.400	0.377	5.8	124	0.00
27 I	Chrysene-d12	0.400	0.400	0.0	114	0.00
28	Pvrene	0.400	0.417	-4.2	128	0.00
29 S	Terphenyl-d14	0.400	0.404	-1.0	121	0.00
30	Benzo(a)anthracene	0.400	0.436	-9.0	140	0.00
31	Chrysene	0.400	0.381	4.8	112	0.01
32	Bis(2-ethylhexyl)phthalate	0.400	0.837	-109.2#	258	0.00
33	Indeno(1,2,3-cd)pyrene	0.400	0.451	-12.7	146	0.00
34 I	Perylene-d12	0.400	0.400	0.0	139	0.00
35	Benzo(b)fluoranthene	0.400	0.353	11.8	129	0.00
36	Benzo(k)fluoranthene	0.400	0.365	8.8	133	0.00
37 C	Benzo(a)pyrene	0.400	0.360	10.0	137	0.00
38	Dibenzo(a,h)anthracene	0.400	0.374	6.5	146	0.00
39	Benzo(a,h,i)perylene	0.400	0.372	7.0	137	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE121319\
Data File : BE100913.D
Acq On : 14 Dec 2019 8:07
Operator : JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 35 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Dec 16 00:44:38 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE120419.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Dec 13 10:42:38 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0				