

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE121418\
 Data File : BE098463.D
 Acq On : 15 Dec 2018 9:15
 Operator : SJ/JU
 Sample : J6358-13
 Misc :
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 BP-MH-SW4001-20181211

Quant Time: Dec 16 01:11:53 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE121018.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Dec 10 14:59:55 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.86	152	935	0.40	ng	0.00
7) Naphthalene-d8	10.65	136	3810	0.40	ng	-0.01
13) Acenaphthene-d10	14.52	164	2474	0.40	ng	-0.01
19) Phenanthrene-d10	17.27	188	6160	0.40	ng	-0.01
27) Chrysene-d12	21.45	240	7636	0.40	ng	-0.01
34) Perylene-d12	24.00	264	7243	0.40	ng	-0.01
System Monitoring Compounds						
4) 2-Fluorophenol	5.46	112	507	0.23	ng	0.01
5) Phenol-d6	7.05	99	423	0.14	ng	0.00
8) Nitrobenzene-d5	9.03	82	1090	0.31	ng	0.01
11) 2-Methylnaphthalene-d10	12.25	152	2347	0.38	ng	-0.01
14) 2,4,6-Tribromophenol	16.02	330	532	0.59	ng	0.00
15) 2-Fluorobiphenyl	13.15	172	3925	0.38	ng	0.00
25) Fluoranthene-d10	19.30	212	35460	0.45	ng	0.00
29) Terphenyl-d14	19.90	244	6247	0.34	ng	-0.01
Target Compounds						
2) 1,4-Dioxane	3.37	88	344	0.213	ng	Qvalue # 59
32) Bis(2-ethylhexyl)phthalate	21.35	149	6669	0.151	ng	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE121418\
Data File : BE098463.D
Acq On : 15 Dec 2018 9:15
Operator : SJ/JU
Sample : J6358-13
Misc :
ALS Vial : 28 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
BP-MH-SW4001-20181211

Quant Time: Dec 16 01:11:53 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE121018.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Dec 10 14:59:55 2018
Response via : Initial Calibration

