

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE121418\
 Data File : BE098435.D
 Acq On : 14 Dec 2018 14:41
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Dec 17 16:46:51 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE121018.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Dec 10 14:59:55 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	92	0.00
2	1,4-Dioxane	0.400	0.328	18.0	76	-0.01
3	n-Nitrosodimethylamine	0.400	0.311	22.3	77	0.00
4 S	2-Fluorophenol	0.400	0.403	-0.8	95	0.00
5 S	Phenol-d6	0.400	0.426	-6.5	97	-0.01
6	bis(2-Chloroethyl)ether	0.400	0.350	12.5	83	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	88	-0.01
8 S	Nitrobenzene-d5	0.400	0.393	1.8	94	0.00
9	Naphthalene	0.400	0.388	3.0	87	0.00
10	Hexachlorobutadiene	0.400	0.417	-4.2	94	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.383	4.3	85	-0.02
12	2-Methylnaphthalene	0.400	0.372	7.0	83	-0.02
13 I	Acenaphthene-d10	0.400	0.400	0.0	87	0.00
14 S	2,4,6-Tribromophenol	0.400	0.448	-12.0	103	0.00
15 S	2-Fluorobiphenyl	0.400	0.358	10.5	86	0.00
16	Acenaphthylene	0.400	0.386	3.5	87	0.00
17	Acenaphthene	0.400	0.386	3.5	85	0.00
18	Fluorene	0.400	0.383	4.3	84	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	86	-0.01
20	4-Bromophenyl-phenylether	0.400	0.365	8.8	81	0.00
21	Hexachlorobenzene	0.400	0.374	6.5	82	-0.01
22	Pentachlorophenol	0.400	0.438	-9.5	97	0.00
23	Phenanthrene	0.400	0.373	6.8	81	0.00
24	Anthracene	0.400	0.347	13.3	75	0.00
25 SURR	Fluoranthene-d10	0.400	0.400	0.0	82	0.00
26	Fluoranthene	0.400	0.401	-0.3	82	0.00
27 I	Chrysene-d12	0.400	0.400	0.0	99	-0.01
28	Pvrene	0.400	0.343	14.2	83	0.00
29 S	Terphenyl-d14	0.400	0.326	18.5	82	0.00
30	Benzo(a)anthracene	0.400	0.349	12.8	89	-0.01
31	Chrysene	0.400	0.384	4.0	96	0.00
32	Bis(2-ethylhexyl)phthalate	0.400	0.351	12.3	88	-0.01
33	Indeno(1,2,3-cd)pyrene	0.400	0.393	1.8	103	-0.02
34 I	Perylene-d12	0.400	0.400	0.0	104	-0.01
35	Benzo(b)fluoranthene	0.400	0.373	6.8	100	-0.01
36	Benzo(k)fluoranthene	0.400	0.406	-1.5	110	-0.01
37 C	Benzo(a)pyrene	0.400	0.391	2.3	104	-0.02
38	Dibenzo(a,h)anthracene	0.400	0.382	4.5	103	-0.02
39	Benzo(a,h,i)perylene	0.400	0.382	4.5	103	-0.02

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE121418\
Data File : BE098435.D
Acq On : 14 Dec 2018 14:41
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Dec 17 16:46:51 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE121018.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Dec 10 14:59:55 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0				