

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE121418\
 Data File : BE098435.D
 Acq On : 14 Dec 2018 14:41
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Dec 17 16:46:51 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE121018.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Dec 10 14:59:55 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	92	0.00
2	1,4-Dioxane	0.692	0.567	18.1	76	-0.01
3	n-Nitrosodimethylamine	0.623	0.484	22.3	77	0.00
4 S	2-Fluorophenol	0.936	0.944	-0.9	95	0.00
5 S	Phenol-d6	1.326	1.411	-6.4	97	-0.01
6	bis(2-Chloroethyl)ether	1.260	1.103	12.5	83	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	88	-0.01
8 S	Nitrobenzene-d5	0.364	0.358	1.6	94	0.00
9	Naphthalene	4.025	3.901	3.1	87	0.00
10	Hexachlorobutadiene	0.256	0.267	-4.3	94	0.00
11 SURR	2-Methylnaphthalene-d10	0.650	0.622	4.3	85	-0.02
12	2-Methylnaphthalene	0.654	0.608	7.0	83	-0.02
13 I	Acenaphthene-d10	1.000	1.000	0.0	87	0.00
14 S	2,4,6-Tribromophenol	0.147	0.164	-11.6	103	0.00
15 S	2-Fluorobiphenyl	1.687	1.511	10.4	86	0.00
16	Acenaphthylene	1.874	1.809	3.5	87	0.00
17	Acenaphthene	1.126	1.085	3.6	85	0.00
18	Fluorene	1.506	1.441	4.3	84	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	86	-0.01
20	4-Bromophenyl-phenylether	0.215	0.196	8.8	81	0.00
21	Hexachlorobenzene	0.231	0.217	6.1	82	-0.01
22	Pentachlorophenol	0.046	0.041	10.9	97	0.00
23	Phenanthrene	0.972	0.907	6.7	81	0.00
24	Anthracene	0.866	0.751	13.3	75	0.00
25 SURR	Fluoranthene-d10	5.131	5.128	0.1	82	0.00
26	Fluoranthene	1.286	1.290	-0.3	82	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	99	-0.01
28	Pvrene	1.255	1.075	14.3	83	0.00
29 S	Terphenyl-d14	0.972	0.792	18.5	82	0.00
30	Benzo(a)anthracene	1.139	0.995	12.6	89	-0.01
31	Chrysene	1.193	1.146	3.9	96	0.00
32	Bis(2-ethylhexyl)phthalate	2.313	2.028	12.3	88	-0.01
33	Indeno(1,2,3-cd)pyrene	1.187	1.166	1.8	103	-0.02
34 I	Perylene-d12	1.000	1.000	0.0	104	-0.01
35	Benzo(b)fluoranthene	1.094	1.020	6.8	100	-0.01
36	Benzo(k)fluoranthene	1.274	1.294	-1.6	110	-0.01
37 C	Benzo(a)pyrene	1.129	1.102	2.4	104	-0.02
38	Dibenzo(a,h)anthracene	1.063	1.016	4.4	103	-0.02
39	Benzo(a,h,i)perylene	1.071	1.022	4.6	103	-0.02

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE121418\
Data File : BE098435.D
Acq On : 14 Dec 2018 14:41
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Dec 17 16:46:51 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE121018.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Dec 10 14:59:55 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			