

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE121418\
 Data File : BE098452.D
 Acq On : 15 Dec 2018 1:14
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Dec 17 17:03:50 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE121018.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Dec 10 14:59:55 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	88	-0.01
2	1,4-Dioxane	0.692	0.514	25.7#	66	0.00
3	n-Nitrosodimethylamine	0.623	0.563	9.6	86	0.00
4 S	2-Fluorophenol	0.936	1.137	-21.5	110	0.00
5 S	Phenol-d6	1.326	1.683	-26.9#	111	-0.01
6	bis(2-Chloroethyl)ether	1.260	1.145	9.1	83	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	88	-0.01
8 S	Nitrobenzene-d5	0.364	0.370	-1.6	98	0.00
9	Naphthalene	4.025	3.919	2.6	88	0.00
10	Hexachlorobutadiene	0.256	0.257	-0.4	91	0.00
11 SURR	2-Methylnaphthalene-d10	0.650	0.611	6.0	84	-0.01
12	2-Methylnaphthalene	0.654	0.623	4.7	86	-0.01
13 I	Acenaphthene-d10	1.000	1.000	0.0	87	-0.01
14 S	2,4,6-Tribromophenol	0.147	0.185	-25.9#	116	0.00
15 S	2-Fluorobiphenyl	1.687	1.522	9.8	87	0.00
16	Acenaphthylene	1.874	1.822	2.8	88	0.00
17	Acenaphthene	1.126	1.071	4.9	84	0.00
18	Fluorene	1.506	1.470	2.4	86	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	85	-0.01
20	4-Bromophenyl-phenylether	0.215	0.197	8.4	81	0.00
21	Hexachlorobenzene	0.231	0.217	6.1	81	-0.01
22	Pentachlorophenol	0.046	0.071	-54.3#	166#	-0.01
23	Phenanthrene	0.972	0.922	5.1	82	0.00
24	Anthracene	0.866	0.780	9.9	77	0.00
25 SURR	Fluoranthene-d10	5.131	5.140	-0.2	82	0.00
26	Fluoranthene	1.286	1.292	-0.5	82	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	101	-0.01
28	Pvrene	1.255	1.081	13.9	85	0.00
29 S	Terphenyl-d14	0.972	0.782	19.5	82	0.00
30	Benzo(a)anthracene	1.139	1.036	9.0	94	-0.01
31	Chrysene	1.193	1.132	5.1	96	0.00
32	Bis(2-ethylhexyl)phthalate	2.313	2.304	0.4	101	-0.01
33	Indeno(1,2,3-cd)pyrene	1.187	1.165	1.9	104	-0.02
34 I	Perylene-d12	1.000	1.000	0.0	107	-0.01
35	Benzo(b)fluoranthene	1.094	1.059	3.2	107	-0.01
36	Benzo(k)fluoranthene	1.274	1.204	5.5	105	-0.01
37 C	Benzo(a)pyrene	1.129	1.093	3.2	106	-0.02
38	Dibenzo(a,h)anthracene	1.063	1.010	5.0	104	-0.02
39	Benzo(a,h,i)perylene	1.071	1.011	5.6	104	-0.02

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE121418\
Data File : BE098452.D
Acq On : 15 Dec 2018 1:14
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 19 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Dec 17 17:03:50 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE121018.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Dec 10 14:59:55 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			