

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE121418\
 Data File : BE098466.D
 Acq On : 15 Dec 2018 11:08
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Dec 16 01:13:58 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE121018.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Dec 10 14:59:55 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	118	-0.01
2	1,4-Dioxane	0.692	0.520	24.9	91	0.00
3	n-Nitrosodimethylamine	0.623	0.525	15.7	108	0.00
4 S	2-Fluorophenol	0.936	1.042	-11.3	136	0.00
5 S	Phenol-d6	1.326	1.537	-15.9	137	-0.01
6	bis(2-Chloroethyl)ether	1.260	1.111	11.8	108	-0.01
7 I	Naphthalene-d8	1.000	1.000	0.0	113	-0.01
8 S	Nitrobenzene-d5	0.364	0.378	-3.8	128	0.00
9	Naphthalene	4.025	3.882	3.6	112	0.00
10	Hexachlorobutadiene	0.256	0.258	-0.8	117	0.00
11 SURR	2-Methylnaphthalene-d10	0.650	0.612	5.8	108	-0.01
12	2-Methylnaphthalene	0.654	0.613	6.3	109	-0.01
13 I	Acenaphthene-d10	1.000	1.000	0.0	111	-0.01
14 S	2,4,6-Tribromophenol	0.147	0.170	-15.6	135	0.00
15 S	2-Fluorobiphenyl	1.687	1.516	10.1	109	0.00
16	Acenaphthylene	1.874	1.805	3.7	110	0.00
17	Acenaphthene	1.126	1.056	6.2	105	0.00
18	Fluorene	1.506	1.438	4.5	106	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	104	-0.01
20	4-Bromophenyl-phenylether	0.215	0.202	6.0	102	0.00
21	Hexachlorobenzene	0.231	0.229	0.9	105	-0.01
22	Pentachlorophenol	0.046	0.063	-37.0#	182#	-0.01
23	Phenanthrene	0.972	0.929	4.4	101	0.00
24	Anthracene	0.866	0.798	7.9	97	0.00
25 SURR	Fluoranthene-d10	5.131	4.938	3.8	97	0.00
26	Fluoranthene	1.286	1.254	2.5	98	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	121	-0.01
28	Pvrene	1.255	1.075	14.3	101	0.00
29 S	Terphenyl-d14	0.972	0.777	20.1	97	-0.01
30	Benzo(a)anthracene	1.139	1.044	8.3	113	-0.01
31	Chrysene	1.193	1.145	4.0	117	-0.01
32	Bis(2-ethylhexyl)phthalate	2.313	2.203	4.8	115	-0.01
33	Indeno(1,2,3-cd)pyrene	1.187	1.219	-2.7	131	-0.02
34 I	Perylene-d12	1.000	1.000	0.0	132	-0.02
35	Benzo(b)fluoranthene	1.094	1.047	4.3	130	-0.01
36	Benzo(k)fluoranthene	1.274	1.272	0.2	137	-0.01
37 C	Benzo(a)pyrene	1.129	1.108	1.9	132	-0.02
38	Dibenzo(a,h)anthracene	1.063	0.999	6.0	127	-0.02
39	Benzo(a,h,i)perylene	1.071	1.029	3.9	130	-0.03

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE121418\
Data File : BE098466.D
Acq On : 15 Dec 2018 11:08
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Dec 16 01:13:58 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE121018.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Dec 10 14:59:55 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			