

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE121418\
 Data File : BE098466.D
 Acq On : 15 Dec 2018 11:08
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Dec 16 01:13:58 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE121018.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Dec 10 14:59:55 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	118	-0.01
2	1,4-Dioxane	0.400	0.301	24.8	91	0.00
3	n-Nitrosodimethylamine	0.400	0.337	15.8	108	0.00
4 S	2-Fluorophenol	0.400	0.445	-11.2	136	0.00
5 S	Phenol-d6	0.400	0.464	-16.0	137	-0.01
6	bis(2-Chloroethyl)ether	0.400	0.353	11.8	108	-0.01
7 I	Naphthalene-d8	0.400	0.400	0.0	113	-0.01
8 S	Nitrobenzene-d5	0.400	0.415	-3.7	128	0.00
9	Naphthalene	0.400	0.386	3.5	112	0.00
10	Hexachlorobutadiene	0.400	0.403	-0.8	117	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.376	6.0	108	-0.01
12	2-Methylnaphthalene	0.400	0.375	6.3	109	-0.01
13 I	Acenaphthene-d10	0.400	0.400	0.0	111	-0.01
14 S	2,4,6-Tribromophenol	0.400	0.462	-15.5	135	0.00
15 S	2-Fluorobiphenyl	0.400	0.359	10.3	109	0.00
16	Acenaphthylene	0.400	0.385	3.8	110	0.00
17	Acenaphthene	0.400	0.375	6.3	105	0.00
18	Fluorene	0.400	0.382	4.5	106	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	104	-0.01
20	4-Bromophenyl-phenylether	0.400	0.376	6.0	102	0.00
21	Hexachlorobenzene	0.400	0.396	1.0	105	-0.01
22	Pentachlorophenol	0.400	0.629	-57.2#	182	-0.01
23	Phenanthrene	0.400	0.382	4.5	101	0.00
24	Anthracene	0.400	0.368	8.0	97	0.00
25 SURR	Fluoranthene-d10	0.400	0.385	3.8	97	0.00
26	Fluoranthene	0.400	0.390	2.5	98	0.00
27 I	Chrysene-d12	0.400	0.400	0.0	121	-0.01
28	Pvrene	0.400	0.343	14.2	101	0.00
29 S	Terphenyl-d14	0.400	0.320	20.0	97	-0.01
30	Benzo(a)anthracene	0.400	0.366	8.5	113	-0.01
31	Chrysene	0.400	0.384	4.0	117	-0.01
32	Bis(2-ethylhexyl)phthalate	0.400	0.381	4.8	115	-0.01
33	Indeno(1,2,3-cd)pyrene	0.400	0.411	-2.7	131	-0.02
34 I	Perylene-d12	0.400	0.400	0.0	132	-0.02
35	Benzo(b)fluoranthene	0.400	0.383	4.3	130	-0.01
36	Benzo(k)fluoranthene	0.400	0.399	0.3	137	-0.01
37 C	Benzo(a)pyrene	0.400	0.393	1.8	132	-0.02
38	Dibenzo(a,h)anthracene	0.400	0.376	6.0	127	-0.02
39	Benzo(a,h,i)perylene	0.400	0.384	4.0	130	-0.03

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE121418\
Data File : BE098466.D
Acq On : 15 Dec 2018 11:08
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Dec 16 01:13:58 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE121018.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Dec 10 14:59:55 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0				