

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE121515\
 Data File : BE091352.D
 Acq On : 15 Dec 2015 9:50
 Operator : UM/SJ
 Sample : SSTDCCC0.4EC
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Dec 15 10:20:50 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SOM02.2-EPA-SIM-BE121515.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Dec 15 04:50:57 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)

1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	151#	-0.02
2 SURR	1,4-Dioxane-d8	0.281	0.262	6.8	135	0.00
3	1,4-Dioxane	0.349	0.290	16.9	128	0.00
4 I	Naphthalene-d8	1.000	1.000	0.0	204#	0.00
5	Naphthalene	0.790	0.853	-8.0	224#	0.00
6 SURR	2-Methylnaphthalene-d10	0.270	0.348	-28.9#	266#	0.00
7	2-Methylnaphthalene	0.310	0.405	-30.6#	268#	-0.01
8 I	Acenaphthene-d10	1.000	1.000	0.0	263#	0.00
9	Acenaphthylene	1.862	1.868	-0.3	269#	0.00
10 C	Acenaphthene	1.459	1.388	4.9	248#	0.00
11	Fluorene	1.852	1.904	-2.8	290#	0.00
12 I	Phenanthrene-d10	1.000	1.000	0.0	265#	-0.01
13	Pentachlorophenol	0.029	0.029	0.0	294#	-0.02
14	Phenanthrene	3.698	4.152	-12.3	300#	-0.01
15	Anthracene	3.001	3.473	-15.7	315#	-0.01
16 SURR	Fluoranthene-d10	0.689	0.733	-6.4	277#	0.00
17 C	Fluoranthene	3.392	3.779	-11.4	305#	0.00
18 I	Chrysene-d12	1.000	1.000	0.0	336#	0.00
19	Pyrene	4.453	3.998	10.2	283#	0.00
20	Benzo(a)anthracene	0.794	0.685	13.7	277#	0.00
21	Chrysene	1.137	1.203	-5.8	366#	0.00
22 I	Perylene-d12	1.000	1.000	0.0	251#	0.00
23	Benzo(b)fluoranthene	0.987	1.166	-18.1	306#	0.00
24	Benzo(k)fluoranthene	1.197	1.371	-14.5	292#	0.00
25 C	Benzo(a)pyrene	1.020	1.106	-8.4	279#	0.00
26	Indeno(1,2,3-cd)pyrene	4.177	4.807	-15.1	292#	0.00
27	Dibenzo(a,h)anthracene	0.781	0.827	-5.9	260#	0.00
28	Benzo(g,h,i)perylene	4.368	4.429	-1.4	252#	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0