

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE121517\
 Data File : BE095002.D
 Acq On : 15 Dec 2017 15:15
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Dec 15 16:32:54 2017
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE121417.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Dec 14 18:42:53 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	94	0.00
2	1,4-Dioxane	0.394	0.377	4.3	85	0.00
3	n-Nitrosodimethylamine	0.466	0.471	-1.1	95	0.00
4 S	2-Fluorophenol	0.621	0.652	-5.0	103	0.00
5 S	Phenol-d6	0.923	0.980	-6.2	103	0.00
6	bis(2-Chloroethyl)ether	0.901	0.901	0.0	95	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	100	-0.02
8 S	Nitrobenzene-d5	0.328	0.366	-11.6	117	0.00
9	Naphthalene	4.697	4.578	2.5	98	-0.02
10	Hexachlorobutadiene	0.214	0.204	4.7	95	-0.01
11 SURR	2-Methylnaphthalene-d10	0.655	0.640	2.3	99	-0.01
12	2-Methylnaphthalene	1.172	1.060	9.6	92	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	100	-0.02
14 S	2,4,6-Tribromophenol	0.093	0.110	-18.3	127	-0.02
15 S	2-Fluorobiphenyl	1.772	1.729	2.4	96	0.00
16	Acenaphthylene	2.138	2.077	2.9	102	-0.02
17	Acenaphthene	1.418	1.390	2.0	99	-0.02
18	Fluorene	1.762	1.693	3.9	98	-0.02
19 I	Phenanthrene-d10	1.000	1.000	0.0	100	-0.02
20	4-Bromophenyl-phenylether	0.234	0.226	3.4	100	-0.02
21	Hexachlorobenzene	0.239	0.226	5.4	96	-0.02
22	Pentachlorophenol	0.046	0.059	-28.3#	135	-0.02
23	Phenanthrene	1.246	1.207	3.1	100	-0.02
24	Anthracene	1.279	1.220	4.6	101	-0.02
25 SURR	Fluoranthene-d10	5.189	4.943	4.7	100	-0.01
26	Fluoranthene	1.538	1.457	5.3	98	-0.01
27 I	Chrysene-d12	1.000	1.000	0.0	103	-0.02
28	Pyrene	1.355	1.299	4.1	96	-0.02
29 S	Terphenyl-d14	0.753	0.718	4.6	98	-0.02
30	Benzo(a)anthracene	1.169	1.115	4.6	100	-0.02
31	Chrysene	1.316	1.235	6.2	97	-0.02
32	Bis(2-ethylhexyl)phthalate	1.489	1.612	-8.3	125	-0.03
33	Indeno(1,2,3-cd)pyrene	1.071	1.052	1.8	99	-0.05
34 I	Perylene-d12	1.000	1.000	0.0	99	-0.03
35	Benzo(b)fluoranthene	1.469	1.396	5.0	94	-0.02
36	Benzo(k)fluoranthene	1.488	1.401	5.8	98	-0.03
37 C	Benzo(a)pyrene	1.303	1.265	2.9	97	-0.02
38	Dibenzo(a,h)anthracene	1.209	1.164	3.7	97	-0.04
39	Benzo(g,h,i)perylene	1.203	1.163	3.3	98	-0.05

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE121517\
Data File : BE095002.D
Acq On : 15 Dec 2017 15:15
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Dec 15 16:32:54 2017
Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE121417.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Dec 14 18:42:53 2017
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			