

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE121518\
 Data File : BE098468.D
 Acq On : 15 Dec 2018 12:29
 Operator : SJ/JU
 Sample : SSTDC00.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDC00.4

Quant Time: Dec 16 02:33:09 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE121018.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Dec 10 14:59:55 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	53	0.00
2	1,4-Dioxane	0.692	0.593	14.3	46#	0.00
3	n-Nitrosodimethylamine	0.623	0.431	30.8#	39#	0.01
4 S	2-Fluorophenol	0.936	1.119	-19.6	65	0.01
5 S	Phenol-d6	1.326	1.641	-23.8	65	0.00
6	bis(2-Chloroethyl)ether	1.260	1.180	6.3	51	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	53	-0.01
8 S	Nitrobenzene-d5	0.364	0.348	4.4	56	0.01
9	Naphthalene	4.025	4.140	-2.9	56	0.00
10	Hexachlorobutadiene	0.256	0.288	-12.5	62	0.00
11 SURR	2-Methylnaphthalene-d10	0.650	0.662	-1.8	55	0.00
12	2-Methylnaphthalene	0.654	0.625	4.4	52	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	52	0.00
14 S	2,4,6-Tribromophenol	0.147	0.173	-17.7	65	0.00
15 S	2-Fluorobiphenyl	1.687	1.624	3.7	55	0.00
16	Acenaphthylene	1.874	1.953	-4.2	56	0.00
17	Acenaphthene	1.126	1.149	-2.0	54	0.00
18	Fluorene	1.506	1.535	-1.9	54	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	49#	-0.01
20	4-Bromophenyl-phenylether	0.215	0.205	4.7	49#	0.00
21	Hexachlorobenzene	0.231	0.244	-5.6	53	0.00
22	Pentachlorophenol	0.046	0.053	-15.2	72	0.00
23	Phenanthrene	0.972	0.992	-2.1	51	0.00
24	Anthracene	0.866	0.805	7.0	46#	0.00
25 SURR	Fluoranthene-d10	5.131	5.429	-5.8	50	0.00
26	Fluoranthene	1.286	1.368	-6.4	50	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	57	0.00
28	Pvrene	1.255	1.195	4.8	53	0.00
29 S	Terphenyl-d14	0.972	0.814	16.3	48#	0.00
30	Benzo(a)anthracene	1.139	1.146	-0.6	59	-0.01
31	Chrysene	1.193	1.214	-1.8	59	0.00
32	Bis(2-ethylhexyl)phthalate	2.313	2.243	3.0	56	-0.01
33	Indeno(1,2,3-cd)pyrene	1.187	1.211	-2.0	61	-0.01
34 I	Perylene-d12	1.000	1.000	0.0	61	-0.01
35	Benzo(b)fluoranthene	1.094	1.104	-0.9	63	0.00
36	Benzo(k)fluoranthene	1.274	1.365	-7.1	68	-0.01
37 C	Benzo(a)pyrene	1.129	1.117	1.1	61	-0.01
38	Dibenzo(a,h)anthracene	1.063	1.020	4.0	60	0.00
39	Benzo(a,h,i)perylene	1.071	1.074	-0.3	63	-0.01

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE121518\
Data File : BE098468.D
Acq On : 15 Dec 2018 12:29
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Dec 16 02:33:09 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE121018.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Dec 10 14:59:55 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			