

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE121518\
 Data File : BE098485.D
 Acq On : 15 Dec 2018 23:09
 Operator : SJ/JU
 Sample : SSTDC00.4
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDC00.4

Quant Time: Dec 16 02:58:19 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE121018.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Dec 10 14:59:55 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	46#	0.00
2	1,4-Dioxane	0.692	0.601	13.2	41#	0.00
3	n-Nitrosodimethylamine	0.623	0.344	44.8#	28#	0.00
4 S	2-Fluorophenol	0.936	1.061	-13.4	54	0.00
5 S	Phenol-d6	1.326	1.380	-4.1	48#	0.00
6	bis(2-Chloroethyl)ether	1.260	1.106	12.2	42#	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	45#	-0.01
8 S	Nitrobenzene-d5	0.364	0.331	9.1	44#	0.01
9	Naphthalene	4.025	4.174	-3.7	48#	0.00
10	Hexachlorobutadiene	0.256	0.299	-16.8	54	0.00
11 SURR	2-Methylnaphthalene-d10	0.650	0.660	-1.5	46#	0.00
12	2-Methylnaphthalene	0.654	0.635	2.9	45#	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	44#	0.00
14 S	2,4,6-Tribromophenol	0.147	0.145	1.4	46#	0.00
15 S	2-Fluorobiphenyl	1.687	1.698	-0.7	49#	0.00
16	Acenaphthylene	1.874	1.964	-4.8	48#	0.00
17	Acenaphthene	1.126	1.136	-0.9	45#	0.00
18	Fluorene	1.506	1.567	-4.1	47#	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	42#	-0.01
20	4-Bromophenyl-phenylether	0.215	0.201	6.5	41#	0.00
21	Hexachlorobenzene	0.231	0.253	-9.5	47#	0.00
22	Pentachlorophenol	0.046	0.034	26.1#	39#	0.01
23	Phenanthrene	0.972	0.980	-0.8	43#	0.00
24	Anthracene	0.866	0.744	14.1	37#	0.00
25 SURR	Fluoranthene-d10	5.131	5.255	-2.4	42#	0.00
26	Fluoranthene	1.286	1.334	-3.7	42#	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	46#	0.00
28	Pvrene	1.255	1.177	6.2	42#	0.00
29 S	Terphenyl-d14	0.972	0.806	17.1	38#	0.00
30	Benzo(a)anthracene	1.139	1.072	5.9	44#	-0.01
31	Chrysene	1.193	1.194	-0.1	46#	0.00
32	Bis(2-ethylhexyl)phthalate	2.313	1.517	34.4#	30#	-0.01
33	Indeno(1,2,3-cd)pyrene	1.187	1.105	6.9	45#	0.00
34 I	Perylene-d12	1.000	1.000	0.0	48#	-0.01
35	Benzo(b)fluoranthene	1.094	1.077	1.6	49#	0.00
36	Benzo(k)fluoranthene	1.274	1.410	-10.7	55	0.00
37 C	Benzo(a)pyrene	1.129	1.145	-1.4	50#	0.00
38	Dibenzo(a,h)anthracene	1.063	0.917	13.7	43#	0.00
39	Benzo(a,h,i)perylene	1.071	1.097	-2.4	51	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE121518\
Data File : BE098485.D
Acq On : 15 Dec 2018 23:09
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 19 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Dec 16 02:58:19 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE121018.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Dec 10 14:59:55 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0	CCC's out = 0		