

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE121518\
 Data File : BE098502.D
 Acq On : 16 Dec 2018 11:08
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Dec 17 02:05:10 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE121018.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Dec 10 14:59:55 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	51	-0.01
2	1,4-Dioxane	0.692	0.610	11.8	45#	0.00
3	n-Nitrosodimethylamine	0.623	0.468	24.9	41#	0.00
4 S	2-Fluorophenol	0.936	1.163	-24.3	65	0.01
5 S	Phenol-d6	1.326	1.634	-23.2	62	0.00
6	bis(2-Chloroethyl)ether	1.260	1.173	6.9	49#	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	51	-0.01
8 S	Nitrobenzene-d5	0.364	0.359	1.4	55	0.00
9	Naphthalene	4.025	4.107	-2.0	54	0.00
10	Hexachlorobutadiene	0.256	0.301	-17.6	62	-0.01
11 SURR	2-Methylnaphthalene-d10	0.650	0.653	-0.5	52	-0.01
12	2-Methylnaphthalene	0.654	0.648	0.9	52	-0.01
13 I	Acenaphthene-d10	1.000	1.000	0.0	49#	-0.01
14 S	2,4,6-Tribromophenol	0.147	0.187	-27.2#	66	0.00
15 S	2-Fluorobiphenyl	1.687	1.731	-2.6	56	0.00
16	Acenaphthylene	1.874	1.925	-2.7	53	0.00
17	Acenaphthene	1.126	1.148	-2.0	51	0.00
18	Fluorene	1.506	1.563	-3.8	52	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	47#	-0.01
20	4-Bromophenyl-phenylether	0.215	0.215	0.0	49#	0.00
21	Hexachlorobenzene	0.231	0.252	-9.1	52	-0.01
22	Pentachlorophenol	0.046	0.056	-21.7	73	0.00
23	Phenanthrene	0.972	1.009	-3.8	49#	0.00
24	Anthracene	0.866	0.824	4.8	45#	0.00
25 SURR	Fluoranthene-d10	5.131	5.325	-3.8	47#	0.00
26	Fluoranthene	1.286	1.349	-4.9	47#	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	54	-0.01
28	Pvrene	1.255	1.224	2.5	51	0.00
29 S	Terphenyl-d14	0.972	0.832	14.4	46#	0.00
30	Benzo(a)anthracene	1.139	1.103	3.2	53	-0.01
31	Chrysene	1.193	1.260	-5.6	57	0.00
32	Bis(2-ethylhexyl)phthalate	2.313	1.778	23.1	41#	-0.01
33	Indeno(1,2,3-cd)pyrene	1.187	1.280	-7.8	61	-0.01
34 I	Perylene-d12	1.000	1.000	0.0	56	-0.01
35	Benzo(b)fluoranthene	1.094	1.153	-5.4	61	0.00
36	Benzo(k)fluoranthene	1.274	1.360	-6.8	63	-0.01
37 C	Benzo(a)pyrene	1.129	1.203	-6.6	61	-0.01
38	Dibenzo(a,h)anthracene	1.063	1.076	-1.2	59	-0.02
39	Benzo(a,h,i)perylene	1.071	1.177	-9.9	64	-0.02

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE121518\
Data File : BE098502.D
Acq On : 16 Dec 2018 11:08
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 34 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Dec 17 02:05:10 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE121018.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Dec 10 14:59:55 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0	CCC's out = 0		