




Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE121719\
Data File : BE100945.D
Acq On : 17 Dec 2019 10:21
Operator : JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Dec 17 17:34:48 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE120419.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Dec 13 10:42:38 2019
Response via : Initial Calibration



(22) Pentachlorophenol

16.782min (-0.003) 40200.86ng

response 32606

Ion	Exp%	Act%
266.00	100	100
264.00	68.90	59.69
268.00	63.40	60.97
0.00	0.00	0.00