

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE121719\
 Data File : BE100949.D
 Acq On : 17 Dec 2019 14:17
 Operator : JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Dec 17 17:48:46 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE120419.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Dec 13 10:42:38 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	90	0.00
2	1,4-Dioxane	0.400	0.427	-6.7	97	0.00
3	n-Nitrosodimethylamine	0.400	0.492	-23.0	128	0.00
4 S	2-Fluorophenol	0.400	0.596	-49.0#	144	-0.01
5 S	Phenol-d6	0.400	0.589	-47.2#	144	0.00
6	bis(2-Chloroethyl)ether	0.400	0.465	-16.3	111	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	101	0.00
8 S	Nitrobenzene-d5	0.400	0.673	-68.3#	195	-0.01
9	Naphthalene	0.400	0.434	-8.5	114	0.00
10	Hexachlorobutadiene	0.400	0.406	-1.5	105	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.464	-16.0	125	0.00
12	2-Methylnaphthalene	0.400	0.474	-18.5	127	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	125	0.00
14 S	2,4,6-Tribromophenol	0.400	0.906	-126.5#	325	0.00
15 S	2-Fluorobiphenyl	0.400	0.360	10.0	118	0.00
16	Acenaphthylene	0.400	0.453	-13.2	164	0.00
17	Acenaphthene	0.400	0.431	-7.7	148	0.00
18	Fluorene	0.400	0.449	-12.2	158	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	131	0.00
20	4-Bromophenyl-phenylether	0.400	0.419	-4.7	151	0.00
21	Hexachlorobenzene	0.400	0.382	4.5	133	0.00
22	Pentachlorophenol	0.400	0.723	-80.7#	1219	0.00
23	Phenanthrene	0.400	0.407	-1.7	141	0.00
24	Anthracene	0.400	0.428	-7.0	162	0.00
25 SURR	Fluoranthene-d10	0.400	0.384	4.0	139	0.00
26	Fluoranthene	0.400	0.383	4.3	139	0.00
27 I	Chrysene-d12	0.400	0.400	0.0	111	0.00
28	Pvrene	0.400	0.481	-20.2	143	0.00
29 S	Terphenyl-d14	0.400	0.429	-7.2	125	0.00
30	Benzo(a)anthracene	0.400	0.500	-25.0	156	0.00
31	Chrysene	0.400	0.450	-12.5	129	0.01
32	Bis(2-ethylhexyl)phthalate	0.400	1.276	-219.0#	423	0.00
33	Indeno(1,2,3-cd)pyrene	0.400	0.537	-34.3#	169	0.00
34 I	Perylene-d12	0.400	0.400	0.0	132	0.00
35	Benzo(b)fluoranthene	0.400	0.437	-9.2	153	0.00
36	Benzo(k)fluoranthene	0.400	0.424	-6.0	148	0.00
37 C	Benzo(a)pyrene	0.400	0.457	-14.2	166	0.00
38	Dibenzo(a,h)anthracene	0.400	0.448	-12.0	167	0.00
39	Benzo(a,h,i)perylene	0.400	0.439	-9.7	154	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE121719\
Data File : BE100949.D
Acq On : 17 Dec 2019 14:17
Operator : JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Dec 17 17:48:46 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE120419.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Dec 13 10:42:38 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0				