

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE121719\
 Data File : BE100949.D
 Acq On : 17 Dec 2019 14:17
 Operator : JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Dec 17 17:48:46 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE120419.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Dec 13 10:42:38 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	90	0.00
2	1,4-Dioxane	0.729	0.779	-6.9	97	0.00
3	n-Nitrosodimethylamine	0.594	0.730	-22.9	128	0.00
4 S	2-Fluorophenol	0.949	1.415	-49.1#	144	-0.01
5 S	Phenol-d6	1.433	2.109	-47.2#	144	0.00
6	bis(2-Chloroethyl)ether	1.385	1.611	-16.3	111	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	101	0.00
8 S	Nitrobenzene-d5	0.190	0.321	-68.9#	195#	-0.01
9	Naphthalene	4.250	4.609	-8.4	114	0.00
10	Hexachlorobutadiene	0.166	0.168	-1.2	105	0.00
11 SURR	2-Methylnaphthalene-d10	0.580	0.672	-15.9	125	0.00
12	2-Methylnaphthalene	0.664	0.787	-18.5	127	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	125	0.00
14 S	2,4,6-Tribromophenol	0.044	0.101	-129.5#	325#	0.00
15 S	2-Fluorobiphenyl	1.587	1.427	10.1	118	0.00
16	Acenaphthylene	1.655	1.872	-13.1	164#	0.00
17	Acenaphthene	1.140	1.229	-7.8	148	0.00
18	Fluorene	1.471	1.651	-12.2	158#	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	131	0.00
20	4-Bromophenyl-phenylether	0.196	0.205	-4.6	151#	0.00
21	Hexachlorobenzene	0.213	0.204	4.2	133	0.00
22	Pentachlorophenol	0.015	0.027	-80.0#	1219#	0.00
23	Phenanthrene	1.212	1.234	-1.8	141	0.00
24	Anthracene	1.013	1.084	-7.0	162#	0.00
25 SURR	Fluoranthene-d10	4.728	4.542	3.9	139	0.00
26	Fluoranthene	1.446	1.386	4.1	139	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	111	0.00
28	Pvrene	1.300	1.563	-20.2	143	0.00
29 S	Terphenyl-d14	0.958	1.028	-7.3	125	0.00
30	Benzo(a)anthracene	1.179	1.474	-25.0#	156#	0.00
31	Chrysene	1.385	1.559	-12.6	129	0.01
32	Bis(2-ethylhexyl)phthalate	1.069	3.088	-188.9#	423#	0.00
33	Indeno(1,2,3-cd)pyrene	1.207	1.619	-34.1#	169#	0.00
34 I	Perylene-d12	1.000	1.000	0.0	132	0.00
35	Benzo(b)fluoranthene	1.095	1.196	-9.2	153#	0.00
36	Benzo(k)fluoranthene	1.228	1.302	-6.0	148	0.00
37 C	Benzo(a)pyrene	0.948	1.084	-14.3	166#	0.00
38	Dibenzo(a,h)anthracene	0.943	1.057	-12.1	167#	0.00
39	Benzo(a,h,i)perylene	1.009	1.107	-9.7	154#	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE121719\
Data File : BE100949.D
Acq On : 17 Dec 2019 14:17
Operator : JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Dec 17 17:48:46 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE120419.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Dec 13 10:42:38 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0	CCC's out = 0		