

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE121817\
 Data File : BE095025.D
 Acq On : 19 Dec 2017 00:10
 Operator : SJ/JU
 Sample : I6900-09DL 5X
 Misc :
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 F9C08DL

Quant Time: Dec 19 06:16:41 2017
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SOM-EPA-SIM-BE121817.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Dec 19 04:08:16 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.46	152	4631	0.40	ng/ul	0.00
2) Naphthalene-d8	11.29	136	15240	0.40	ng/ul	0.00
6) Acenaphthene-d10	15.05	164	9191	0.40	ng/ul	0.00
10) Phenanthrene-d10	17.75	188	20836	0.40	ng/ul	0.00
16) Chrysene-d12	21.84	240	21289	0.40	ng/ul	0.00
20) Perylene-d12	24.50	264	20621	0.40	ng/ul	0.00
System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	12.84	152	1546	0.06	ng/ul	0.00
14) Fluoranthene-d10	19.72	212	4356	0.07	ng/ul	0.00
Target Compounds				Qvalue		
3) Naphthalene	11.34	128	17520	0.433	ng/ul#	89
8) Acenaphthene	15.11	153	7891	0.244	ng/ul	95
9) Fluorene	16.07	166	3918	0.109	ng/ul	91
12) Phenanthrene	17.78	178	5515	0.100	ng/ul#	84
15) Fluoranthene	19.76	202	22192	0.310	ng/ul	84
17) Pyrene	20.11	202	14568	0.197	ng/ul#	82
18) Benzo(a)anthracene	21.82	228	2428	0.040	ng/ul#	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE121817\
Data File : BE095025.D
Acq On : 19 Dec 2017 00:10
Operator : SJ/JU
Sample : I6900-09DL 5X
Misc :
ALS Vial : 22 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
F9C08DL

Quant Time: Dec 19 06:16:41 2017
Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SOM-EPA-SIM-BE121817.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Dec 19 04:08:16 2017
Response via : Initial Calibration

