

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE121818\
 Data File : BE098509.D
 Acq On : 18 Dec 2018 10:10
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Dec 18 11:27:12 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE121018.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Dec 10 14:59:55 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	69	0.00
2	1,4-Dioxane	0.400	0.394	1.5	69	-0.01
3	n-Nitrosodimethylamine	0.400	0.497	-24.2	92	-0.02
4 S	2-Fluorophenol	0.400	0.484	-21.0	86	0.00
5 S	Phenol-d6	0.400	0.527	-31.8#	90	-0.01
6	bis(2-Chloroethyl)ether	0.400	0.479	-19.7	85	-0.01
7 I	Naphthalene-d8	0.400	0.400	0.0	76	-0.01
8 S	Nitrobenzene-d5	0.400	0.351	12.3	73	0.00
9	Naphthalene	0.400	0.428	-7.0	83	0.00
10	Hexachlorobutadiene	0.400	0.387	3.3	76	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.428	-7.0	83	-0.01
12	2-Methylnaphthalene	0.400	0.452	-13.0	88	-0.01
13 I	Acenaphthene-d10	0.400	0.400	0.0	72	0.00
14 S	2,4,6-Tribromophenol	0.400	0.468	-17.0	89	0.00
15 S	2-Fluorobiphenyl	0.400	0.410	-2.5	81	0.00
16	Acenaphthylene	0.400	0.384	4.0	72	0.00
17	Acenaphthene	0.400	0.407	-1.7	74	0.00
18	Fluorene	0.400	0.405	-1.3	73	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	60	-0.01
20	4-Bromophenyl-phenylether	0.400	0.456	-14.0	71	0.00
21	Hexachlorobenzene	0.400	0.419	-4.7	64	0.00
22	Pentachlorophenol	0.400	0.837	-109.2#	148	-0.01
23	Phenanthrene	0.400	0.446	-11.5	68	0.00
24	Anthracene	0.400	0.435	-8.7	66	-0.01
25 SURR	Fluoranthene-d10	0.400	0.410	-2.5	59	0.00
26	Fluoranthene	0.400	0.419	-4.7	60	0.00
27 I	Chrysene-d12	0.400	0.400	0.0	64	-0.01
28	Pvrene	0.400	0.399	0.3	62	0.00
29 S	Terphenyl-d14	0.400	0.385	3.8	62	0.00
30	Benzo(a)anthracene	0.400	0.424	-6.0	69	-0.01
31	Chrysene	0.400	0.416	-4.0	67	-0.01
32	Bis(2-ethylhexyl)phthalate	0.400	6.979	-1644.7#	1114	-0.01
33	Indeno(1,2,3-cd)pyrene	0.400	0.360	10.0	60	-0.01
34 I	Perylene-d12	0.400	0.400	0.0	65	-0.02
35	Benzo(b)fluoranthene	0.400	0.479	-19.7	80	-0.01
36	Benzo(k)fluoranthene	0.400	0.441	-10.2	74	-0.02
37 C	Benzo(a)pyrene	0.400	0.422	-5.5	70	-0.01
38	Dibenzo(a,h)anthracene	0.400	0.360	10.0	60	-0.02
39	Benzo(a,h,i)perylene	0.400	0.343	14.2	57	-0.01

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE121818\
Data File : BE098509.D
Acq On : 18 Dec 2018 10:10
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Dec 18 11:27:12 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE121018.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Dec 10 14:59:55 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0				