

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE121818\
 Data File : BE098509.D
 Acq On : 18 Dec 2018 10:10
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Dec 18 11:27:12 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE121018.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Dec 10 14:59:55 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	69	0.00
2	1,4-Dioxane	0.692	0.681	1.6	69	-0.01
3	n-Nitrosodimethylamine	0.623	0.773	-24.1	92	-0.02
4 S	2-Fluorophenol	0.936	1.133	-21.0	86	0.00
5 S	Phenol-d6	1.326	1.745	-31.6#	90	-0.01
6	bis(2-Chloroethyl)ether	1.260	1.509	-19.8	85	-0.01
7 I	Naphthalene-d8	1.000	1.000	0.0	76	-0.01
8 S	Nitrobenzene-d5	0.364	0.319	12.4	73	0.00
9	Naphthalene	4.025	4.306	-7.0	83	0.00
10	Hexachlorobutadiene	0.256	0.248	3.1	76	0.00
11 SURR	2-Methylnaphthalene-d10	0.650	0.697	-7.2	83	-0.01
12	2-Methylnaphthalene	0.654	0.739	-13.0	88	-0.01
13 I	Acenaphthene-d10	1.000	1.000	0.0	72	0.00
14 S	2,4,6-Tribromophenol	0.147	0.172	-17.0	89	0.00
15 S	2-Fluorobiphenyl	1.687	1.728	-2.4	81	0.00
16	Acenaphthylene	1.874	1.797	4.1	72	0.00
17	Acenaphthene	1.126	1.146	-1.8	74	0.00
18	Fluorene	1.506	1.524	-1.2	73	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	60	-0.01
20	4-Bromophenyl-phenylether	0.215	0.245	-14.0	71	0.00
21	Hexachlorobenzene	0.231	0.242	-4.8	64	0.00
22	Pentachlorophenol	0.046	0.090	-95.7#	148	-0.01
23	Phenanthrene	0.972	1.085	-11.6	68	0.00
24	Anthracene	0.866	0.943	-8.9	66	-0.01
25 SURR	Fluoranthene-d10	5.131	5.259	-2.5	59	0.00
26	Fluoranthene	1.286	1.348	-4.8	60	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	64	-0.01
28	Pvrene	1.255	1.252	0.2	62	0.00
29 S	Terphenyl-d14	0.972	0.936	3.7	62	0.00
30	Benzo(a)anthracene	1.139	1.208	-6.1	69	-0.01
31	Chrysene	1.193	1.242	-4.1	67	-0.01
32	Bis(2-ethylhexyl)phthalate	2.313	40.365	-1645.1#	1114#	-0.01
33	Indeno(1,2,3-cd)pyrene	1.187	1.067	10.1	60	-0.01
34 I	Perylene-d12	1.000	1.000	0.0	65	-0.02
35	Benzo(b)fluoranthene	1.094	1.309	-19.7	80	-0.01
36	Benzo(k)fluoranthene	1.274	1.406	-10.4	74	-0.02
37 C	Benzo(a)pyrene	1.129	1.192	-5.6	70	-0.01
38	Dibenzo(a,h)anthracene	1.063	0.956	10.1	60	-0.02
39	Benzo(a,h,i)perylene	1.071	0.918	14.3	57	-0.01

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE121818\
Data File : BE098509.D
Acq On : 18 Dec 2018 10:10
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Dec 18 11:27:12 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE121018.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Dec 10 14:59:55 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			