

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE121818\
 Data File : BE098520.D
 Acq On : 18 Dec 2018 19:39
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Dec 19 03:25:31 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE121018.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Dec 10 14:59:55 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	73	-0.01
2	1,4-Dioxane	0.692	0.645	6.8	69	-0.01
3	n-Nitrosodimethylamine	0.623	0.452	27.4#	57	0.00
4 S	2-Fluorophenol	0.936	1.014	-8.3	81	0.00
5 S	Phenol-d6	1.326	1.680	-26.7#	91	-0.01
6	bis(2-Chloroethyl)ether	1.260	1.217	3.4	73	-0.01
7 I	Naphthalene-d8	1.000	1.000	0.0	85	-0.01
8 S	Nitrobenzene-d5	0.364	0.310	14.8	79	0.00
9	Naphthalene	4.025	3.995	0.7	87	-0.01
10	Hexachlorobutadiene	0.256	0.257	-0.4	88	-0.01
11 SURR	2-Methylnaphthalene-d10	0.650	0.668	-2.8	89	-0.01
12	2-Methylnaphthalene	0.654	0.652	0.3	87	-0.01
13 I	Acenaphthene-d10	1.000	1.000	0.0	82	-0.01
14 S	2,4,6-Tribromophenol	0.147	0.135	8.2	79	-0.01
15 S	2-Fluorobiphenyl	1.687	1.672	0.9	89	-0.01
16	Acenaphthylene	1.874	1.656	11.6	75	-0.01
17	Acenaphthene	1.126	1.087	3.5	80	-0.01
18	Fluorene	1.506	1.386	8.0	76	-0.01
19 I	Phenanthrene-d10	1.000	1.000	0.0	69	-0.02
20	4-Bromophenyl-phenylether	0.215	0.211	1.9	70	-0.01
21	Hexachlorobenzene	0.231	0.244	-5.6	74	-0.01
22	Pentachlorophenol	0.046	0.038	17.4	72	-0.01
23	Phenanthrene	0.972	0.934	3.9	67	-0.01
24	Anthracene	0.866	0.766	11.5	62	-0.01
25 SURR	Fluoranthene-d10	5.131	4.817	6.1	62	-0.01
26	Fluoranthene	1.286	1.188	7.6	61	-0.01
27 I	Chrysene-d12	1.000	1.000	0.0	74	-0.01
28	Pvrene	1.255	1.096	12.7	63	-0.01
29 S	Terphenyl-d14	0.972	0.827	14.9	64	-0.01
30	Benzo(a)anthracene	1.139	1.056	7.3	70	-0.01
31	Chrysene	1.193	1.133	5.0	71	-0.01
32	Bis(2-ethylhexyl)phthalate	2.313	1.909	17.5	61	-0.01
33	Indeno(1,2,3-cd)pyrene	1.187	0.845	28.8#	56	-0.02
34 I	Perylene-d12	1.000	1.000	0.0	76	-0.03
35	Benzo(b)fluoranthene	1.094	1.065	2.7	77	-0.02
36	Benzo(k)fluoranthene	1.274	1.325	-4.0	83	-0.03
37 C	Benzo(a)pyrene	1.129	1.103	2.3	76	-0.03
38	Dibenzo(a,h)anthracene	1.063	0.824	22.5	61	-0.03
39	Benzo(a,h,i)perylene	1.071	0.758	29.2#	56	-0.03

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE121818\
Data File : BE098520.D
Acq On : 18 Dec 2018 19:39
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Dec 19 03:25:31 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE121018.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Dec 10 14:59:55 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			