

Data Path : Z:\HPCHEM1\BNA E\DATA\BE121916\
 Data File : BE092608.D
 Acq On : 19 Dec 2016 15:18
 Operator : UM/SJ
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 ICVBE121916

Quant Time: Dec 19 16:20:29 2016
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE121916.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Dec 19 16:16:32 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	117	0.00
2	1,4-Dioxane	0.733	0.800	-9.1	130	0.00
3	n-Nitrosodimethylamine	0.613	0.535	12.7	133	0.00
4 S	2-Fluorophenol	0.761	0.758	0.4	136	0.00
5 S	Phenol-d6	1.090	1.027	5.8	140	-0.02
6	bis(2-Chloroethyl)ether	1.132	0.975	13.9	126	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	114	0.00
8 S	Nitrobenzene-d5	0.280	0.255	8.9	131	0.00
9	Naphthalene	1.026	1.020	0.6	123	0.00
10	Hexachlorobutadiene	0.156	0.156	0.0	122	0.00
11	2-Methylnaphthalene	0.653	0.617	5.5	121	0.00
12 I	Acenaphthene-d10	1.000	1.000	0.0	113	0.00
13 S	2,4,6-Tribromophenol	0.113	0.098	13.3	123	0.00
14 S	2-Fluorobiphenyl	1.196	1.170	2.2	121	0.00
15	Acenaphthylene	6.899	6.624	4.0	122	0.00
16	Acenaphthene	1.137	1.096	3.6	117	0.00
17	Fluorene	1.397	1.336	4.4	119	-0.01
18 I	Phenanthrene-d10	1.000	1.000	0.0	112	0.00
19	Hexachlorobenzene	0.214	0.206	3.7	117	0.00
20	Pentachlorophenol	0.036	0.029	19.4	130	0.00
21	Phenanthrene	1.261	1.174	6.9	119	0.00
22	Anthracene	1.152	1.032	10.4	119	0.00
23	Fluoranthene	5.447	5.098	6.4	120	0.00
24 I	Chrysene-d12	1.000	1.000	0.0	108	0.00
25	Pyrene	4.403	4.226	4.0	118	0.00
26 S	Terphenyl-d14	0.613	0.600	2.1	119	0.00
27	Benzo(a)anthracene	4.490	4.329	3.6	122	0.00
28	Chrysene	5.204	5.292	-1.7	119	0.00
29	Bis(2-ethylhexyl)phthalate	0.486	0.385	20.8	121	0.00
30	Indeno(1,2,3-cd)pyrene	4.702	4.540	3.4	122	0.00
31 I	Perylene-d12	1.000	1.000	0.0	112	0.00
32	Benzo(b)fluoranthene	1.150	1.101	4.3	123	0.00
33	Benzo(k)fluoranthene	1.282	1.186	7.5	119	0.00
34 C	Benzo(a)pyrene	1.110	1.029	7.3	124	0.00
35	Dibenzo(a,h)anthracene	1.054	0.959	9.0	121	0.00
36	Benzo(g,h,i)perylene	4.515	4.248	5.9	120	-0.02

(#) = Out of Range

SPCC's out = 0 CCC's out = 0