

Data Path : Z:\HPCHEM1\BNA E\DATA\BE121916\
 Data File : BE092615.D
 Acq On : 19 Dec 2016 19:46
 Operator : UM/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Dec 20 01:01:23 2016
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE121916.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Dec 19 16:16:32 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	104	0.00
2	1,4-Dioxane	0.733	0.784	-7.0	113	0.00
3	n-Nitrosodimethylamine	0.613	0.549	10.4	121	0.00
4 S	2-Fluorophenol	0.761	0.689	9.5	110	0.00
5 S	Phenol-d6	1.090	0.920	15.6	112	-0.02
6	bis(2-Chloroethyl)ether	1.132	0.937	17.2	108	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	102	0.00
8 S	Nitrobenzene-d5	0.280	0.240	14.3	110	0.00
9	Naphthalene	1.026	0.965	5.9	104	0.00
10	Hexachlorobutadiene	0.156	0.147	5.8	102	0.00
11	2-Methylnaphthalene	0.653	0.574	12.1	100	0.00
12 I	Acenaphthene-d10	1.000	1.000	0.0	100	0.00
13 S	2,4,6-Tribromophenol	0.113	0.089	21.2	99	0.00
14 S	2-Fluorobiphenyl	1.196	1.107	7.4	101	0.00
15	Acenaphthylene	6.899	6.249	9.4	101	0.00
16	Acenaphthene	1.137	1.072	5.7	101	0.00
17	Fluorene	1.397	1.272	8.9	100	-0.01
18 I	Phenanthrene-d10	1.000	1.000	0.0	99	0.00
19	Hexachlorobenzene	0.214	0.199	7.0	100	0.00
20	Pentachlorophenol	0.036	0.034	5.6	133	0.00
21	Phenanthrene	1.261	1.126	10.7	100	0.00
22	Anthracene	1.152	0.992	13.9	101	0.00
23	Fluoranthene	5.447	4.854	10.9	101	0.00
24 I	Chrysene-d12	1.000	1.000	0.0	95	0.00
25	Pyrene	4.403	4.074	7.5	100	0.00
26 S	Terphenyl-d14	0.613	0.573	6.5	100	-0.02
27	Benzo(a)anthracene	4.490	4.127	8.1	102	0.00
28	Chrysene	5.204	5.059	2.8	101	0.00
29	Bis(2-ethylhexyl)phthalate	0.486	0.402	17.3	111	0.00
30	Indeno(1,2,3-cd)pyrene	4.702	4.283	8.9	101	0.00
31 I	Perylene-d12	1.000	1.000	0.0	100	0.00
32	Benzo(b)fluoranthene	1.150	1.017	11.6	101	0.00
33	Benzo(k)fluoranthene	1.282	1.125	12.2	101	0.00
34 C	Benzo(a)pyrene	1.110	0.936	15.7	100	0.00
35	Dibenzo(a,h)anthracene	1.054	0.909	13.8	102	0.00
36	Benzo(g,h,i)perylene	4.515	4.025	10.9	101	-0.02

(#) = Out of Range

SPCC's out = 0 CCC's out = 0