



n-Nitrosodimethylamine

Response Ratio




Resp Ratio = $8.72 \times 10^{-1} \times \text{Amt} - 3.46 \times 10^{-2}$
Coef of Det (r^2) = 0.999 Curve Fit: Linear

Method Name: Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE121916.M
Calibration Table Last Updated: Mon Dec 19 16:16:32 2016




Method Name: Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE121916.M
Calibration Table Last Updated: Mon Dec 19 16:16:32 2016



Method Name: Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE121916.M
Calibration Table Last Updated: Mon Dec 19 16:16:32 2016


bis(2-Chloroethyl)ether

Response Ratio




Resp Ratio = 1.47e+000 * Amt - 3.09e-002
Coef of Det (r^2) = 0.999 Curve Fit: Linear


Method Name: Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE121916.M
Calibration Table Last Updated: Mon Dec 19 16:16:32 2016



Method Name: Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE121916.M
Calibration Table Last Updated: Mon Dec 19 16:16:32 2016




Method Name: Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE121916.M
Calibration Table Last Updated: Mon Dec 19 16:16:32 2016



Method Name: Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE121916.M
Calibration Table Last Updated: Mon Dec 19 16:16:32 2016

Bis(2-ethylhexyl)phthalate

Response Ratio




$R = 3.37e-001 A^2 + 4.06e-001 A - 4.44e-003$
Coef of Det (r^2) = 1.000 Curve Fit: Quadratic

Method Name: Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE121916.M
Calibration Table Last Updated: Mon Dec 19 16:16:32 2016

1,4-Dioxane

Response Ratio



Resp Ratio = $5.49 \times 10^{-1} \times \text{Amt} + 2.00 \times 10^{-2}$
Coef of Det (r^2) = 0.999 Curve Fit: Linear

Method Name: Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE121916.M
Calibration Table Last Updated: Mon Dec 19 16:16:32 2016