


Data Path : Z:\HPCHEM1\BNA_E\Data\BE121917\
 Data File : BE095027.D
 Acq On : 19 Dec 2017 10:40
 Operator : SJ/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE121417.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Thu Dec 14 18:42:53 2017



AutoFind: Scans 2646, 2647, 2648; Background Corrected with Scan 2638

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	53.5	24881	PASS
68	69	0.00	2	0.0	0	PASS
69	198	0.00	100	55.3	25724	PASS
70	69	0.00	2	0.1	32	PASS
127	198	10	80	53.3	24787	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	46546	PASS
199	198	5	9	6.7	3097	PASS
275	198	10	60	25.5	11883	PASS
365	198	1	100	3.1	1430	PASS
441	442	0.01	24	16.8	5042	PASS
442	198	50	100	64.4	29979	PASS
443	442	15	24	19.1	5735	PASS