

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE122419\
 Data File : BE100972.D
 Acq On : 24 Dec 2019 15:54
 Operator : JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 ICVBE122419

Quant Time: Dec 24 16:48:09 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE122419.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Dec 24 14:28:42 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	102	0.00
2	1,4-Dioxane	1.413	1.256	11.1	102	0.00
3	n-Nitrosodimethylamine	0.330	0.285	13.6	141	-0.02
4 S	2-Fluorophenol	0.676	0.728	-7.7	122	-0.01
5 S	Phenol-d6	0.950	0.991	-4.3	126	-0.03
6	bis(2-Chloroethyl)ether	1.446	1.383	4.4	116	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	105	0.00
8 S	Nitrobenzene-d5	0.192	0.191	0.5	118	-0.01
9	Naphthalene	4.253	4.311	-1.4	105	0.00
10	Hexachlorobutadiene	0.188	0.188	0.0	102	0.00
11 SURR	2-Methylnaphthalene-d10	0.588	0.616	-4.8	107	0.00
12	2-Methylnaphthalene	0.644	0.625	3.0	99	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	105	0.00
14 S	2,4,6-Tribromophenol	0.088	0.074	15.9	111	0.00
15 S	2-Fluorobiphenyl	1.421	1.434	-0.9	104	0.00
16	Acenaphthylene	1.495	1.388	7.2	105	0.00
17	Acenaphthene	1.086	1.081	0.5	103	0.00
18	Fluorene	1.354	1.293	4.5	103	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	112	0.00
20	4-Bromophenyl-phenylether	0.176	0.151	14.2	109	0.00
21	Hexachlorobenzene	0.207	0.189	8.7	106	0.00
22	Pentachlorophenol	0.031	0.020	35.5#	109	-0.03
23	Phenanthrene	1.031	0.956	7.3	116	0.00
24	Anthracene	1.043	0.967	7.3	114	0.00
25 SURR	Fluoranthene-d10	4.854	4.445	8.4	110	0.00
26	Fluoranthene	1.406	1.252	11.0	110	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	106	0.00
28	Pvrene	1.399	1.439	-2.9	107	0.00
29 S	Terphenyl-d14	0.919	0.937	-2.0	110	0.00
30	Benzo(a)anthracene	0.979	0.886	9.5	110	0.00
31	Chrysene	1.823	1.979	-8.6	104	0.00
32	Bis(2-ethylhexyl)phthalate	2.154	1.783	17.2	111	0.00
33	Indeno(1,2,3-cd)pyrene	1.372	1.279	6.8	103	0.00
34 I	Perylene-d12	1.000	1.000	0.0	103	0.00
35	Benzo(b)fluoranthene	0.842	0.798	5.2	100	0.00
36	Benzo(k)fluoranthene	1.597	1.478	7.5	95	0.00
37 C	Benzo(a)pyrene	0.983	0.966	1.7	104	0.00
38	Dibenzo(a,h)anthracene	0.891	0.855	4.0	120	0.00
39	Benzo(a,h,i)perylene	1.049	1.018	3.0	101	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE122419\
Data File : BE100972.D
Acq On : 24 Dec 2019 15:54
Operator : JU
Sample : SSTDICV0.4
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
ICVBE122419

Quant Time: Dec 24 16:48:09 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE122419.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Dec 24 14:28:42 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			