

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE122419\
 Data File : BE100971.D
 Acq On : 24 Dec 2019 12:55
 Operator : JU
 Sample : SSTDICC5.0
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 SSTDICC5.0

Manual Integrations
 APPROVED

mohammad
 12/26/2019 4:02:29 PM

Quant Time: Dec 24 14:34:36 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE122419.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Dec 24 11:32:12 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.75	152	2767	0.40	ng	0.00
7) Naphthalene-d8	10.52	136	13673	0.40	ng	0.00
13) Acenaphthene-d10	14.37	164	8357	0.40	ng	-0.02
19) Phenanthrene-d10	17.11	188	18463	0.40	ng	-0.01
27) Chrysene-d12	21.28	240	22183	0.40	ng	-0.01
34) Perylene-d12	23.72	264	24276	0.40	ng	0.00

System Monitoring Compounds						
4) 2-Fluorophenol	5.35	112	29281m	5.93	ng	-0.01
5) Phenol-d6	6.92	99	50391m	8.37	ng	-0.07
8) Nitrobenzene-d5	8.89	82	50665	7.01	ng	-0.03
11) 2-Methylnaphthalene-d10	12.12	152	103048	5.01	ng	0.00
14) 2,4,6-Tribromophenol	15.86	330	14822	7.22	ng	-0.01
15) 2-Fluorobiphenyl	13.00	172	151585	5.13	ng	-0.02
25) Fluoranthene-d10	19.14	212	1238419	5.36	ng	0.00
29) Terphenyl-d14	19.75	244	260639	5.15	ng	-0.01

Target Compounds				Qvalue		
3) n-Nitrosodimethylamine	3.61	42	17732	5.854	ng	# 82
6) bis(2-Chloroethyl)ether	7.17	93	50621	7.157	ng	88
9) Naphthalene	10.56	128	733610	4.967	ng	99
10) Hexachlorobutadiene	10.86	225	29061	4.604	ng	98
12) 2-Methylnaphthalene	12.19	142	112565	5.417	ng	97
16) Acenaphthylene	14.10	152	193312	5.614	ng	100
17) Acenaphthene	14.44	154	121605	5.193	ng	99
18) Fluorene	15.41	166	160284	5.409	ng	99
20) 4-Bromophenyl-phenylether	16.30	248	47055	5.730	ng	95
21) Hexachlorobenzene	16.42	284	48337	5.164	ng	100
22) Pentachlorophenol	16.77	266	12933m	9.342	ng	
23) Phenanthrene	17.15	178	266505	5.538	ng	99
24) Anthracene	17.24	178	263634m	8.152	ng	
26) Fluoranthene	19.17	202	366424	5.485	ng	98
28) Pyrene	19.53	202	376039	4.742	ng	99
30) Benzo(a)anthracene	21.27	228	327838	5.638	ng	99
31) Chrysene	21.32	228	438541m	5.984	ng	
32) Bis(2-ethylhexyl)phthalate	21.21	149	753590	4.119	ng	100
33) Indeno(1,2,3-cd)pyrene	26.28	276	395181	5.154	ng	# 91
36) Benzo(k)fluoranthene	23.02	252	461519m	6.424	ng	
37) Benzo(a)pyrene	23.61	252	331404	5.405	ng	# 92
38) Dibenzo(a,h)anthracene	26.31	278	316269	6.124	ng	# 95
39) Benzo(g,h,i)perylene	27.06	276	346263	5.425	ng	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE122419\
Data File : BE100971.D
Acq On : 24 Dec 2019 12:55
Operator : JU
Sample : SSTDIC5.0
Misc :
ALS Vial : 8 Sample Multiplier: 1


Instrument :
BNA_E
ClientSampleId :
SSTDICC5.0

Manual Integrations
APPROVED

mohammad
12/26/2019 4:02:29 PM

Quant Time: Dec 24 14:34:36 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE122419.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Dec 24 11:32:12 2019
Response via : Initial Calibration





#1


1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.75 min Scan# 424
Delta R.T. 0.00 min
Lab File: BE100971.D
Acq: 24 Dec 2019 12:55

Instrument :
BNA_E
ClientSampleId :
SSTDICC5.0

Tot Ion	Ratio	Lower	Upper
152	100		
150	164.0	132.6	199.0
115	66.7	64.1	96.1

Manual Integrations
APPROVED

mohammad
12/26/2019 4:02:29 PM






Abundance

Ion 152.00 (151.70 to 152.70): E

Ion 150.00 (149.70 to 150.70): E


Ion 115.00 (114.70 to 115.70): E



#3

n-Nitrosodimethylamine
Concen: 5.854 ng
RT: 3.61 min Scan# 65
Delta R.T. -0.06 min
Lab File: BE100971.D
Acq: 24 Dec 2019 12:55

Tgt Ion	Ratio	Lower	Upper
42	100		
74	123.4	116.6	174.8
44	7.4	0.0	0.0#






Abundance

Ion 42.00 (41.70 to 42.70): BE1

Ion 74.00 (73.70 to 74.70): BE1

Ion 44.00 (43.70 to 44.70): BE1





#4

2-Fluorophenol

Concen: 5.932 ng m

RT: 5.35 min Scan# 216

Delta R.T. -0.01 min

Lab File: BE100971.D

Acq: 24 Dec 2019 12:55

Instrument :

BNA_E





ClientSampleId :

SSTDICC5.0

Tot Ion:	112	Resp:	29281
Ion Ratio	Lower	Upper	
112	100		
64	62.3	59.0	88.6
63	133.1	121.4	182.0

Manual Integrations
APPROVED

mohammad
12/26/2019 4:02:29 PM



#5

Phenol-d6

Concen: 8.367 ng m




RT: 6.92 min Scan# 352


Delta R.T. -0.07 min

Lab File: BE100971.D

Acq: 24 Dec 2019 12:55

Tgt Ion:	99	Resp:	50391
Ion Ratio	Lower	Upper	
99	100		
42	0.0	0.0	0.0
71	21.3	0.0	0.0#





#6

bis(2-Chloroethyl)ether

Concen: 7.157 ng

RT: 7.17 min Scan# 374

Delta R.T. -0.02 min

Lab File: BE100971.D

Acq: 24 Dec 2019 12:55

Instrument :

BNA_E





ClientSampleId :

SSTDICC5.0

Tot Ion:	93	Resp:	50621
Ion	Ratio	Lower	Upper
93	100		
63	260.9	228.3	342.5
95	29.1	25.9	38.9

Manual Integrations
APPROVED

mohammad
12/26/2019 4:02:29 PM



#7

Naphthalene-d8

Concen: 0.400 ng




RT: 10.52 min Scan# 643


Delta R.T. -0.00 min

Lab File: BE100971.D

Acq: 24 Dec 2019 12:55

Tgt Ion:	136	Resp:	13673
Ion	Ratio	Lower	Upper
136	100		
137	11.4	9.0	13.6
54	29.6	34.4	51.6#
68	6.3	7.8	11.6#





#8

Nitrobenzene-d5

Concen: 7.010 ng

RT: 8.89 min Scan# 517

Delta R.T. -0.03 min

Lab File: BE100971.D

Acq: 24 Dec 2019 12:55

Instrument :

BNA_E





ClientSampleId :

SSTDICC5.0

Tot Ion	Ratio	Lower	Upper
82	100		
128	139.4	140.2	210.2#
54	179.6	183.8	275.6#

Manual Integrations
APPROVED

mohammad
12/26/2019 4:02:29 PM



#9

Naphthalene

Concen: 4.967 ng




RT: 10.56 min Scan# 646


Delta R.T. -0.01 min

Lab File: BE100971.D

Acq: 24 Dec 2019 12:55

Tgt Ion	Ratio	Lower	Upper
128	100		
129	2.7	2.6	3.8
127	3.3	2.8	4.2






#10
Hexachlorobutadiene
Concen: 4.604 ng
RT: 10.86 min Scan# 669
Delta R.T. -0.00 min
Lab File: BE100971.D
Acq: 24 Dec 2019 12:55

Instrument :
BNA_E
ClientSampleId :
SSTDICC5.0

Tot Ion	Ratio	Lower	Upper
225	100		
223	62.5	51.4	77.0
227	63.8	50.2	75.2

Manual Integrations
APPROVED

mohammad
12/26/2019 4:02:29 PM






Abundance

Ion 225.00 (224.70 to 225.70): E


Ion 223.00 (222.70 to 223.70): E

Ion 227.00 (226.70 to 227.70): E



#11
2-Methylnaphthalene-d10
Concen: 5.007 ng
RT: 12.12 min Scan# 757
Delta R.T. -0.00 min
Lab File: BE100971.D
Acq: 24 Dec 2019 12:55



Tgt Ion	Ratio	Lower	Upper
152	100		
151	19.1	15.1	22.7




Abundance

Ion 152.00 (151.70 to 152.70): E

Ion 151.00 (150.70 to 151.70): E






#12
2-Methylnaphthalene
Concen: 5.417 ng
RT: 12.19 min Scan# 762
Delta R.T. -0.00 min
Lab File: BE100971.D
Acq: 24 Dec 2019 12:55

Instrument :
BNA_E
ClientSampleId :
SSTDICC5.0

Tot Ion	Ratio	Lower	Upper
142	100		
141	87.3	70.8	106.2
115	36.8	32.0	48.0

Manual Integrations
APPROVED

mohammad
12/26/2019 4:02:29 PM






Abundance

Ion 142.00 (141.70 to 142.70): E


Ion 141.00 (140.70 to 141.70): E

Ion 115.00 (114.70 to 115.70): E



#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.37 min Scan# 913
Delta R.T. -0.02 min
Lab File: BE100971.D
Acq: 24 Dec 2019 12:55

Tgt Ion	Ratio	Lower	Upper
164	100		
162	109.3	81.6	122.4
160	51.4	37.7	56.5






Abundance

Ion 164.00 (163.70 to 164.70): E

Ion 162.00 (161.70 to 162.70): E

Ion 160.00 (159.70 to 160.70): E









#14
2,4,6-Tribromophenol
Concen: 7.224 ng
RT: 15.86 min Scan# 1018
Delta R.T. -0.01 min
Lab File: BE100971.D
Acq: 24 Dec 2019 12:55

Instrument :
BNA_E
ClientSampleId :
SSTDICC5.0

Tot Ion	Ratio	Lower	Upper
330	100		
332	95.7	72.4	108.6
141	45.5	25.4	38.2




Manual Integrations
APPROVED


mohammad
12/26/2019 4:02:29 PM



#15
2-Fluorobiphenyl
Concen: 5.127 ng
RT: 13.00 min Scan# 818
Delta R.T. -0.02 min
Lab File: BE100971.D
Acq: 24 Dec 2019 12:55

Tgt Ion	Ratio	Lower	Upper
172	100		
171	34.7	29.0	43.6
170	24.0	20.0	30.0






#16
Acenaphthylene
Concen: 5.614 ng
RT: 14.10 min Scan# 894
Delta R.T. -0.00 min
Lab File: BE100971.D
Acq: 24 Dec 2019 12:55

Instrument :
BNA_E
ClientSampleId :
SSTDICC5.0




Tot Ion	Ratio	Lower	Upper
152	100		
151	19.6	15.5	23.3
153	13.1	10.5	15.7

Manual Integrations
APPROVED

mohammad
12/26/2019 4:02:29 PM




Abundance Ion 152.00 (151.70 to 152.70): E
Ion 151.00 (150.70 to 151.70): E
Ion 153.00 (152.70 to 153.70): E






#17
Acenaphthene
Concen: 5.193 ng
RT: 14.44 min Scan# 918
Delta R.T. -0.00 min
Lab File: BE100971.D
Acq: 24 Dec 2019 12:55

Tgt Ion	Ratio	Lower	Upper
154	100		
153	113.5	91.7	137.5
152	55.6	44.5	66.7



Abundance Ion 154.00 (153.70 to 154.70): E
Ion 153.00 (152.70 to 153.70): E
Ion 152.00 (151.70 to 152.70): E









#18
Fluorene
 Concen: 5.409 ng
 RT: 15.41 min Scan# 984
 Delta R.T. -0.01 min
 Lab File: BE100971.D
 Acq: 24 Dec 2019 12:55

Instrument :
 BNA_E
 ClientSampleId :
 SSTDICC5.0

Tot Ion	Ratio	Resp	Lower	Upper
166	100	160284		
165	98.0	79.4	119.0	
167	13.4	10.6	15.8	




Manual Integrations
 APPROVED


mohammad
 12/26/2019 4:02:29 PM



#19
Phenanthrene-d10
 Concen: 0.400 ng
 RT: 17.11 min Scan# 1111
 Delta R.T. -0.01 min
 Lab File: BE100971.D
 Acq: 24 Dec 2019 12:55

Tgt Ion	Ratio	Resp	Lower	Upper
188	100	18463		
94	0.0	0.0	0.0	
80	13.5	9.7	14.5	









#20
4-Bromophenyl-phenylether
Concen: 5.730 ng
RT: 16.30 min Scan# 1051
Delta R.T. -0.01 min
Lab File: BE100971.D
Acq: 24 Dec 2019 12:55

Instrument :
BNA_E
ClientSampleId :
SSTDICC5.0

Tot Ion	Ratio	Lower	Upper
248	100		
250	91.4	76.0	114.0
141	82.3	61.0	91.6




Manual Integrations
APPROVED


mohammad
12/26/2019 4:02:29 PM



#21
Hexachlorobenzene
Concen: 5.164 ng
RT: 16.42 min Scan# 1060
Delta R.T. -0.00 min
Lab File: BE100971.D
Acq: 24 Dec 2019 12:55

Tgt Ion	Ratio	Lower	Upper
284	100		
142	42.1	33.3	49.9
249	34.0	27.2	40.8





#22

Pentachlorophenol

Concen: 9.342 ng m

RT: 16.77 min Scan# 1086

Delta R.T. -0.05 min

Lab File: BE100971.D

Acq: 24 Dec 2019 12:55

Instrument :

BNA_E





ClientSampleId :

SSTDICC5.0

Tot Ion:	266	Resp:	12933
Ion Ratio	Lower	Upper	
266	100		
264	60.5	62.3	93.5#
268	62.7	66.1	99.1#

Manual Integrations
APPROVED

mohammad
12/26/2019 4:02:29 PM



#23

Phenanthrene

Concen: 5.538 ng




RT: 17.15 min Scan# 1114


Delta R.T. -0.01 min

Lab File: BE100971.D

Acq: 24 Dec 2019 12:55

Tgt Ion:	178	Resp:	266505
Ion Ratio	Lower	Upper	
178	100		
176	19.3	15.7	23.5
179	15.2	12.6	18.8









#24
 Anthracene
 Concen: 8.152 ng/ml
 RT: 17.24 min Scan# 1121
 Delta R.T. -0.01 min
 Lab File: BE100971.D
 Acq: 24 Dec 2019 12:55

Instrument :
 BNA_E
 ClientSampleId :
 SSTDICC5.0

Tot Ion:	178	Resp:	263634
Ion Ratio	Lower	Upper	
178	100		
176	16.3	15.0	22.4
179	13.4	12.5	18.7




Manual Integrations
 APPROVED


mohammad
 12/26/2019 4:02:29 PM



#25
 Fluoranthene-d10
 Concen: 5.359 ng
 RT: 19.14 min Scan# 1370
 Delta R.T. -0.01 min
 Lab File: BE100971.D
 Acq: 24 Dec 2019 12:55

Tgt Ion:	212	Resp:	1238419
Ion Ratio	Lower	Upper	
212	100		
106	3.5	2.4	3.6
104	1.9	1.4	2.0









#26
Fluoranthene
Concen: 5.485 ng
RT: 19.17 min Scan# 1375
Delta R.T. -0.01 min
Lab File: BE100971.D
Acq: 24 Dec 2019 12:55

Instrument :
BNA_E
ClientSampleId :
SSTDICC5.0

Tot Ion	Ratio	Lower	Upper
202	100		
101	12.2	9.0	13.4
203	17.0	13.2	19.8




Manual Integrations
APPROVED


mohammad
12/26/2019 4:02:29 PM



#27
Chrysene-d12
Concen: 0.400 ng
RT: 21.28 min Scan# 1623
Delta R.T. -0.01 min
Lab File: BE100971.D
Acq: 24 Dec 2019 12:55

Tgt Ion	Ratio	Lower	Upper
240	100		
120	15.8	10.2	15.4#
236	28.0	22.6	34.0









#28
Pvrene
Concen: 4.742 ng
RT: 19.53 min Scan# 1438
Delta R.T. -0.01 min
Lab File: BE100971.D
Acq: 24 Dec 2019 12:55

Instrument :
BNA_E
ClientSampleId :
SSTDICC5.0

Tot Ion	Ratio	Resp	Lower	Upper
202	100	376039		
200	21.3		16.6	25.0
203	17.6		13.8	20.6




Manual Integrations
APPROVED


mohammad
12/26/2019 4:02:29 PM



#29
Terphenyl-d14
Concen: 5.151 ng
RT: 19.75 min Scan# 1476
Delta R.T. -0.01 min
Lab File: BE100971.D
Acq: 24 Dec 2019 12:55

Tgt Ion	Ratio	Resp	Lower	Upper
244	100	260639		
212	30.7		29.3	43.9
122	10.9		10.2	15.4









#30
Benzo(a)anthracene
Concen: 5.638 ng
RT: 21.27 min Scan# 1622
Delta R.T. -0.01 min
Lab File: BE100971.D
Acq: 24 Dec 2019 12:55

Instrument :
BNA_E
ClientSampleId :
SSTDICC5.0

Tot Ion	Ratio	Lower	Upper
228	100		
226	27.6	21.8	32.6
229	19.7	15.4	23.0




Manual Integrations
APPROVED


mohammad
12/26/2019 4:02:29 PM



#31
Chrysene
Concen: 5.984 ng m
RT: 21.32 min Scan# 1626
Delta R.T. -0.01 min
Lab File: BE100971.D
Acq: 24 Dec 2019 12:55

Tgt Ion	Ratio	Lower	Upper
228	100		
226	31.1	23.6	35.4
229	19.1	15.7	23.5





#32

Bis(2-ethylhexyl)phthalate

Concen: 4.119 ng

RT: 21.21 min Scan# 1617

Delta R.T. -0.01 min

Lab File: BE100971.D

Acq: 24 Dec 2019 12:55

Instrument :

BNA_E


ClientSampleId :

SSTDICC5.0




Tot Ion:	149	Resp:	753590
Ion Ratio	Lower	Upper	
149	100		
167	6.5	5.1	7.7
279	0.8	0.6	1.0

Manual Integrations
APPROVED

mohammad
12/26/2019 4:02:29 PM



Abundance Ion 149.00 (148.70 to 149.70): E
Ion 167.00 (166.70 to 167.70): E
Ion 279.00 (278.70 to 279.70): E



#33

Indeno(1,2,3-cd)pyrene

Concen: 5.154 ng

RT: 26.28 min Scan# 2873




Delta R.T. -0.01 min


Lab File: BE100971.D

Acq: 24 Dec 2019 12:55

Tgt Ion:	276	Resp:	395181
Ion Ratio	Lower	Upper	
276	100		
138	24.9	14.8	22.2#
277	24.9	18.0	27.0

Abundance Ion 276.00 (275.70 to 276.70): E
Ion 138.00 (137.70 to 138.70): E
Ion 277.00 (276.70 to 277.70): E









#34
Pervlene-d12
Concen: 0.400 ng
RT: 23.72 min Scan# 2159
Delta R.T. -0.00 min
Lab File: BE100971.D
Acq: 24 Dec 2019 12:55

Instrument :
BNA_E
ClientSampleId :
SSTDICC5.0

Tot Ion	Ratio	Lower	Upper
264	100		
260	24.3	19.1	28.7
265	27.5	26.6	40.0




Manual Integrations
APPROVED


mohammad
12/26/2019 4:02:29 PM



#36
Benzo(k)fluoranthene
Concen: 6.424 ng m
RT: 23.02 min Scan# 1963
Delta R.T. -0.01 min
Lab File: BE100971.D
Acq: 24 Dec 2019 12:55

Tgt Ion	Ratio	Lower	Upper
252	100		
253	21.9	18.9	28.3
125	10.9	11.2	16.8









#37
Benzo(a)pyrene
Concen: 5.405 ng
RT: 23.61 min Scan# 2128
Delta R.T. -0.01 min
Lab File: BE100971.D
Acq: 24 Dec 2019 12:55

Instrument :
BNA_E
ClientSampleId :
SSTDICC5.0

Tot Ion	Ratio	Lower	Upper
252	100		
253	21.9	20.1	30.1
125	11.5	12.6	18.8#




Manual Integrations
APPROVED




mohammad
12/26/2019 4:02:29 PM



#38
Dibenzo(a,h)anthracene
Concen: 6.124 ng
RT: 26.31 min Scan# 2879
Delta R.T. -0.01 min
Lab File: BE100971.D
Acq: 24 Dec 2019 12:55

Tgt Ion	Ratio	Lower	Upper
278	100		
139	16.2	16.4	24.6#
279	23.5	19.3	28.9





#39

Benzo(a,h,i)perylene

Concen: 5.425 ng

RT: 27.06 min Scan# 3091

Delta R.T. -0.01 min

Lab File: BE100971.D

Acq: 24 Dec 2019 12:55

Instrument :

BNA_E

ClientSampleId :

SSTDICC5.0

Tot Ion:	276	Resp:	346263
Ion Ratio	Lower	Upper	
276	100		
277	23.8	20.0	30.0
138	21.9	18.6	27.8

Manual Integrations
APPROVED

mohammad
12/26/2019 4:02:29 PM

