

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE122617\
 Data File : BE095047.D
 Acq On : 26 Dec 2017 9:39
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Dec 26 10:52:14 2017
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE121417.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Dec 14 18:42:53 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	86	-0.01
2	1,4-Dioxane	0.394	0.377	4.3	77	0.00
3	n-Nitrosodimethylamine	0.466	0.472	-1.3	86	0.00
4 S	2-Fluorophenol	0.621	0.605	2.6	87	-0.01
5 S	Phenol-d6	0.923	0.876	5.1	84	0.00
6	bis(2-Chloroethyl)ether	0.901	0.927	-2.9	89	-0.01
7 I	Naphthalene-d8	1.000	1.000	0.0	93	-0.02
8 S	Nitrobenzene-d5	0.328	0.311	5.2	92	0.00
9	Naphthalene	4.697	4.474	4.7	89	-0.02
10	Hexachlorobutadiene	0.214	0.206	3.7	89	-0.02
11 SURR	2-Methylnaphthalene-d10	0.655	0.630	3.8	90	-0.02
12	2-Methylnaphthalene	1.172	1.019	13.1	82	-0.02
13 I	Acenaphthene-d10	1.000	1.000	0.0	89	-0.02
14 S	2,4,6-Tribromophenol	0.093	0.075	19.4	77	-0.03
15 S	2-Fluorobiphenyl	1.772	1.746	1.5	87	-0.02
16	Acenaphthylene	2.138	1.982	7.3	87	-0.03
17	Acenaphthene	1.418	1.365	3.7	87	-0.02
18	Fluorene	1.762	1.691	4.0	88	-0.03
19 I	Phenanthrene-d10	1.000	1.000	0.0	90	-0.03
20	4-Bromophenyl-phenylether	0.234	0.219	6.4	87	-0.03
21	Hexachlorobenzene	0.239	0.230	3.8	88	-0.03
22	Pentachlorophenol	0.046	0.040	13.0	81	-0.02
23	Phenanthrene	1.246	1.210	2.9	90	-0.03
24	Anthracene	1.279	1.195	6.6	89	-0.03
25 SURR	Fluoranthene-d10	5.189	5.025	3.2	91	-0.03
26	Fluoranthene	1.538	1.490	3.1	90	-0.03
27 I	Chrysene-d12	1.000	1.000	0.0	89	-0.03
28	Pyrene	1.355	1.657	-22.3	105	-0.03
29 S	Terphenyl-d14	0.753	0.752	0.1	88	-0.03
30	Benzo(a)anthracene	1.169	1.192	-2.0	92	-0.03
31	Chrysene	1.316	1.374	-4.4	93	-0.03
32	Bis(2-ethylhexyl)phthalate	1.489	1.326	10.9	89	-0.05
33	Indeno(1,2,3-cd)pyrene	1.071	1.133	-5.8	92	-0.07
34 I	Perylene-d12	1.000	1.000	0.0	98	-0.05
35	Benzo(b)fluoranthene	1.469	1.427	2.9	95	-0.04
36	Benzo(k)fluoranthene	1.488	1.408	5.4	97	-0.05
37 C	Benzo(a)pyrene	1.303	1.287	1.2	97	-0.05
38	Dibenzo(a,h)anthracene	1.209	1.036	14.3	85	-0.07
39	Benzo(g,h,i)perylene	1.203	1.127	6.3	94	-0.08

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE122617\
Data File : BE095047.D
Acq On : 26 Dec 2017 9:39
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Dec 26 10:52:14 2017
Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE121417.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Dec 14 18:42:53 2017
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0	CCC's out = 0		