

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE122617\
 Data File : BE095063.D
 Acq On : 26 Dec 2017 19:22
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Dec 27 01:18:56 2017
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE121417.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Dec 14 18:42:53 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	95	-0.01
2	1,4-Dioxane	0.394	0.369	6.3	84	0.00
3	n-Nitrosodimethylamine	0.466	0.463	0.6	94	0.00
4 S	2-Fluorophenol	0.621	0.705	-13.5	113	-0.01
5 S	Phenol-d6	0.923	1.030	-11.6	110	-0.01
6	bis(2-Chloroethyl)ether	0.901	0.942	-4.6	101	-0.01
7 I	Naphthalene-d8	1.000	1.000	0.0	105	-0.02
8 S	Nitrobenzene-d5	0.328	0.354	-7.9	119	-0.02
9	Naphthalene	4.697	4.610	1.9	104	-0.03
10	Hexachlorobutadiene	0.214	0.209	2.3	103	-0.02
11 SURR	2-Methylnaphthalene-d10	0.655	0.656	-0.2	107	-0.02
12	2-Methylnaphthalene	1.172	0.788	32.8#	72	-0.02
13 I	Acenaphthene-d10	1.000	1.000	0.0	106	-0.03
14 S	2,4,6-Tribromophenol	0.093	0.122	-31.2#	149	-0.03
15 S	2-Fluorobiphenyl	1.772	1.757	0.8	103	-0.02
16	Acenaphthylene	2.138	2.140	-0.1	111	-0.03
17	Acenaphthene	1.418	1.381	2.6	104	-0.02
18	Fluorene	1.762	1.750	0.7	107	-0.03
19 I	Phenanthrene-d10	1.000	1.000	0.0	105	-0.03
20	4-Bromophenyl-phenylether	0.234	0.225	3.8	105	-0.03
21	Hexachlorobenzene	0.239	0.228	4.6	102	-0.03
22	Pentachlorophenol	0.046	0.068	-47.8#	162#	-0.03
23	Phenanthrene	1.246	1.204	3.4	105	-0.03
24	Anthracene	1.279	1.081	15.5	94	-0.03
25 SURR	Fluoranthene-d10	5.189	4.983	4.0	106	-0.03
26	Fluoranthene	1.538	1.481	3.7	105	-0.03
27 I	Chrysene-d12	1.000	1.000	0.0	97	-0.03
28	Pyrene	1.355	1.660	-22.5	116	-0.03
29 S	Terphenyl-d14	0.753	0.803	-6.6	103	-0.03
30	Benzo(a)anthracene	1.169	1.309	-12.0	111	-0.03
31	Chrysene	1.316	1.205	8.4	89	-0.05
32	Bis(2-ethylhexyl)phthalate	1.489	2.145	-44.1#	157#	-0.05
33	Indeno(1,2,3-cd)pyrene	1.071	1.241	-15.9	111	-0.09
34 I	Perylene-d12	1.000	1.000	0.0	109	-0.06
35	Benzo(b)fluoranthene	1.469	1.391	5.3	103	-0.05
36	Benzo(k)fluoranthene	1.488	1.399	6.0	108	-0.06
37 C	Benzo(a)pyrene	1.303	1.303	0.0	110	-0.06
38	Dibenzo(a,h)anthracene	1.209	1.166	3.6	107	-0.09
39	Benzo(g,h,i)perylene	1.203	1.164	3.2	108	-0.10

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE122617\
Data File : BE095063.D
Acq On : 26 Dec 2017 19:22
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Dec 27 01:18:56 2017
Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE121417.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Dec 14 18:42:53 2017
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0	CCC's out = 0		