

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE122617\
 Data File : BE095063.D
 Acq On : 26 Dec 2017 19:22
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Dec 27 01:18:56 2017
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE121417.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Dec 14 18:42:53 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	95	-0.01
2	1,4-Dioxane	0.400	0.374	6.5	84	0.00
3	n-Nitrosodimethylamine	0.400	0.398	0.5	94	0.00
4 S	2-Fluorophenol	0.400	0.454	-13.5	113	-0.01
5 S	Phenol-d6	0.400	0.447	-11.7	110	-0.01
6	bis(2-Chloroethyl)ether	0.400	0.418	-4.5	101	-0.01
7 I	Naphthalene-d8	0.400	0.400	0.0	105	-0.02
8 S	Nitrobenzene-d5	0.400	0.431	-7.7	119	-0.02
9	Naphthalene	0.400	0.393	1.8	104	-0.03
10	Hexachlorobutadiene	0.400	0.392	2.0	103	-0.02
11 SURR	2-Methylnaphthalene-d10	0.400	0.401	-0.3	107	-0.02
12	2-Methylnaphthalene	0.400	0.269	32.8#	72	-0.02
13 I	Acenaphthene-d10	0.400	0.400	0.0	106	-0.03
14 S	2,4,6-Tribromophenol	0.400	0.526	-31.5#	149	-0.03
15 S	2-Fluorobiphenyl	0.400	0.396	1.0	103	-0.02
16	Acenaphthylene	0.400	0.400	0.0	111	-0.03
17	Acenaphthene	0.400	0.389	2.8	104	-0.02
18	Fluorene	0.400	0.397	0.8	107	-0.03
19 I	Phenanthrene-d10	0.400	0.400	0.0	105	-0.03
20	4-Bromophenyl-phenylether	0.400	0.385	3.8	105	-0.03
21	Hexachlorobenzene	0.400	0.381	4.8	102	-0.03
22	Pentachlorophenol	0.400	0.538	-34.5#	162	-0.03
23	Phenanthrene	0.400	0.386	3.5	105	-0.03
24	Anthracene	0.400	0.338	15.5	94	-0.03
25 SURR	Fluoranthene-d10	0.400	0.384	4.0	106	-0.03
26	Fluoranthene	0.400	0.385	3.8	105	-0.03
27 I	Chrysene-d12	0.400	0.400	0.0	97	-0.03
28	Pyrene	0.400	0.490	-22.5	116	-0.03
29 S	Terphenyl-d14	0.400	0.426	-6.5	103	-0.03
30	Benzo(a)anthracene	0.400	0.448	-12.0	111	-0.03
31	Chrysene	0.400	0.366	8.5	89	-0.05
32	Bis(2-ethylhexyl)phthalate	0.400	0.576	-44.0#	157	-0.05
33	Indeno(1,2,3-cd)pyrene	0.400	0.463	-15.8	111	-0.09
34 I	Perylene-d12	0.400	0.400	0.0	109	-0.06
35	Benzo(b)fluoranthene	0.400	0.379	5.3	103	-0.05
36	Benzo(k)fluoranthene	0.400	0.376	6.0	108	-0.06
37 C	Benzo(a)pyrene	0.400	0.400	0.0	110	-0.06
38	Dibenzo(a,h)anthracene	0.400	0.386	3.5	107	-0.09
39	Benzo(g,h,i)perylene	0.400	0.387	3.3	108	-0.10

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE122617\
Data File : BE095063.D
Acq On : 26 Dec 2017 19:22
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Dec 27 01:18:56 2017
Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE121417.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Dec 14 18:42:53 2017
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0				