

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE122618\
 Data File : BE098559.D
 Acq On : 26 Dec 2018 10:07
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Dec 26 10:40:59 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE121018.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Dec 10 14:59:55 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	51	-0.01
2	1,4-Dioxane	0.400	0.419	-4.7	54	-0.01
3	n-Nitrosodimethylamine	0.400	0.348	13.0	48	0.00
4 S	2-Fluorophenol	0.400	0.416	-4.0	55	0.00
5 S	Phenol-d6	0.400	0.441	-10.2	56	-0.01
6	bis(2-Chloroethyl)ether	0.400	0.362	9.5	48	-0.01
7 I	Naphthalene-d8	0.400	0.400	0.0	51	-0.01
8 S	Nitrobenzene-d5	0.400	0.387	3.3	54	0.00
9	Naphthalene	0.400	0.417	-4.2	54	-0.01
10	Hexachlorobutadiene	0.400	0.460	-15.0	60	-0.01
11 SURR	2-Methylnaphthalene-d10	0.400	0.423	-5.7	55	-0.01
12	2-Methylnaphthalene	0.400	0.415	-3.7	54	-0.01
13 I	Acenaphthene-d10	0.400	0.400	0.0	55	-0.01
14 S	2,4,6-Tribromophenol	0.400	0.457	-14.2	66	-0.01
15 S	2-Fluorobiphenyl	0.400	0.389	2.8	59	-0.01
16	Acenaphthylene	0.400	0.367	8.3	52	-0.01
17	Acenaphthene	0.400	0.388	3.0	54	-0.01
18	Fluorene	0.400	0.396	1.0	55	-0.01
19 I	Phenanthrene-d10	0.400	0.400	0.0	51	-0.02
20	4-Bromophenyl-phenylether	0.400	0.385	3.8	51	-0.01
21	Hexachlorobenzene	0.400	0.438	-9.5	57	-0.01
22	Pentachlorophenol	0.400	0.531	-32.8#	73	-0.01
23	Phenanthrene	0.400	0.408	-2.0	53	-0.01
24	Anthracene	0.400	0.366	8.5	47	-0.01
25 SURR	Fluoranthene-d10	0.400	0.443	-10.7	54	-0.02
26	Fluoranthene	0.400	0.436	-9.0	53	-0.02
27 I	Chrysene-d12	0.400	0.400	0.0	59	-0.01
28	Pvrene	0.400	0.372	7.0	53	-0.02
29 S	Terphenyl-d14	0.400	0.367	8.3	54	-0.02
30	Benzo(a)anthracene	0.400	0.385	3.8	58	-0.03
31	Chrysene	0.400	0.395	1.3	59	-0.01
32	Bis(2-ethylhexyl)phthalate	0.400	0.338	15.5	50	-0.01
33	Indeno(1,2,3-cd)pyrene	0.400	0.423	-5.7	66	-0.04
34 I	Perylene-d12	0.400	0.400	0.0	62	-0.03
35	Benzo(b)fluoranthene	0.400	0.402	-0.5	65	-0.03
36	Benzo(k)fluoranthene	0.400	0.391	2.3	64	-0.03
37 C	Benzo(a)pyrene	0.400	0.403	-0.8	64	-0.03
38	Dibenzo(a,h)anthracene	0.400	0.392	2.0	63	-0.04
39	Benzo(a,h,i)perylene	0.400	0.406	-1.5	65	-0.04

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE122618\
Data File : BE098559.D
Acq On : 26 Dec 2018 10:07
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Dec 26 10:40:59 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE121018.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Dec 10 14:59:55 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0				