

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE122619\
 Data File : BE100980.D
 Acq On : 26 Dec 2019 20:40
 Operator : JU
 Sample : K6392-02
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 MW-2

Manual Integrations
 APPROVED

mohammad
 12/27/2019 4:21:07 PM

Quant Time: Dec 27 10:56:23 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE122419.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Dec 24 14:28:42 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.75	152	2473	0.40	ng	0.00
7) Naphthalene-d8	10.52	136	11890	0.40	ng	0.00
13) Acenaphthene-d10	14.37	164	7246	0.40	ng	-0.01
19) Phenanthrene-d10	17.11	188	17742	0.40	ng	-0.01
27) Chrysene-d12	21.29	240	20269	0.40	ng	0.00
34) Perylene-d12	23.71	264	23728	0.40	ng	-0.01
System Monitoring Compounds						
4) 2-Fluorophenol	5.35	112	1030	0.25	ng	0.00
5) Phenol-d6	6.93	99	811m	0.29	ng	-0.06
8) Nitrobenzene-d5	8.90	82	2365	0.48	ng	-0.01
11) 2-Methylnaphthalene-d10	12.12	152	7009	0.40	ng	0.00
14) 2,4,6-Tribromophenol	15.86	330	620	0.47	ng	-0.01
15) 2-Fluorobiphenyl	13.00	172	12880	0.50	ng	-0.01
25) Fluoranthene-d10	19.14	212	105408	0.49	ng	0.00
29) Terphenyl-d14	19.74	244	23077	0.50	ng	-0.02
Target Compounds						
2) 1,4-Dioxane	3.31	88	51066	10.664	ng	# 74
9) Naphthalene	10.57	128	2819	0.022	ng	# 88
23) Phenanthrene	17.15	178	1885	0.041	ng	97
26) Fluoranthene	19.17	202	1269	0.020	ng	# 80
32) Bis(2-ethylhexyl)phthalate	21.21	149	89562	0.978	ng	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE122619\
Data File : BE100980.D
Acq On : 26 Dec 2019 20:40
Operator : JU
Sample : K6392-02
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
MW-2

Manual Integrations
APPROVED

mohammad
12/27/2019 4:21:07 PM

Quant Time: Dec 27 10:56:23 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE122419.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Dec 24 14:28:42 2019
Response via : Initial Calibration

