

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE122619\
 Data File : BE100981.D
 Acq On : 26 Dec 2019 21:16
 Operator : JU
 Sample : K6392-03
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 MW-3

Manual Integrations
 APPROVED

mohammad
 12/27/2019 4:21:10 PM

Quant Time: Dec 27 10:57:12 2019

Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE122419.M

Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION

QLast Update : Tue Dec 24 14:28:42 2019

Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.75	152	2679	0.40	ng	0.00
7) Naphthalene-d8	10.52	136	12484	0.40	ng	0.00
13) Acenaphthene-d10	14.37	164	7530	0.40	ng	-0.02
19) Phenanthrene-d10	17.11	188	18446	0.40	ng	-0.01
27) Chrysene-d12	21.28	240	20233	0.40	ng	-0.01
34) Perylene-d12	23.71	264	23731	0.40	ng	-0.01
System Monitoring Compounds						
4) 2-Fluorophenol	5.35	112	1106	0.24	ng	-0.01
5) Phenol-d6	6.93	99	1056m	0.31	ng	-0.06
8) Nitrobenzene-d5	8.89	82	2517	0.48	ng	-0.03
11) 2-Methylnaphthalene-d10	12.12	152	7407	0.40	ng	0.00
14) 2,4,6-Tribromophenol	15.86	330	687	0.49	ng	-0.01
15) 2-Fluorobiphenyl	13.00	172	12007	0.45	ng	-0.02
25) Fluoranthene-d10	19.14	212	110533	0.49	ng	0.00
29) Terphenyl-d14	19.74	244	22669	0.49	ng	-0.02
Target Compounds						
2) 1,4-Dioxane	3.31	88	53381	10.280	ng	# 74
9) Naphthalene	10.58	128	2879	0.022	ng	# 86
23) Phenanthrene	17.15	178	2040	0.043	ng	96
26) Fluoranthene	19.17	202	1486	0.023	ng	# 92
32) Bis(2-ethylhexyl)phthalate	21.21	149	62024	0.697	ng	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE122619\
Data File : BE100981.D
Acq On : 26 Dec 2019 21:16
Operator : JU
Sample : K6392-03
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
MW-3

Manual Integrations
APPROVED

mohammad
12/27/2019 4:21:10 PM

Quant Time: Dec 27 10:57:12 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE122419.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Dec 24 14:28:42 2019
Response via : Initial Calibration

