

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE122619\
 Data File : BE100974.D
 Acq On : 26 Dec 2019 16:17
 Operator : JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Dec 27 03:06:11 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE122419.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Dec 24 14:28:42 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	145	0.00
2	1,4-Dioxane	0.400	0.439	-9.7	158	0.00
3	n-Nitrosodimethylamine	0.400	0.547	-36.8#	211	0.00
4 S	2-Fluorophenol	0.400	0.371	7.3	149	0.00
5 S	Phenol-d6	0.400	0.371	7.3	111	-0.04
6	bis(2-Chloroethyl)ether	0.400	0.300	25.0#	129	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	148	0.00
8 S	Nitrobenzene-d5	0.400	0.396	1.0	123	-0.01
9	Naphthalene	0.400	0.426	-6.5	157	0.00
10	Hexachlorobutadiene	0.400	0.412	-3.0	149	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.419	-4.7	151	0.00
12	2-Methylnaphthalene	0.400	0.391	2.3	142	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	137	0.00
14 S	2,4,6-Tribromophenol	0.400	0.433	-8.2	141	0.00
15 S	2-Fluorobiphenyl	0.400	0.443	-10.7	150	0.00
16	Acenaphthylene	0.400	0.342	14.5	126	0.00
17	Acenaphthene	0.400	0.441	-10.2	150	0.00
18	Fluorene	0.400	0.458	-14.5	162	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	137	-0.01
20	4-Bromophenyl-phenylether	0.400	0.393	1.8	153	0.00
21	Hexachlorobenzene	0.400	0.389	2.8	138	0.00
22	Pentachlorophenol	0.400	0.326	18.5	96	-0.04
23	Phenanthrene	0.400	0.422	-5.5	162	-0.01
24	Anthracene	0.400	0.379	5.3	143	0.00
25 SURR	Fluoranthene-d10	0.400	0.379	5.3	140	0.00
26	Fluoranthene	0.400	0.385	3.8	146	0.00
27 I	Chrysene-d12	0.400	0.400	0.0	132	0.00
28	Pvrene	0.400	0.437	-9.2	141	0.00
29 S	Terphenyl-d14	0.400	0.442	-10.5	148	-0.01
30	Benzo(a)anthracene	0.400	0.425	-6.2	161	0.00
31	Chrysene	0.400	0.465	-16.3	138	0.00
32	Bis(2-ethylhexyl)phthalate	0.400	0.377	5.8	127	-0.01
33	Indeno(1,2,3-cd)pyrene	0.400	0.469	-17.2	162	-0.01
34 I	Perylene-d12	0.400	0.400	0.0	141	0.00
35	Benzo(b)fluoranthene	0.400	0.448	-12.0	161	0.00
36	Benzo(k)fluoranthene	0.400	0.425	-6.2	149	0.00
37 C	Benzo(a)pyrene	0.400	0.409	-2.2	147	0.00
38	Dibenzo(a,h)anthracene	0.400	0.412	-3.0	175	-0.01
39	Benzo(a,h,i)perylene	0.400	0.455	-13.7	160	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE122619\
Data File : BE100974.D
Acq On : 26 Dec 2019 16:17
Operator : JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Dec 27 03:06:11 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE122419.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Dec 24 14:28:42 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0				