

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE122619\
 Data File : BE100974.D
 Acq On : 26 Dec 2019 16:17
 Operator : JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Dec 27 03:06:11 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE122419.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Dec 24 14:28:42 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	145	0.00
2	1,4-Dioxane	1.413	1.365	3.4	158#	0.00
3	n-Nitrosodimethylamine	0.330	0.300	9.1	211#	0.00
4 S	2-Fluorophenol	0.676	0.628	7.1	149	0.00
5 S	Phenol-d6	0.950	0.615	35.3#	111	-0.04
6	bis(2-Chloroethyl)ether	1.446	1.086	24.9	129	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	148	0.00
8 S	Nitrobenzene-d5	0.192	0.140	27.1#	123	-0.01
9	Naphthalene	4.253	4.531	-6.5	157#	0.00
10	Hexachlorobutadiene	0.188	0.193	-2.7	149	0.00
11 SURR	2-Methylnaphthalene-d10	0.588	0.616	-4.8	151#	0.00
12	2-Methylnaphthalene	0.644	0.630	2.2	142	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	137	0.00
14 S	2,4,6-Tribromophenol	0.088	0.073	17.0	141	0.00
15 S	2-Fluorobiphenyl	1.421	1.576	-10.9	150	0.00
16	Acenaphthylene	1.495	1.278	14.5	126	0.00
17	Acenaphthene	1.086	1.197	-10.2	150	0.00
18	Fluorene	1.354	1.550	-14.5	162#	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	137	-0.01
20	4-Bromophenyl-phenylether	0.176	0.173	1.7	153#	0.00
21	Hexachlorobenzene	0.207	0.202	2.4	138	0.00
22	Pentachlorophenol	0.031	0.014	54.8#	96	-0.04
23	Phenanthrene	1.031	1.087	-5.4	162#	-0.01
24	Anthracene	1.043	0.988	5.3	143	0.00
25 SURR	Fluoranthene-d10	4.854	4.603	5.2	140	0.00
26	Fluoranthene	1.406	1.355	3.6	146	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	132	0.00
28	Pvrene	1.399	1.530	-9.4	141	0.00
29 S	Terphenyl-d14	0.919	1.016	-10.6	148	-0.01
30	Benzo(a)anthracene	0.979	1.040	-6.2	161#	0.00
31	Chrysene	1.823	2.120	-16.3	138	0.00
32	Bis(2-ethylhexyl)phthalate	2.154	1.636	24.0	127	-0.01
33	Indeno(1,2,3-cd)pyrene	1.372	1.610	-17.3	162#	-0.01
34 I	Perylene-d12	1.000	1.000	0.0	141	0.00
35	Benzo(b)fluoranthene	0.842	0.943	-12.0	161#	0.00
36	Benzo(k)fluoranthene	1.597	1.698	-6.3	149	0.00
37 C	Benzo(a)pyrene	0.983	1.005	-2.2	147	0.00
38	Dibenzo(a,h)anthracene	0.891	0.916	-2.8	175#	-0.01
39	Benzo(a,h,i)perylene	1.049	1.193	-13.7	160#	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE122619\
Data File : BE100974.D
Acq On : 26 Dec 2019 16:17
Operator : JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Dec 27 03:06:11 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE122419.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Dec 24 14:28:42 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0	CCC's out = 0		