

Data Path : Z:\HPCHEM1\BNA E\DATA\BE122716\
 Data File : BE092617.D
 Acq On : 27 Dec 2016 10:04
 Operator : UM/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Dec 28 01:29:27 2016
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE121916.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Dec 19 16:16:32 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	134	0.00
2	1,4-Dioxane	0.733	0.786	-7.2	146	0.00
3	n-Nitrosodimethylamine	0.613	0.618	-0.8	176#	-0.01
4 S	2-Fluorophenol	0.761	0.700	8.0	144	0.00
5 S	Phenol-d6	1.090	1.046	4.0	164#	-0.02
6	bis(2-Chloroethyl)ether	1.132	1.019	10.0	151#	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	130	0.00
8 S	Nitrobenzene-d5	0.280	0.262	6.4	153#	0.00
9	Naphthalene	1.026	1.023	0.3	140	0.00
10	Hexachlorobutadiene	0.156	0.157	-0.6	140	0.00
11	2-Methylnaphthalene	0.653	0.625	4.3	139	-0.01
12 I	Acenaphthene-d10	1.000	1.000	0.0	130	0.00
13 S	2,4,6-Tribromophenol	0.113	0.096	15.0	139	0.00
14 S	2-Fluorobiphenyl	1.196	1.162	2.8	138	0.00
15	Acenaphthylene	6.899	6.424	6.9	135	0.00
16	Acenaphthene	1.137	1.083	4.7	133	0.00
17	Fluorene	1.397	1.338	4.2	137	-0.01
18 I	Phenanthrene-d10	1.000	1.000	0.0	128	0.00
19	Hexachlorobenzene	0.214	0.224	-4.7	145	0.00
20	Pentachlorophenol	0.036	0.028	22.2	143	0.00
21	Phenanthrene	1.261	1.165	7.6	134	0.00
22	Anthracene	1.152	1.055	8.4	138	0.00
23	Fluoranthene	5.447	5.048	7.3	135	0.00
24 I	Chrysene-d12	1.000	1.000	0.0	129	0.00
25	Pyrene	4.403	4.016	8.8	134	0.00
26 S	Terphenyl-d14	0.613	0.567	7.5	134	0.00
27	Benzo(a)anthracene	4.490	4.351	3.1	146	0.00
28	Chrysene	5.204	5.138	1.3	138	0.00
29	Bis(2-ethylhexyl)phthalate	0.486	0.345	29.0#	129	0.00
30	Indeno(1,2,3-cd)pyrene	4.702	4.287	8.8	137	0.00
31 I	Perylene-d12	1.000	1.000	0.0	126	0.01
32	Benzo(b)fluoranthene	1.150	1.076	6.4	135	0.00
33	Benzo(k)fluoranthene	1.282	1.271	0.9	144	0.00
34 C	Benzo(a)pyrene	1.110	1.042	6.1	141	0.00
35	Dibenzo(a,h)anthracene	1.054	0.953	9.6	136	0.02
36	Benzo(g,h,i)perylene	4.515	4.262	5.6	135	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0