

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE122719\
 Data File : BE100987.D
 Acq On : 27 Dec 2019 9:58
 Operator : JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Dec 27 23:56:57 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE122419.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Dec 27 17:39:05 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	103	0.00
2	1,4-Dioxane	1.413	1.235	12.6	102	0.00
3	n-Nitrosodimethylamine	0.330	0.401	-21.5	201#	-0.02
4 S	2-Fluorophenol	0.676	0.711	-5.2	120	0.00
5 S	Phenol-d6	0.950	0.794	16.4	102	-0.04
6	bis(2-Chloroethyl)ether	1.446	1.251	13.5	106	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	108	0.00
8 S	Nitrobenzene-d5	0.192	0.181	5.7	115	-0.01
9	Naphthalene	4.253	4.379	-3.0	110	0.00
10	Hexachlorobutadiene	0.188	0.195	-3.7	109	0.00
11 SURR	2-Methylnaphthalene-d10	0.588	0.637	-8.3	113	0.00
12	2-Methylnaphthalene	0.644	0.600	6.8	98	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	105	-0.01
14 S	2,4,6-Tribromophenol	0.088	0.079	10.2	118	-0.01
15 S	2-Fluorobiphenyl	1.421	1.544	-8.7	112	0.00
16	Acenaphthylene	1.495	1.356	9.3	102	0.00
17	Acenaphthene	1.086	1.131	-4.1	108	0.00
18	Fluorene	1.354	1.407	-3.9	112	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	104	-0.01
20	4-Bromophenyl-phenylether	0.176	0.168	4.5	113	0.00
21	Hexachlorobenzene	0.207	0.202	2.4	105	0.00
22	Pentachlorophenol	0.031	0.018	41.9#	92	-0.01
23	Phenanthrene	1.031	1.025	0.6	116	0.00
24	Anthracene	1.043	0.983	5.8	108	0.00
25 SURR	Fluoranthene-d10	4.854	4.845	0.2	112	0.00
26	Fluoranthene	1.406	1.363	3.1	112	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	101	0.00
28	Pvrene	1.399	1.574	-12.5	111	0.00
29 S	Terphenyl-d14	0.919	1.040	-13.2	116	0.00
30	Benzo(a)anthracene	0.979	1.032	-5.4	122	0.00
31	Chrysene	1.823	1.981	-8.7	99	0.00
32	Bis(2-ethylhexyl)phthalate	2.154	1.706	20.8	101	-0.01
33	Indeno(1,2,3-cd)pyrene	1.372	1.420	-3.5	109	-0.01
34 I	Perylene-d12	1.000	1.000	0.0	92	0.00
35	Benzo(b)fluoranthene	0.842	1.013	-20.3	113	0.00
36	Benzo(k)fluoranthene	1.597	1.685	-5.5	96	0.00
37 C	Benzo(a)pyrene	0.983	1.053	-7.1	100	0.00
38	Dibenzo(a,h)anthracene	0.891	0.949	-6.5	118	-0.01
39	Benzo(a,h,i)perylene	1.049	1.166	-11.2	102	-0.01

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE122719\
Data File : BE100987.D
Acq On : 27 Dec 2019 9:58
Operator : JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Dec 27 23:56:57 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE122419.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Dec 27 17:39:05 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			