

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE122719\
 Data File : BE100999.D
 Acq On : 27 Dec 2019 18:44
 Operator : JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Dec 27 23:40:37 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE122419.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Dec 27 17:39:05 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	109	0.00
2	1,4-Dioxane	1.413	0.942	33.3#	82	0.00
3	n-Nitrosodimethylamine	0.330	0.345	-4.5	184#	-0.01
4 S	2-Fluorophenol	0.676	0.685	-1.3	123	0.00
5 S	Phenol-d6	0.950	0.902	5.1	123	-0.03
6	bis(2-Chloroethyl)ether	1.446	1.037	28.3#	93	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	111	0.00
8 S	Nitrobenzene-d5	0.192	0.176	8.3	116	0.00
9	Naphthalene	4.253	4.390	-3.2	114	0.00
10	Hexachlorobutadiene	0.188	0.198	-5.3	115	0.00
11 SURR	2-Methylnaphthalene-d10	0.588	0.640	-8.8	118	0.00
12	2-Methylnaphthalene	0.644	0.603	6.4	102	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	107	-0.01
14 S	2,4,6-Tribromophenol	0.088	0.070	20.5	107	-0.01
15 S	2-Fluorobiphenyl	1.421	1.535	-8.0	114	0.00
16	Acenaphthylene	1.495	1.306	12.6	101	0.00
17	Acenaphthene	1.086	1.121	-3.2	110	0.00
18	Fluorene	1.354	1.344	0.7	110	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	102	-0.01
20	4-Bromophenyl-phenylether	0.176	0.165	6.2	109	0.00
21	Hexachlorobenzene	0.207	0.206	0.5	105	0.00
22	Pentachlorophenol	0.031	0.014	54.8#	68	-0.04
23	Phenanthrene	1.031	1.016	1.5	113	0.00
24	Anthracene	1.043	0.931	10.7	101	0.00
25 SURR	Fluoranthene-d10	4.854	4.495	7.4	102	0.00
26	Fluoranthene	1.406	1.290	8.3	104	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	92	0.00
28	Pvrene	1.399	1.571	-12.3	102	0.00
29 S	Terphenyl-d14	0.919	1.063	-15.7	108	-0.01
30	Benzo(a)anthracene	0.979	1.005	-2.7	109	0.00
31	Chrysene	1.823	1.917	-5.2	87	0.00
32	Bis(2-ethylhexyl)phthalate	2.154	1.639	23.9	89	-0.01
33	Indeno(1,2,3-cd)pyrene	1.372	1.427	-4.0	100	-0.01
34 I	Perylene-d12	1.000	1.000	0.0	88	0.00
35	Benzo(b)fluoranthene	0.842	0.946	-12.4	101	0.00
36	Benzo(k)fluoranthene	1.597	1.658	-3.8	90	0.00
37 C	Benzo(a)pyrene	0.983	0.964	1.9	88	-0.01
38	Dibenzo(a,h)anthracene	0.891	0.793	11.0	94	-0.01
39	Benzo(a,h,i)perylene	1.049	1.158	-10.4	97	-0.01

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE122719\
Data File : BE100999.D
Acq On : 27 Dec 2019 18:44
Operator : JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Dec 27 23:40:37 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE122419.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Dec 27 17:39:05 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			