

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE123114\
 Data File : BE089297.D
 Acq On : 31 Dec 2014 17:51
 Operator : TP/IZ
 Sample : F5245-05RX
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 Y9RC1RX

Quant Time: Jan 02 12:35:35 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SOM01.2-EPA-SIM-BE123014.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jan 01 03:02:36 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.95	152	1384	0.40	ng/ul	0.00
4) Naphthalene-d8	8.51	136	5464	0.40	ng/ul	0.00
8) Acenaphthene-d10	10.65	164	2668	0.40	ng/ul	0.00
12) Phenanthrene-d10	12.71	188	3964	0.40	ng/ul	0.00
18) Chrysene-d12	18.06	240	2602	0.40	ng/ul	0.00
22) Perylene-d12	21.12	264	1880	0.40	ng/ul	0.00
System Monitoring Compounds						
2) 1,4-Dioxane-d8	2.15	96	658	0.32	ng/uL	0.00
6) 2-Methylnaphthalene-d10	0.00	152	0	0.00	ng/ul	
16) Fluoranthene-d10	0.00	212	0d	0.00	ng/ul	
Target Compounds						
3) 1,4-Dioxane	2.18	88	1361	0.59	ng/ul#	Qvalue 84
17) Fluoranthene	14.96	202	999	0.03	ng/ul#	54
19) Pyrene	15.41	202	1053	0.03	ng/ul#	57

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE123114\
Data File : BE089297.D
Acq On : 31 Dec 2014 17:51
Operator : TP/IZ
Sample : F5245-05RX
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
Y9RC1RX

Quant Time: Jan 02 12:35:35 2015
Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SOM01.2-EPA-SIM-BE123014.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jan 01 03:02:36 2015
Response via : Initial Calibration

