

Data Path : Z:\HPCHEM1\BNA_E\Data\BE123114\
 Data File : BE089300.D
 Acq On : 31 Dec 2014 19:32
 Operator : TP/IZ
 Sample : SSTD0.454
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTD0.454

Quant Time: Jan 01 03:57:26 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SOM01.2-EPA-SIM-BE123014.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jan 01 03:02:36 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	102	0.00
2 SURR	1,4-Dioxane-d8	0.601	0.639	-6.3	101	0.00
3	1,4-Dioxane	0.670	0.697	-4.0	101	0.00
4 I	Naphthalene-d8	1.000	1.000	0.0	104	0.00
5	Naphthalene	0.950	1.010	-6.3	102	0.00
6 SURR	2-Methylnaphthalene-d10	0.459	0.491	-7.0	103	0.00
7	2-Methylnaphthalene	0.580	0.628	-8.3	103	0.00
8 I	Acenaphthene-d10	1.000	1.000	0.0	103	0.00
9	Acenaphthylene	6.314	7.017	-11.1	106	0.00
10 C	Acenaphthene	4.328	4.562	-5.4	103	0.00
11	Fluorene	1.146	1.212	-5.8	102	0.00
12 I	Phenanthrene-d10	1.000	1.000	0.0	103	-0.01
13	Pentachlorophenol	0.041	0.047	-14.6	101	0.00
14	Phenanthrene	4.281	4.547	-6.2	102	-0.01
15	Anthracene	3.824	4.066	-6.3	102	0.00
16 SURR	Fluoranthene-d10	0.693	0.687	0.9	95	0.00
17 C	Fluoranthene	3.812	3.920	-2.8	98	0.00
18 I	Chrysene-d12	1.000	1.000	0.0	94	0.00
19	Pyrene	5.160	5.672	-9.9	96	0.00
20	Benzo(a)anthracene	0.902	0.906	-0.4	87	0.00
21	Chrysene	1.093	1.190	-8.9	95	0.00
22 I	Perylene-d12	1.000	1.000	0.0	92	0.00
23	Benzo(b)fluoranthene	1.026	0.915	10.8	75	0.00
24	Benzo(k)fluoranthene	1.319	1.407	-6.7	92	0.00
25 C	Benzo(a)pyrene	1.026	1.026	0.0	83	0.00
26	Indeno(1,2,3-cd)pyrene	2.822	2.347	16.8	70	0.00
27	Dibenzo(a,h)anthracene	0.832	0.761	8.5	79	0.00
28	Benzo(g,h,i)perylene	4.060	4.004	1.4	85	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0