

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE123114\
 Data File : BE089295.D
 Acq On : 31 Dec 2014 16:45
 Operator : TP/IZ
 Sample : F5245-01RX
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 Y9R96RX

Quant Time: Jan 02 12:32:35 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SOM01.2-EPA-SIM-BE123014.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jan 01 03:02:36 2015
 Response via : Initial Calibration


Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.95	152	1462	0.40	ng/ul	0.00
4) Naphthalene-d8	8.51	136	5705	0.40	ng/ul	0.00
8) Acenaphthene-d10	10.65	164	2772	0.40	ng/ul	0.00
12) Phenanthrene-d10	12.71	188	4174	0.40	ng/ul	0.00
18) Chrysene-d12	18.06	240	2634	0.40	ng/ul	0.00
22) Perylene-d12	21.12	264	2015	0.40	ng/ul	0.00
System Monitoring Compounds						
2) 1,4-Dioxane-d8	2.15	96	692	0.31	ng/uL	0.00
6) 2-Methylnaphthalene-d10	0.00	152	0	0.00	ng/ul	
16) Fluoranthene-d10	0.00	212	0d	0.00	ng/ul	
Target Compounds						
3) 1,4-Dioxane	2.19	88	907	0.37	ng/ul#	Qvalue 84
17) Fluoranthene	14.96	202	1411	0.04	ng/ul	66
19) Pyrene	15.41	202	1495	0.04	ng/ul#	66


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE123114\
Data File : BE089295.D
Acq On : 31 Dec 2014 16:45
Operator : TP/IZ
Sample : F5245-01RX
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
Y9R96RX

Quant Time: Jan 02 12:32:35 2015
Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SOM01.2-EPA-SIM-BE123014.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jan 01 03:02:36 2015
Response via : Initial Calibration





#3

1,4-Dioxane

Concen: 0.37 ng/ul

RT: 2.19 min Scan# 131

Delta R.T. -0.00 min

Lab File: BE089295.D

Acq: 31 Dec 2014 16:45

Instrument :

BNA_E

ClientSampleId :

Y9R96RX


Tgt Ion: 88 Resp: 907

Ion Ratio Lower Upper




88 100

58 76.2 61.5 92.3

43 0.0 23.9 35.9#



Abundance Ion 88.00 (87.70 to 88.70): BE0
Ion 58.00 (57.70 to 58.70): BE0
Ion 43.00 (42.70 to 43.70): BE0



#17

Fluoranthene

Concen: 0.04 ng/ul

RT: 14.96 min Scan# 2031

Delta R.T. 0.01 min

Lab File: BE089295.D

Acq: 31 Dec 2014 16:45


Tgt Ion: 202 Resp: 1411

Ion Ratio Lower Upper



202 100


101 8.8 0.0 24.6

203 35.3 0.0 37.8



Abundance Ion 202.00 (201.70 to 202.70): E
Ion 101.00 (100.70 to 101.70): E
Ion 203.00 (202.70 to 203.70): E





#19

Pyrene

Concen: 0.04 ng/ul

RT: 15.41 min Scan# 2100

Delta R.T. 0.01 min

Lab File: BE089295.D

Acq: 31 Dec 2014 16:45

Instrument :

BNA_E

ClientSampleId :

Y9R96RX


Tgt Ion: 202 Resp: 1495

Ion Ratio Lower Upper

202 100

200 8.9 4.2 6.4#

203 36.0 14.4 21.6#



Abundance Ion 202.00 (201.70 to 202.70): E

Ion 200.00 (199.70 to 200.70): E

Ion 203.00 (202.70 to 203.70): E

