

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE123114\
 Data File : BE089299.D
 Acq On : 31 Dec 2014 18:56
 Operator : TP/IZ
 Sample : F5245-07RX
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 Y9RC3RX

Quant Time: Jan 02 12:37:37 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SOM01.2-EPA-SIM-BE123014.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jan 01 03:02:36 2015
 Response via : Initial Calibration


Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.95	152	1544	0.40	ng/ul	0.00
4) Naphthalene-d8	8.51	136	6029	0.40	ng/ul	0.00
8) Acenaphthene-d10	10.65	164	2943	0.40	ng/ul	0.00
12) Phenanthrene-d10	12.71	188	4383	0.40	ng/ul	0.00
18) Chrysene-d12	18.06	240	2745	0.40	ng/ul	0.00
22) Perylene-d12	21.12	264	2144	0.40	ng/ul	0.00
System Monitoring Compounds						
2) 1,4-Dioxane-d8	2.16	96	670	0.29	ng/uL	0.00
6) 2-Methylnaphthalene-d10	0.00	152	0	0.00	ng/ul	
16) Fluoranthene-d10	0.00	212	0d	0.00	ng/ul	
Target Compounds						
3) 1,4-Dioxane	2.20	88	727	0.28	ng/ul#	Qvalue 82
17) Fluoranthene	14.96	202	1170	0.03	ng/ul#	58
19) Pyrene	15.41	202	1200	0.03	ng/ul#	60


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE123114\
Data File : BE089299.D
Acq On : 31 Dec 2014 18:56
Operator : TP/IZ
Sample : F5245-07RX
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
Y9RC3RX

Quant Time: Jan 02 12:37:37 2015
Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SOM01.2-EPA-SIM-BE123014.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jan 01 03:02:36 2015
Response via : Initial Calibration





#3

1,4-Dioxane

Concen: 0.28 ng/ul

RT: 2.20 min Scan# 133

Delta R.T. 0.01 min

Lab File: BE089299.D

Acq: 31 Dec 2014 18:56

Instrument :

BNA_E

ClientSampleId :

Y9RC3RX


Tgt Ion: 88 Resp: 727

Ion Ratio Lower Upper




88 100

58 74.1 61.5 92.3

43 0.0 23.9 35.9#



Abundance Ion 88.00 (87.70 to 88.70): BE0
Ion 58.00 (57.70 to 58.70): BE0
Ion 43.00 (42.70 to 43.70): BE0



#17

Fluoranthene

Concen: 0.03 ng/ul

RT: 14.96 min Scan# 2031

Delta R.T. 0.01 min

Lab File: BE089299.D

Acq: 31 Dec 2014 18:56


Tgt Ion: 202 Resp: 1170

Ion Ratio Lower Upper



202 100




101 9.3 0.0 24.6

203 39.2 0.0 37.8#



Abundance Ion 202.00 (201.70 to 202.70): E
Ion 101.00 (100.70 to 101.70): E
Ion 203.00 (202.70 to 203.70): E





#19

Pyrene

Concen: 0.03 ng/ul

RT: 15.41 min Scan# 2100

Delta R.T. 0.01 min

Lab File: BE089299.D

Acq: 31 Dec 2014 18:56

Instrument :

BNA_E

ClientSampleId :

Y9RC3RX

Tgt Ion: 202 Resp: 1200

Ion Ratio Lower Upper

202 100

200 10.0 4.2 6.4#

203 39.0 14.4 21.6#

