


Data Path : Z:\svoasrv\HPCHEM1\BNA\_F\Data\BF010219\  
 Data File : BF111803.D  
 Acq On : 2 Jan 2019 11:05  
 Operator : JU/SJ  
 Sample : DFTPP  
 Misc :  
 ALS Vial : 1 Sample Multiplier: 1

Instrument :  
 BNA\_F  
 ClientSampleId :  
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA\_F\METHODS\8270-BF121918.M  
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 Last Update : Wed Dec 19 17:39:39 2018



AutoFind: Scans 1630, 1631, 1632; Background Corrected with Scan 1625

| Target<br>Mass | Rel. to<br>Mass | Lower<br>Limit% | Upper<br>Limit% | Rel.<br>Abn% | Raw<br>Abn | Result<br>Pass/Fail |
|----------------|-----------------|-----------------|-----------------|--------------|------------|---------------------|
| 51             | 198             | 10              | 80              | 41.7         | 44581      | PASS                |
| 68             | 69              | 0.00            | 2               | 1.7          | 669        | PASS                |
| 69             | 198             | 0.00            | 100             | 36.9         | 39408      | PASS                |
| 70             | 69              | 0.00            | 2               | 0.9          | 349        | PASS                |
| 127            | 198             | 10              | 80              | 47.5         | 50757      | PASS                |
| 197            | 198             | 0.00            | 2               | 0.0          | 0          | PASS                |
| 198            | 198             | 100             | 100             | 100.0        | 106808     | PASS                |
| 199            | 198             | 5               | 9               | 6.4          | 6851       | PASS                |
| 275            | 198             | 10              | 60              | 26.9         | 28759      | PASS                |
| 365            | 198             | 1               | 100             | 3.5          | 3708       | PASS                |
| 441            | 198             | 0.01            | 100             | 13.1         | 13967      | PASS                |
| 442            | 198             | 50              | 100             | 85.2         | 91010      | PASS                |
| 443            | 442             | 15              | 24              | 18.9         | 17204      | PASS                |