

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF010421\
 Data File : BF122909.D
 Acq On : 4 Jan 2021 14:45
 Operator : JU/CG
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Jan 04 15:35:29 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF120520.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Dec 31 15:53:00 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	67	0.00
2	1,4-Dioxane	40.000	37.830	5.4	63	0.00
3	Pyridine	40.000	37.150	7.1	60	0.00
4	n-Nitrosodimethylamine	40.000	33.450	16.4	54	-0.01
5 S	2-Fluorophenol	80.000	78.055	2.4	65	0.00
6	Aniline	40.000	36.004	10.0	59	0.00
7 S	Phenol-d6	80.000	74.264	7.2	62	0.00
8	2-Chlorophenol	40.000	38.330	4.2	63	0.00
9	Benzaldehyde	40.000	38.519	3.7	73	0.00
10 C	Phenol	40.000	36.547	8.6	61	0.00
11	bis(2-Chloroethyl)ether	40.000	37.149	7.1	61	0.00
12	1,3-Dichlorobenzene	40.000	38.076	4.8	63	0.00
13 C	1,4-Dichlorobenzene	40.000	38.131	4.7	64	0.00
14	1,2-Dichlorobenzene	40.000	35.829	10.4	63	0.00
15	Benzyl Alcohol	40.000	36.631	8.4	59	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	32.568	18.6	54	0.00
17	2-Methylphenol	40.000	38.195	4.5	62	0.00
18	Hexachloroethane	40.000	37.675	5.8	62	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	34.481	13.8	57	0.00
20	3+4-Methylphenols	40.000	37.030	7.4	63	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	67	0.00
22	Acetophenone	40.000	37.241	6.9	63	0.00
23 S	Nitrobenzene-d5	80.000	73.934	7.6	61	0.00
24	Nitrobenzene	40.000	36.897	7.8	61	0.00
25	Isophorone	40.000	35.754	10.6	59	0.00
26 C	2-Nitrophenol	40.000	40.514	-1.3	65	0.00
27	2,4-Dimethylphenol	40.000	37.339	6.7	61	0.00
28	bis(2-Chloroethoxy)methane	40.000	36.552	8.6	60	0.00
29 C	2,4-Dichlorophenol	40.000	37.579	6.1	62	0.00
30	1,2,4-Trichlorobenzene	40.000	36.979	7.6	62	0.00
31	Naphthalene	40.000	37.411	6.5	62	0.00
32	Benzoic acid	40.000	31.515	21.2	48	-0.01
33	4-Chloroaniline	40.000	37.582	6.0	62	0.00
34 C	Hexachlorobutadiene	40.000	36.956	7.6	61	0.00
35	Caprolactam	40.000	37.760	5.6	62	-0.01
36 C	4-Chloro-3-methylphenol	40.000	37.569	6.1	61	0.00
37	2-Methylnaphthalene	40.000	37.063	7.3	62	0.00
38	1-Methylnaphthalene	40.000	36.754	8.1	62	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	66	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	37.616	6.0	61	0.00
41 P	Hexachlorocyclopentadiene	40.000	38.562	3.6	59	0.00
42 S	2,4,6-Tribromophenol	80.000	73.403	8.2	59	0.00
43 C	2,4,6-Trichlorophenol	40.000	36.428	8.9	58	0.00
44	2,4,5-Trichlorophenol	40.000	36.226	9.4	59	0.00
45 S	2-Fluorobiphenyl	80.000	69.163	13.5	61	0.00
46	1,1'-Biphenyl	40.000	37.126	7.2	61	0.00
47	2-Chloronaphthalene	40.000	37.508	6.2	62	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF010421\
 Data File : BF122909.D
 Acq On : 4 Jan 2021 14:45
 Operator : JU/CG
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Jan 04 15:35:29 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF120520.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Dec 31 15:53:00 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	37.435	6.4	59	0.00
49	Acenaphthylene	40.000	37.187	7.0	61	0.00
50	Dimethylphthalate	40.000	37.535	6.2	61	0.00
51	2,6-Dinitrotoluene	40.000	39.145	2.1	62	0.00
52 C	Acenaphthene	40.000	34.638	13.4	56	0.00
53	3-Nitroaniline	40.000	38.299	4.3	61	0.00
54 P	2,4-Dinitrophenol	40.000	48.103	-20.3	73	0.00
55	Dibenzofuran	40.000	36.220	9.5	60	0.00
56 P	4-Nitrophenol	40.000	38.731	3.2	60	0.00
57	2,4-Dinitrotoluene	40.000	37.584	6.0	59	0.00
58	Fluorene	40.000	36.363	9.1	63	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	35.405	11.5	56	0.00
60	Diethylphthalate	40.000	36.617	8.5	60	0.00
61	4-Chlorophenyl-phenylether	40.000	37.181	7.0	62	0.00
62	4-Nitroaniline	40.000	38.813	3.0	62	0.00
63	Azobenzene	40.000	35.570	11.1	58	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	66	0.00
65	4,6-Dinitro-2-methylphenol	40.000	39.004	2.5	60	0.00
66 c	n-Nitrosodiphenylamine	40.000	36.911	7.7	60	0.00
67	4-Bromophenyl-phenylether	40.000	37.360	6.6	61	0.00
68	Hexachlorobenzene	40.000	37.133	7.2	61	0.00
69	Atrazine	40.000	32.613	18.5	54	0.00
70 C	Pentachlorophenol	40.000	37.809	5.5	57	0.00
71	Phenanthrene	40.000	36.781	8.0	61	0.00
72	Anthracene	40.000	37.493	6.3	62	0.00
73	Carbazole	40.000	37.465	6.3	62	0.00
74	Di-n-butylphthalate	40.000	36.206	9.5	60	0.00
75 C	Fluoranthene	40.000	36.348	9.1	60	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	63	0.00
77	Benzydine	40.000	45.774	-14.4	63	0.00
78	Pyrene	40.000	38.241	4.4	59	0.00
79 S	Terphenyl-d14	80.000	74.635	6.7	61	0.00
80	Butylbenzylphthalate	40.000	37.082	7.3	57	0.00
81	Benzo(a)anthracene	40.000	38.128	4.7	60	0.00
82	3,3'-Dichlorobenzidine	40.000	36.701	8.2	57	0.00
83	Chrysene	40.000	37.447	6.4	59	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	35.888	10.3	57	0.00
85 c	Di-n-octyl phthalate	40.000	33.887	15.3	53	0.00
86 I	Perylene-d12	20.000	20.000	0.0	65	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	39.435	1.4	65	0.00
88	Benzo(b)fluoranthene	40.000	35.851	10.4	55	0.00
89	Benzo(k)fluoranthene	40.000	38.547	3.6	64	0.00
90 C	Benzo(a)pyrene	40.000	37.326	6.7	60	0.00
91	Dibenzo(a,h)anthracene	40.000	39.591	1.0	66	0.00
92	Benzo(g,h,i)perylene	40.000	40.763	-1.9	69	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF010421\
Data File : BF122909.D
Acq On : 4 Jan 2021 14:45
Operator : JU/CG
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Jan 04 15:35:29 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF120520.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Dec 31 15:53:00 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0