

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF010422\
 Data File : BF126479.D
 Acq On : 4 Jan 2022 21:08
 Operator : JU/CG
 Sample : M5342-12MS
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :

Quant Time: Jan 05 02:01:39 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF122921.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jan 05 01:57:07 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.781	152	132907	20.000	ng	0.00
21) Naphthalene-d8	8.069	136	560180	20.000	ng	0.00
39) Acenaphthene-d10	9.828	164	260169	20.000	ng	0.00
64) Phenanthrene-d10	11.310	188	421865	20.000	ng	0.00
76) Chrysene-d12	13.951	240	231408	20.000	ng	0.00
86) Perylene-d12	15.386	264	227186	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.410	112	897263	96.949	ng	0.02
7) Phenol-d6	6.422	99	1155454	96.391	ng	0.00
23) Nitrobenzene-d5	7.351	82	739679	67.799	ng	0.00
42) 2,4,6-Tribromophenol	10.616	330	221749	110.299	ng	0.00
45) 2-Fluorobiphenyl	9.145	172	1012625	68.492	ng	0.00
79) Terphenyl-d14	12.898	244	853455	71.657	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	2.599	88	217029	46.763	ng	99
3) Pyridine	3.316	79	479885	38.333	ng	98
4) n-Nitrosodimethylamine	3.269	42	229636	44.615	ng	95
6) Aniline	6.445	93	270285	18.133	ng	99
8) 2-Chlorophenol	6.569	128	476415	51.618	ng	96
9) Benzaldehyde	6.334	77	310005	41.776	ng	92
10) Phenol	6.434	94	597700	44.680	ng	98
11) bis(2-Chloroethyl)ether	6.522	93	488059	47.781	ng	97
12) 1,3-Dichlorobenzene	6.728	146	455008	49.333	ng	97
13) 1,4-Dichlorobenzene	6.804	146	453904	49.303	ng	97
14) 1,2-Dichlorobenzene	6.951	146	446630	49.790	ng	97
15) Benzyl Alcohol	6.928	79	416120	49.615	ng	98
16) 2,2'-oxybis(1-Chloropr...	7.063	45	769316	47.956	ng	99
17) 2-Methylphenol	7.045	107	395338	48.695	ng	99
18) Hexachloroethane	7.298	117	184137	50.201	ng	96
19) n-Nitroso-di-n-propyla...	7.204	70	311545	45.546	ng	99
20) 3+4-Methylphenols	7.198	107	427922	44.272	ng	91
22) Acetophenone	7.198	105	616360	47.594	ng	97
24) Nitrobenzene	7.369	77	501383	45.926	ng	95
25) Isophorone	7.610	82	924681	47.346	ng	98
26) 2-Nitrophenol	7.687	139	254375	52.560	ng	94
27) 2,4-Dimethylphenol	7.728	122	415521	54.908	ng	98
28) bis(2-Chloroethoxy)met...	7.822	93	599139	46.961	ng	98
29) 2,4-Dichlorophenol	7.928	162	346443	50.137	ng	98
30) 1,2,4-Trichlorobenzene	8.010	180	357507	49.643	ng	99
31) Naphthalene	8.092	128	1312276	49.631	ng	100
32) Benzoic acid	7.869	122	337454	59.483	ng	95
33) 4-Chloroaniline	8.139	127	91903	8.297	ng	93
34) Hexachlorobutadiene	8.210	225	175124	50.032	ng	98
35) Caprolactam	8.469	113	3725	2.119	ng	94
36) 4-Chloro-3-methylphenol	8.628	107	413447	49.028	ng	99
37) 2-Methylnaphthalene	8.781	142	827307	53.041	ng	99
38) 1-Methylnaphthalene	8.881	142	761378	50.489	ng	99
40) 1,2,4,5-Tetrachloroben...	8.951	216	279510	49.441	ng	98
41) Hexachlorocyclopentadiene	8.939	237	294949	128.141	ng	97
43) 2,4,6-Trichlorophenol	9.063	196	212491	50.824	ng	96

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF010422\
 Data File : BF126479.D
 Acq On : 4 Jan 2022 21:08
 Operator : JU/CG
 Sample : M5342-12MS
 Misc :
 ALS Vial : 13 Sample Multiplier: 1


Instrument :
 BNA_F
 ClientSampleId :


Quant Time: Jan 05 02:01:39 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF122921.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jan 05 01:57:07 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
44) 2,4,5-Trichlorophenol	9.104	196	222021	49.334	ng	98
46) 1,1'-Biphenyl	9.251	154	940027	49.263	ng	99
47) 2-Chloronaphthalene	9.275	162	690050	47.744	ng	98
48) 2-Nitroaniline	9.369	65	254786	43.655	ng	88
49) Acenaphthylene	9.686	152	1086635	49.572	ng	99
50) Dimethylphthalate	9.557	163	805417	49.589	ng	100
51) 2,6-Dinitrotoluene	9.616	165	186489	53.167	ng	86
52) Acenaphthene	9.863	154	680191	47.331	ng	99
53) 3-Nitroaniline	9.775	138	93566	19.455	ng	90
54) 2,4-Dinitrophenol	9.892	184	178904	98.288	ng	# 81
55) Dibenzofuran	10.033	168	919916	47.351	ng	97
56) 4-Nitrophenol	9.951	139	336819	98.476	ng	97
57) 2,4-Dinitrotoluene	10.016	165	242417	53.568	ng	97
58) Fluorene	10.375	166	663240	48.284	ng	99
59) 2,3,4,6-Tetrachlorophenol	10.151	232	170658	52.833	ng	92
60) Diethylphthalate	10.251	149	769962	49.189	ng	99
61) 4-Chlorophenyl-phenyle...	10.369	204	287386	47.943	ng	95
62) 4-Nitroaniline	10.398	138	197843	43.680	ng	92
63) Azobenzene	10.528	77	915816	45.422	ng	98
65) 4,6-Dinitro-2-methylph...	10.428	198	117464	51.417	ng	98
66) n-Nitrosodiphenylamine	10.486	169	626135	47.839	ng	100
67) 4-Bromophenyl-phenylether	10.857	248	188145	50.246	ng	99
68) Hexachlorobenzene	10.922	284	222962	51.190	ng	98
69) Atrazine	11.016	200	190675	84.604	ng	98
70) Pentachlorophenol	11.122	266	222559	105.133	ng	98
71) Phenanthrene	11.339	178	1032392	47.867	ng	99
72) Anthracene	11.386	178	1059996	49.117	ng	100
73) Carbazole	11.545	167	958221	45.343	ng	99
74) Di-n-butylphthalate	11.874	149	1202487	49.980	ng	99
75) Fluoranthene	12.522	202	953760	48.452	ng	98
77) Benzidine	12.645	184	299314	80.857	ng	99
78) Pyrene	12.751	202	986347	50.437	ng	100
80) Butylbenzylphthalate	13.374	149	453809	52.369	ng	89
81) Benzo(a)anthracene	13.939	228	614569	46.180	ng	99
82) 3,3'-Dichlorobenzidine	13.898	252	161077	25.944	ng	98
83) Chrysene	13.974	228	674654	49.422	ng	99
84) Bis(2-ethylhexyl)phtha...	13.933	149	492364	54.825	ng	100
85) Di-n-octyl phthalate	14.545	149	916860	58.867	ng	97
87) Indeno(1,2,3-cd)pyrene	16.809	276	804217	52.894	ng	99
88) Benzo(b)fluoranthene	14.974	252	687566	50.804	ng	98
89) Benzo(k)fluoranthene	15.004	252	639259	48.760	ng	98
90) Benzo(a)pyrene	15.327	252	668591	58.918	ng	99
91) Dibenzo(a,h)anthracene	16.827	278	656837	53.538	ng	99
92) Benzo(g,h,i)perylene	17.239	276	717771	54.942	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Jan 05 02:01:39 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF122921.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jan 05 01:57:07 2022
Response via : Initial Calibration









#1
 1,4-Dichlorobenzene-d4
 Concen: 20.000 ng
 RT: 6.781 min Scan# 79
 Delta R.T. -0.005 min
 Lab File: BF126479.D
 Acq: 4 Jan 2022 21:08




Instrument :
 BNA_F
 ClientSampleId :


Tgt Ion:152 Resp: 132907
 Ion Ratio Lower Upper
 152 100
 150 159.9 128.1 192.1
 115 66.5 56.7 85.1



#2
 1,4-Dioxane
 Concen: 46.763 ng
 RT: 2.599 min Scan# 87
 Delta R.T. 0.112 min
 Lab File: BF126479.D
 Acq: 4 Jan 2022 21:08


Tgt Ion: 88 Resp: 217029
 Ion Ratio Lower Upper
 88 100
 58 78.5 62.8 94.2
 43 26.1 22.3 33.5








#3
Pyridine
Concen: 38.333 ng
RT: 3.316 min Scan# 201
Delta R.T. 0.106 min
Lab File: BF126479.D
Acq: 4 Jan 2022 21:08


Instrument :
BNA_F
ClientSampleId :





Tgt Ion: 79 Resp: 479885
Ion Ratio Lower Upper
79 100
52 71.8 56.2 84.4
51 32.7 26.7 40.1




#4
n-Nitrosodimethylamine
Concen: 44.615 ng
RT: 3.269 min Scan# 201
Delta R.T. 0.106 min
Lab File: BF126479.D
Acq: 4 Jan 2022 21:08



Tgt Ion: 42 Resp: 229636
Ion Ratio Lower Upper
42 100
74 163.0 125.6 188.4
44 8.1 5.7 8.5









#5
2-Fluorophenol
Concen: 96.949 ng
RT: 5.410 min Scan# 50
Delta R.T. 0.018 min
Lab File: BF126479.D
Acq: 4 Jan 2022 21:08




Instrument :
BNA_F
ClientSampleId :


Tgt Ion:112 Resp: 897263
Ion Ratio Lower Upper
112 100
64 60.1 49.5 74.3
63 28.9 23.9 35.9



#6
Aniline
Concen: 18.133 ng
RT: 6.445 min Scan# 741
Delta R.T. 0.000 min
Lab File: BF126479.D
Acq: 4 Jan 2022 21:08

Tgt Ion: 93 Resp: 270285
Ion Ratio Lower Upper
93 100
66 40.4 32.2 48.4
65 21.3 16.3 24.5





#7

Phenol-d6

Concen: 96.391 ng

RT: 6.422 min Scan# 71

Delta R.T. 0.000 min

Lab File: BF126479.D

Acq: 4 Jan 2022 21:08





Instrument :

BNA_F

ClientSampleId :

Tgt Ion: 99 Resp: 1155454

Ion	Ratio	Lower	Upper
99	100		
42	17.7	14.2	21.4
71	30.0	24.3	36.5



#8

2-Chlorophenol

Concen: 51.618 ng

RT: 6.569 min Scan# 762




Delta R.T. 0.000 min


Lab File: BF126479.D

Acq: 4 Jan 2022 21:08

Tgt Ion:128 Resp: 476415

Ion	Ratio	Lower	Upper
128	100		
130	33.2	12.7	52.7
64	51.4	35.6	75.6









#9
Benzaldehyde
Concen: 41.776 ng
RT: 6.334 min Scan# 719
Delta R.T. 0.000 min
Lab File: BF126479.D
Acq: 4 Jan 2022 21:08




Instrument :
BNA_F
ClientSampleId :


Tgt Ion: 77 Resp: 310005
Ion Ratio Lower Upper
77 100
105 102.0 74.9 114.9
106 100.8 72.4 112.4



#10
Phenol
Concen: 44.680 ng
RT: 6.434 min Scan# 739
Delta R.T. 0.000 min
Lab File: BF126479.D
Acq: 4 Jan 2022 21:08

Tgt Ion: 94 Resp: 597700
Ion Ratio Lower Upper
94 100
65 27.8 7.2 47.2
66 36.4 14.5 54.5









#11
bis(2-Chloroethyl)ether
Concen: 47.781 ng
RT: 6.522 min Scan# 754
Delta R.T. 0.000 min
Lab File: BF126479.D
Acq: 4 Jan 2022 21:08




Instrument :
BNA_F
ClientSampleId :


Tgt Ion: 93 Resp: 488059
Ion Ratio Lower Upper
93 100
63 77.0 60.5 100.5
95 33.4 12.9 52.9



#12
1,3-Dichlorobenzene
Concen: 49.333 ng
RT: 6.728 min Scan# 789
Delta R.T. 0.000 min
Lab File: BF126479.D
Acq: 4 Jan 2022 21:08

Tgt Ion:146 Resp: 455008
Ion Ratio Lower Upper
146 100
148 64.1 50.9 76.3
75 31.1 28.4 42.6





#13

1,4-Dichlorobenzene

Concen: 49.303 ng

RT: 6.804 min Scan# 804

Delta R.T. 0.000 min

Lab File: BF126479.D

Acq: 4 Jan 2022 21:08

Instrument :

BNA_F

ClientSampleId :





Tgt Ion:146 Resp: 453904

Ion Ratio Lower Upper

146 100

148 65.6 49.8 74.6

111 44.7 36.1 54.1



#14

1,2-Dichlorobenzene

Concen: 49.790 ng

RT: 6.951 min Scan# 827

Delta R.T. -0.006 min

Lab File: BF126479.D

Acq: 4 Jan 2022 21:08




Tgt Ion:146 Resp: 446630


Ion Ratio Lower Upper

146 100

148 62.3 51.6 77.4


111 47.4 40.2 60.2








#15
Benzyl Alcohol
Concen: 49.615 ng
RT: 6.928 min Scan# 81
Delta R.T. 0.000 min
Lab File: BF126479.D
Acq: 4 Jan 2022 21:08


Instrument :
BNA_F
ClientSampleId :





Tgt Ion: 79 Resp: 416120
Ion Ratio Lower Upper
79 100
108 82.4 63.2 94.8
77 62.1 49.4 74.2




#16
2,2'-oxybis(1-Chloropropane)
Concen: 47.956 ng
RT: 7.063 min Scan# 846
Delta R.T. 0.000 min
Lab File: BF126479.D
Acq: 4 Jan 2022 21:08




Tgt Ion: 45 Resp: 769316
Ion Ratio Lower Upper
45 100
77 13.0 0.0 32.4
79 9.9 0.0 29.8








#17
2-Methylphenol
Concen: 48.695 ng
RT: 7.045 min Scan# 84
Delta R.T. 0.000 min
Lab File: BF126479.D
Acq: 4 Jan 2022 21:08

Instrument :
BNA_F
ClientSampleId :







Tgt Ion:107 Resp: 395338
Ion Ratio Lower Upper
107 100
108 116.8 93.4 140.0
77 44.1 36.1 54.1
79 42.3 34.9 52.3



#18
Hexachloroethane
Concen: 50.201 ng
RT: 7.298 min Scan# 886
Delta R.T. 0.000 min
Lab File: BF126479.D
Acq: 4 Jan 2022 21:08


Tgt Ion:117 Resp: 184137
Ion Ratio Lower Upper
117 100
119 96.2 75.4 113.0
201 77.4 56.8 85.2








#19
n-Nitroso-di-n-propylamine
Concen: 45.546 ng
RT: 7.204 min Scan# 81
Delta R.T. 0.000 min
Lab File: BF126479.D
Acq: 4 Jan 2022 21:08

Instrument :
BNA_F
ClientSampleId :






Tgt Ion:	70	Resp:	311545
Ion	Ratio	Lower	Upper
70	100		
42	57.4	45.8	68.8
101	8.9	6.9	10.3
130	19.1	14.4	21.6



#20
3+4-Methylphenols
Concen: 44.272 ng
RT: 7.198 min Scan# 869
Delta R.T. 0.000 min
Lab File: BF126479.D
Acq: 4 Jan 2022 21:08


Tgt Ion:	107	Resp:	427922
Ion	Ratio	Lower	Upper
107	100		
108	93.5	72.4	112.4
77	122.0	83.8	123.8
79	28.8	8.5	48.5








#21
Naphthalene-d8
Concen: 20.000 ng
RT: 8.069 min Scan# 1019
Delta R.T. 0.000 min
Lab File: BF126479.D
Acq: 4 Jan 2022 21:08


Instrument :
BNA_F
ClientSampleId :





Tgt Ion:136 Resp: 560180
Ion Ratio Lower Upper
136 100
137 11.3 8.6 13.0
54 8.6 7.8 11.6
68 7.0 6.3 9.5




#22
Acetophenone
Concen: 47.594 ng
RT: 7.198 min Scan# 869
Delta R.T. 0.000 min
Lab File: BF126479.D
Acq: 4 Jan 2022 21:08



Tgt Ion:105 Resp: 616360
Ion Ratio Lower Upper
105 100
71 6.9 5.4 8.0
51 36.9 27.4 41.0
120 21.4 17.0 25.6





#23

Nitrobenzene-d5

Concen: 67.799 ng

RT: 7.351 min Scan# 895

Delta R.T. 0.000 min

Lab File: BF126479.D

Acq: 4 Jan 2022 21:08

Instrument :

BNA_F

ClientSampleId :





Tgt Ion: 82 Resp: 739679

Ion Ratio Lower Upper

82 100

128 44.9 32.7 49.1

54 50.1 41.7 62.5



#24

Nitrobenzene

Concen: 45.926 ng

RT: 7.369 min Scan# 898

Delta R.T. 0.000 min

Lab File: BF126479.D

Acq: 4 Jan 2022 21:08




Tgt Ion: 77 Resp: 501383


Ion Ratio Lower Upper

77 100

123 47.5 34.6 52.0

65 13.0 10.6 15.8









#25
Isophorone
Concen: 47.346 ng
RT: 7.610 min Scan# 91
Delta R.T. 0.000 min
Lab File: BF126479.D
Acq: 4 Jan 2022 21:08




Instrument :
BNA_F
ClientSampleId :


Tgt Ion: 82 Resp: 924681
Ion Ratio Lower Upper
82 100
95 6.0 5.0 7.4
138 16.6 12.3 18.5



#26
2-Nitrophenol
Concen: 52.560 ng
RT: 7.687 min Scan# 952
Delta R.T. 0.000 min
Lab File: BF126479.D
Acq: 4 Jan 2022 21:08

Tgt Ion:139 Resp: 254375
Ion Ratio Lower Upper
139 100
109 23.7 20.0 30.0
65 45.7 40.8 61.2









#27
2,4-Dimethylphenol
Concen: 54.908 ng
RT: 7.728 min Scan# 91
Delta R.T. 0.000 min
Lab File: BF126479.D
Acq: 4 Jan 2022 21:08




Instrument :
BNA_F
ClientSampleId :


Tgt Ion:122 Resp: 415521
Ion Ratio Lower Upper
122 100
107 110.7 91.2 136.8
121 56.0 44.4 66.6



#28
bis(2-Chloroethoxy)methane
Concen: 46.961 ng
RT: 7.822 min Scan# 975
Delta R.T. 0.000 min
Lab File: BF126479.D
Acq: 4 Jan 2022 21:08


Tgt Ion: 93 Resp: 599139
Ion Ratio Lower Upper
93 100
95 32.8 25.4 38.2
123 10.9 8.9 13.3








#29
2,4-Dichlorophenol
Concen: 50.137 ng
RT: 7.928 min Scan# 993
Delta R.T. 0.000 min
Lab File: BF126479.D
Acq: 4 Jan 2022 21:08

Instrument :
BNA_F
ClientSampleId :







Tgt Ion:162 Resp: 346443
Ion Ratio Lower Upper
162 100
164 62.3 43.8 83.8
98 44.7 23.6 63.6



#30
1,2,4-Trichlorobenzene
Concen: 49.643 ng
RT: 8.010 min Scan# 1007
Delta R.T. 0.000 min
Lab File: BF126479.D
Acq: 4 Jan 2022 21:08

Tgt Ion:180 Resp: 357507
Ion Ratio Lower Upper
180 100
182 94.0 75.5 113.3
145 31.1 26.2 39.2









#31
Naphthalene
Concen: 49.631 ng
RT: 8.092 min Scan# 1023
Delta R.T. 0.000 min
Lab File: BF126479.D
Acq: 4 Jan 2022 21:08




Instrument :
BNA_F
ClientSampleId :


Tgt Ion:128 Resp: 1312276
Ion Ratio Lower Upper
128 100
129 11.1 8.6 13.0
127 12.9 10.3 15.5



#32
Benzoic acid
Concen: 59.483 ng
RT: 7.869 min Scan# 983
Delta R.T. 0.024 min
Lab File: BF126479.D
Acq: 4 Jan 2022 21:08

Tgt Ion:122 Resp: 337454
Ion Ratio Lower Upper
122 100
105 117.7 101.0 141.0
77 86.6 74.5 114.5









#33
4-Chloroaniline
Concen: 8.297 ng
RT: 8.139 min Scan# 1029
Delta R.T. 0.000 min
Lab File: BF126479.D
Acq: 4 Jan 2022 21:08




Instrument :
BNA_F
ClientSampleId :


Tgt Ion:127	Resp:	91903
Ion Ratio	Lower	Upper
127	100	
129	31.0	26.5 39.7
65	26.6	26.6 39.8
92	17.2	16.1 24.1



#34
Hexachlorobutadiene
Concen: 50.032 ng
RT: 8.210 min Scan# 1041
Delta R.T. 0.000 min
Lab File: BF126479.D
Acq: 4 Jan 2022 21:08

Tgt Ion:225	Resp:	175124
Ion Ratio	Lower	Upper
225	100	
223	61.8	48.6 73.0
227	62.3	51.4 77.0









#35
 Caprolactam
 Concen: 2.119 ng
 RT: 8.469 min Scan# 1092
 Delta R.T. -0.035 min
 Lab File: BF126479.D
 Acq: 4 Jan 2022 21:08




Instrument :
 BNA_F
 ClientSampleId :


Tgt Ion:113 Resp: 3725
 Ion Ratio Lower Upper
 113 100
 55 214.1 183.9 223.9
 56 157.0 129.9 169.9



#36
 4-Chloro-3-methylphenol
 Concen: 49.028 ng
 RT: 8.628 min Scan# 1112
 Delta R.T. 0.006 min
 Lab File: BF126479.D
 Acq: 4 Jan 2022 21:08

Tgt Ion:107 Resp: 413447
 Ion Ratio Lower Upper
 107 100
 144 23.8 18.5 27.7
 142 73.6 58.1 87.1





#37

2-Methylnaphthalene

Concen: 53.041 ng

RT: 8.781 min Scan# 1138

Delta R.T. 0.000 min

Lab File: BF126479.D

Acq: 4 Jan 2022 21:08

Instrument :

BNA_F





ClientSampleId :

Tgt Ion:142 Resp: 827307

Ion Ratio Lower Upper

142 100

141 85.1 68.6 102.8



#38

1-Methylnaphthalene

Concen: 50.489 ng

RT: 8.881 min Scan# 1155

Delta R.T. 0.000 min

Lab File: BF126479.D




Acq: 4 Jan 2022 21:08


Tgt Ion:142 Resp: 761378

Ion Ratio Lower Upper

142 100

141 91.4 72.1 108.1





#39

Acenaphthene-d10

Concen: 20.000 ng

RT: 9.828 min Scan# 1167

Delta R.T. 0.000 min


Lab File: BF126479.D

Acq: 4 Jan 2022 21:08

Instrument :

BNA_F

ClientSampleId :






Tgt Ion:164 Resp: 260169

Ion Ratio Lower Upper

164 100

162 103.6 81.4 122.0

160 46.2 36.6 54.8



#40

1,2,4,5-Tetrachlorobenzene

Concen: 49.441 ng

RT: 8.951 min Scan# 1167

Delta R.T. 0.000 min

Lab File: BF126479.D

Acq: 4 Jan 2022 21:08

Tgt Ion:216 Resp: 279510




Ion Ratio Lower Upper


216 100

214 79.0 64.6 96.8

179 22.9 19.5 29.3


108 26.4 21.1 31.7








#41
Hexachlorocyclopentadiene
Concen: 128.141 ng
RT: 8.939 min Scan# 1165
Delta R.T. 0.000 min
Lab File: BF126479.D
Acq: 4 Jan 2022 21:08


Instrument :
BNA_F
ClientSampleId :





Tgt Ion: 237 Resp: 294949
Ion Ratio Lower Upper
237 100
235 61.5 38.9 78.9
272 14.1 0.0 33.5




#42
2,4,6-Tribromophenol
Concen: 110.299 ng
RT: 10.616 min Scan# 1450
Delta R.T. 0.000 min
Lab File: BF126479.D
Acq: 4 Jan 2022 21:08




Tgt Ion: 330 Resp: 221749
Ion Ratio Lower Upper
330 100
332 95.0 76.0 114.0
141 34.9 30.1 45.1








#43
2,4,6-Trichlorophenol
Concen: 50.824 ng
RT: 9.063 min Scan# 1186
Delta R.T. 0.000 min
Lab File: BF126479.D
Acq: 4 Jan 2022 21:08

Instrument :
BNA_F
ClientSampleId :







Tgt Ion:196 Resp: 212491
Ion Ratio Lower Upper
196 100
198 101.8 78.5 117.7
200 33.0 24.9 37.3



#44
2,4,5-Trichlorophenol
Concen: 49.334 ng
RT: 9.104 min Scan# 1193
Delta R.T. 0.006 min
Lab File: BF126479.D
Acq: 4 Jan 2022 21:08

Tgt Ion:196 Resp: 222021
Ion Ratio Lower Upper
196 100
198 99.7 82.0 123.0
97 65.3 53.5 80.3
132 35.6 28.7 43.1









#45
2-Fluorobiphenyl
Concen: 68.492 ng
RT: 9.145 min Scan# 11
Delta R.T. -0.006 min
Lab File: BF126479.D
Acq: 4 Jan 2022 21:08




Instrument :
BNA_F
ClientSampleId :


Tgt Ion:172 Resp: 1012625
Ion Ratio Lower Upper
172 100
171 33.9 28.3 42.5
170 23.5 19.0 28.6



#46
1,1'-Biphenyl
Concen: 49.263 ng
RT: 9.251 min Scan# 1218
Delta R.T. 0.000 min
Lab File: BF126479.D
Acq: 4 Jan 2022 21:08

Tgt Ion:154 Resp: 940027
Ion Ratio Lower Upper
154 100
153 39.1 19.0 59.0
76 13.8 0.0 34.5





#47

2-Chloronaphthalene

Concen: 47.744 ng

RT: 9.275 min Scan# 11

Delta R.T. 0.000 min


Lab File: BF126479.D

Acq: 4 Jan 2022 21:08

Instrument :

BNA_F

ClientSampleId :






Tgt Ion:162 Resp: 690050

Ion Ratio Lower Upper

162 100

127 42.7 33.5 50.3

164 33.9 26.2 39.4



#48

2-Nitroaniline

Concen: 43.655 ng

RT: 9.369 min Scan# 1238

Delta R.T. 0.000 min

Lab File: BF126479.D

Acq: 4 Jan 2022 21:08




Tgt Ion: 65 Resp: 254786


Ion Ratio Lower Upper

65 100

92 72.6 54.4 81.6

138 106.6 72.5 108.7





#49

Acenaphthylene

Concen: 49.572 ng

RT: 9.686 min Scan# 11

Delta R.T. 0.000 min

Lab File: BF126479.D

Acq: 4 Jan 2022 21:08

Instrument :

BNA_F

ClientSampleId :





Tgt Ion:152 Resp: 1086635

Ion Ratio Lower Upper

152 100

151 20.0 15.7 23.5

153 14.4 11.4 17.0



#50

Dimethylphthalate

Concen: 49.589 ng

RT: 9.557 min Scan# 1270

Delta R.T. 0.006 min

Lab File: BF126479.D

Acq: 4 Jan 2022 21:08




Tgt Ion:163 Resp: 805417


Ion Ratio Lower Upper

163 100

194 3.3 2.6 3.8

164 9.6 7.8 11.8









#51
2,6-Dinitrotoluene
Concen: 53.167 ng
RT: 9.616 min Scan# 1280
Delta R.T. 0.006 min
Lab File: BF126479.D
Acq: 4 Jan 2022 21:08




Instrument :
BNA_F
ClientSampleId :


Tgt Ion:165 Resp: 186489
Ion Ratio Lower Upper
165 100
63 50.1 48.4 72.6
89 52.1 50.5 75.7



#52
Acenaphthene
Concen: 47.331 ng
RT: 9.863 min Scan# 1322
Delta R.T. 0.000 min
Lab File: BF126479.D
Acq: 4 Jan 2022 21:08


Tgt Ion:154 Resp: 680191
Ion Ratio Lower Upper
154 100
153 110.6 87.8 131.6
152 49.5 39.4 59.2








#53
3-Nitroaniline
Concen: 19.455 ng
RT: 9.775 min Scan# 11
Delta R.T. -0.006 min
Lab File: BF126479.D
Acq: 4 Jan 2022 21:08

Instrument :
BNA_F
ClientSampleId :







Tgt Ion:138 Resp: 93566
Ion Ratio Lower Upper
138 100
108 9.0 8.6 13.0
92 119.2 105.0 157.6



#54
2,4-Dinitrophenol
Concen: 98.288 ng
RT: 9.892 min Scan# 1327
Delta R.T. 0.006 min
Lab File: BF126479.D
Acq: 4 Jan 2022 21:08

Tgt Ion:184 Resp: 178904
Ion Ratio Lower Upper
184 100
63 62.6 70.4 105.6#
154 56.1 49.7 74.5









#55
Dibenzenofuran
Concen: 47.351 ng
RT: 10.033 min Scan# 1351
Delta R.T. 0.000 min
Lab File: BF126479.D
Acq: 4 Jan 2022 21:08




Instrument :
BNA_F
ClientSampleId :


Tgt Ion:168 Resp: 919916
Ion Ratio Lower Upper
168 100
139 39.9 30.6 46.0
169 13.4 10.0 15.0



#56
4-Nitrophenol
Concen: 98.476 ng
RT: 9.951 min Scan# 1337
Delta R.T. 0.012 min
Lab File: BF126479.D
Acq: 4 Jan 2022 21:08


Tgt Ion:139 Resp: 336819
Ion Ratio Lower Upper
139 100
109 63.4 44.1 84.1
65 101.1 85.5 125.5








#57
2,4-Dinitrotoluene
Concen: 53.568 ng
RT: 10.016 min Scan# 11
Delta R.T. 0.000 min
Lab File: BF126479.D
Acq: 4 Jan 2022 21:08


Instrument :
BNA_F
ClientSampleId :





Tgt Ion:165 Resp: 242417
Ion Ratio Lower Upper
165 100
63 53.1 43.0 64.6
89 85.1 71.0 106.4
182 2.1 1.6 2.4




#58
Fluorene
Concen: 48.284 ng
RT: 10.375 min Scan# 1409
Delta R.T. 0.000 min
Lab File: BF126479.D
Acq: 4 Jan 2022 21:08




Tgt Ion:166 Resp: 663240
Ion Ratio Lower Upper
166 100
165 97.8 77.7 116.5
167 13.7 10.7 16.1








#59
2,3,4,6-Tetrachlorophenol
Concen: 52.833 ng
RT: 10.151 min Scan# 1371
Delta R.T. 0.000 min
Lab File: BF126479.D
Acq: 4 Jan 2022 21:08

Instrument :
BNA_F
ClientSampleId :







Tgt Ion:232 Resp: 170658
Ion Ratio Lower Upper
232 100
131 54.0 50.0 75.0
130 2.4 2.2 3.2
166 34.8 29.0 43.4



#60
Diethylphthalate
Concen: 49.189 ng
RT: 10.251 min Scan# 1388
Delta R.T. 0.000 min
Lab File: BF126479.D
Acq: 4 Jan 2022 21:08


Tgt Ion:149 Resp: 769962
Ion Ratio Lower Upper
149 100
177 21.0 16.7 25.1
150 12.9 9.8 14.8








#61
4-Chlorophenyl-phenylether
Concen: 47.943 ng
RT: 10.369 min Scan# 1408
Delta R.T. 0.000 min
Lab File: BF126479.D
Acq: 4 Jan 2022 21:08


Instrument :
BNA_F
ClientSampleId :





Tgt Ion:204 Resp: 287386
Ion Ratio Lower Upper
204 100
206 33.0 26.3 39.5
141 67.4 59.0 88.6




#62
4-Nitroaniline
Concen: 43.680 ng
RT: 10.398 min Scan# 1413
Delta R.T. 0.006 min
Lab File: BF126479.D
Acq: 4 Jan 2022 21:08



Tgt Ion:138 Resp: 197843
Ion Ratio Lower Upper
138 100
92 50.9 36.2 76.2
108 59.9 46.7 86.7









#63
Azobenzene
Concen: 45.422 ng
RT: 10.528 min Scan# 1435
Delta R.T. 0.000 min
Lab File: BF126479.D
Acq: 4 Jan 2022 21:08




Instrument :
BNA_F
ClientSampleId :


Tgt Ion: 77	Resp: 915816
Ion Ratio	Lower Upper
77 100	
182 20.2	0.0 39.0
105 23.9	3.2 43.2
51 28.8	8.2 48.2



#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 11.310 min Scan# 1568
Delta R.T. 0.000 min
Lab File: BF126479.D
Acq: 4 Jan 2022 21:08


Tgt Ion:188	Resp: 421865
Ion Ratio	Lower Upper
188 100	
94 13.1	10.1 15.1
80 14.3	11.8 17.6








#65
4,6-Dinitro-2-methylphenol
Concen: 51.417 ng
RT: 10.428 min Scan# 1418
Delta R.T. 0.006 min
Lab File: BF126479.D
Acq: 4 Jan 2022 21:08


Instrument :
BNA_F
ClientSampleId :





Tgt Ion:198 Resp: 117464
Ion Ratio Lower Upper
198 100
51 54.3 32.0 72.0
105 53.9 32.7 72.7




#66
n-Nitrosodiphenylamine
Concen: 47.839 ng
RT: 10.486 min Scan# 1428
Delta R.T. 0.000 min
Lab File: BF126479.D
Acq: 4 Jan 2022 21:08



Tgt Ion:169 Resp: 626135
Ion Ratio Lower Upper
169 100
168 66.4 53.4 80.0
167 36.3 29.4 44.0





#67

4-Bromophenyl-phenylether

Concen: 50.246 ng

RT: 10.857 min Scan# 1491

Delta R.T. 0.000 min

Lab File: BF126479.D

Acq: 4 Jan 2022 21:08

Instrument :

BNA_F

ClientSampleId :





Tgt Ion:248 Resp: 188145

Ion Ratio Lower Upper

248 100

250 95.8 76.2 114.4

141 84.4 66.2 99.4



#68

Hexachlorobenzene

Concen: 51.190 ng

RT: 10.922 min Scan# 1502

Delta R.T. 0.000 min

Lab File: BF126479.D

Acq: 4 Jan 2022 21:08




Tgt Ion:284 Resp: 222962


Ion Ratio Lower Upper

284 100

142 39.4 30.7 46.1

249 27.3 21.1 31.7









#69
 Atrazine
 Concen: 84.604 ng
 RT: 11.016 min Scan# 1518
 Delta R.T. 0.006 min
 Lab File: BF126479.D
 Acq: 4 Jan 2022 21:08




Instrument :
 BNA_F
 ClientSampleId :


Tgt Ion:200 Resp: 190675
 Ion Ratio Lower Upper
 200 100
 173 29.6 9.2 49.2
 215 45.3 23.2 63.2



#70
 Pentachlorophenol
 Concen: 105.133 ng
 RT: 11.122 min Scan# 1536
 Delta R.T. 0.006 min
 Lab File: BF126479.D
 Acq: 4 Jan 2022 21:08


Tgt Ion:266 Resp: 222559
 Ion Ratio Lower Upper
 266 100
 268 66.9 52.3 78.5
 264 61.0 50.4 75.6








#71
Phenanthrene
Concen: 47.867 ng
RT: 11.339 min Scan# 1573
Delta R.T. 0.006 min
Lab File: BF126479.D
Acq: 4 Jan 2022 21:08

Instrument :
BNA_F
ClientSampleId :







Tgt Ion:178 Resp: 1032392
Ion Ratio Lower Upper
178 100
176 19.1 15.7 23.5
179 15.6 12.3 18.5



#72
Anthracene
Concen: 49.117 ng
RT: 11.386 min Scan# 1581
Delta R.T. 0.000 min
Lab File: BF126479.D
Acq: 4 Jan 2022 21:08

Tgt Ion:178 Resp: 1059996
Ion Ratio Lower Upper
178 100
176 18.7 15.0 22.6
179 15.7 12.5 18.7





#73

Carbazole

Concen: 45.343 ng

RT: 11.545 min Scan# 1

Delta R.T. 0.006 min

Lab File: BF126479.D

Acq: 4 Jan 2022 21:08

Instrument :

BNA_F

ClientSampleId :





Tgt Ion:167 Resp: 958221

Ion Ratio Lower Upper

167 100

166 20.9 16.9 25.3

139 13.7 11.3 16.9



#74

Di-n-butylphthalate

Concen: 49.980 ng

RT: 11.874 min Scan# 1664

Delta R.T. 0.000 min

Lab File: BF126479.D

Acq: 4 Jan 2022 21:08




Tgt Ion:149 Resp: 1202487


Ion Ratio Lower Upper

149 100

150 9.3 7.2 10.8


104 5.3 4.2 6.2








#75
Fluoranthene
Concen: 48.452 ng
RT: 12.522 min Scan# 11
Delta R.T. 0.000 min
Lab File: BF126479.D
Acq: 4 Jan 2022 21:08

Instrument :
BNA_F
ClientSampleId :







Tgt Ion:202 Resp: 953760
Ion Ratio Lower Upper
202 100
101 16.5 0.0 35.7
203 17.7 0.0 37.1



#76
Chrysene-d12
Concen: 20.000 ng
RT: 13.951 min Scan# 2017
Delta R.T. 0.000 min
Lab File: BF126479.D
Acq: 4 Jan 2022 21:08

Tgt Ion:240 Resp: 231408
Ion Ratio Lower Upper
240 100
120 14.8 14.4 21.6
236 24.5 20.5 30.7





#77

Benzidine

Concen: 80.857 ng

RT: 12.645 min Scan# 1

Delta R.T. 0.006 min


Lab File: BF126479.D

Acq: 4 Jan 2022 21:08

Instrument :

BNA_F

ClientSampleId :






Tgt Ion:184 Resp: 299314

Ion Ratio Lower Upper

184 100

185 14.0 11.8 17.8

183 10.2 8.2 12.2



#78

Pyrene

Concen: 50.437 ng

RT: 12.751 min Scan# 1813

Delta R.T. 0.000 min

Lab File: BF126479.D

Acq: 4 Jan 2022 21:08




Tgt Ion:202 Resp: 986347


Ion Ratio Lower Upper

202 100

200 20.9 16.6 24.8


203 17.6 13.8 20.8








#79
 Terphenyl-d14
 Concen: 71.657 ng
 RT: 12.898 min Scan# 1838
 Delta R.T. 0.000 min
 Lab File: BF126479.D
 Acq: 4 Jan 2022 21:08

Instrument :
 BNA_F
 ClientSampleId :







Tgt Ion:244 Resp: 853455
 Ion Ratio Lower Upper
 244 100
 212 6.3 5.0 7.4
 122 12.9 13.3 19.9#



#80
 Butylbenzylphthalate
 Concen: 52.369 ng
 RT: 13.374 min Scan# 1919
 Delta R.T. 0.000 min
 Lab File: BF126479.D
 Acq: 4 Jan 2022 21:08


Tgt Ion:149 Resp: 453809
 Ion Ratio Lower Upper
 149 100
 91 77.5 70.5 105.7
 206 18.4 12.6 19.0








#81
Benzo(a)anthracene
Concen: 46.180 ng
RT: 13.939 min Scan# 2008
Delta R.T. 0.000 min
Lab File: BF126479.D
Acq: 4 Jan 2022 21:08

Instrument :
BNA_F
ClientSampleId :







Tgt Ion:228 Resp: 614569
Ion Ratio Lower Upper
228 100
226 26.3 21.4 32.0
229 19.2 15.4 23.2



#82
3,3'-Dichlorobenzidine
Concen: 25.944 ng
RT: 13.898 min Scan# 2008
Delta R.T. -0.006 min
Lab File: BF126479.D
Acq: 4 Jan 2022 21:08

Tgt Ion:252 Resp: 161077
Ion Ratio Lower Upper
252 100
254 65.3 51.0 76.4
126 15.2 13.4 20.0





#83

Chrysene

Concen: 49.422 ng

RT: 13.974 min Scan# 20

Delta R.T. 0.000 min

Lab File: BF126479.D

Acq: 4 Jan 2022 21:08

Instrument :

BNA_F

ClientSampleId :





Tgt Ion:228 Resp: 674654

Ion Ratio Lower Upper

228 100

226 30.4 23.7 35.5

229 19.5 15.5 23.3



#84

Bis(2-ethylhexyl)phthalate

Concen: 54.825 ng

RT: 13.933 min Scan# 2014

Delta R.T. 0.000 min

Lab File: BF126479.D

Acq: 4 Jan 2022 21:08




Tgt Ion:149 Resp: 492364


Ion Ratio Lower Upper

149 100

167 26.5 21.4 32.2

279 2.9 2.4 3.6





#85

Di-n-octyl phthalate

Concen: 58.867 ng

RT: 14.545 min Scan# 21

Delta R.T. 0.000 min

Lab File: BF126479.D

Acq: 4 Jan 2022 21:08

Instrument :

BNA_F

ClientSampleId :





Tgt Ion:149 Resp: 916860

Ion Ratio Lower Upper

149 100

167 1.5 1.1 1.7

43 11.1 9.9 14.9



#86

Perylene-d12

Concen: 20.000 ng

RT: 15.386 min Scan# 2261

Delta R.T. 0.000 min

Lab File: BF126479.D

Acq: 4 Jan 2022 21:08




Tgt Ion:264 Resp: 227186


Ion Ratio Lower Upper

264 100

260 23.8 19.2 28.8


265 20.7 17.4 26.2








#87
 Indeno(1,2,3-cd)pyrene
 Concen: 52.894 ng
 RT: 16.809 min Scan# 2191
 Delta R.T. 0.000 min
 Lab File: BF126479.D
 Acq: 4 Jan 2022 21:08

Instrument :
 BNA_F
 ClientSampleId :







Tgt Ion:276 Resp: 804217
 Ion Ratio Lower Upper
 276 100
 138 35.3 29.2 43.8
 277 24.8 19.8 29.8



#88
 Benzo(b)fluoranthene
 Concen: 50.804 ng
 RT: 14.974 min Scan# 2191
 Delta R.T. 0.000 min
 Lab File: BF126479.D
 Acq: 4 Jan 2022 21:08

Tgt Ion:252 Resp: 687566
 Ion Ratio Lower Upper
 252 100
 253 22.2 17.8 26.8
 125 13.6 12.4 18.6









#89
Benzo(k)fluoranthene
Concen: 48.760 ng
RT: 15.004 min Scan# 2197
Delta R.T. 0.000 min
Lab File: BF126479.D
Acq: 4 Jan 2022 21:08




Instrument :
BNA_F
ClientSampleId :


Tgt Ion:252 Resp: 639259
Ion Ratio Lower Upper
252 100
253 21.9 17.5 26.3
125 13.1 11.8 17.6



#90
Benzo(a)pyrene
Concen: 58.918 ng
RT: 15.327 min Scan# 2251
Delta R.T. 0.000 min
Lab File: BF126479.D
Acq: 4 Jan 2022 21:08

Tgt Ion:252 Resp: 668591
Ion Ratio Lower Upper
252 100
253 21.2 17.1 25.7
125 15.2 13.0 19.6









#91
Dibenzo(a,h)anthracene
Concen: 53.538 ng
RT: 16.827 min Scan# 21
Delta R.T. 0.006 min
Lab File: BF126479.D
Acq: 4 Jan 2022 21:08

Instrument :
BNA_F
ClientSampleId :

Tgt Ion:278 Resp: 656837
Ion Ratio Lower Upper
278 100
139 22.3 18.0 27.0
279 22.5 18.6 27.8



#92
Benzo(g,h,i)perylene
Concen: 54.942 ng
RT: 17.239 min Scan# 2576
Delta R.T. 0.006 min
Lab File: BF126479.D
Acq: 4 Jan 2022 21:08

Tgt Ion:276 Resp: 717771
Ion Ratio Lower Upper
276 100
277 24.1 18.5 27.7
138 32.4 25.9 38.9

