

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF010423\
 Data File : BF131896.D
 Acq On : 04 Jan 2023 13:27
 Operator : CG\JU
 Sample : N6242-04
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 MM-4

Quant Time: Jan 05 05:40:14 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF122422.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Dec 30 02:06:40 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.810	152	63302	20.000	ng	0.00
21) Naphthalene-d8	8.098	136	234130	20.000	ng	-0.01
39) Acenaphthene-d10	9.863	164	134711	20.000	ng	-0.01
64) Phenanthrene-d10	11.357	188	244003	20.000	ng	0.00
76) Chrysene-d12	14.009	240	123458	20.000	ng	0.00
86) Perylene-d12	15.486	264	142352	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.428	112	469143	122.820	ng	0.00
7) Phenol-d6	6.451	99	611473	121.844	ng	0.00
23) Nitrobenzene-d5	7.380	82	430928	73.485	ng	-0.01
42) 2,4,6-Tribromophenol	10.657	330	173495	118.537	ng	-0.01
45) 2-Fluorobiphenyl	9.186	172	761315	77.410	ng	0.00
79) Terphenyl-d14	12.951	244	706461	96.793	ng	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF010423\
Data File : BF131896.D
Acq On : 04 Jan 2023 13:27
Operator : CG\JU
Sample : N6242-04
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
MM-4

Quant Time: Jan 05 05:40:14 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF122422.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Dec 30 02:06:40 2022
Response via : Initial Calibration

